

# Manual Supplement

Manual Title:	Fluke 123/124/125 Service	Supplement Issue:	1
Part Number:	4822 872 05398	Issue Date:	2/14
Print Date:	March 2007	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 3-26, replace **3.5.18 Diode Test Function Test** with:

### **3.5.18 Diode Test Function Test**

Proceed as follows to test the Diode Test function:

1. Connect the test tool to the 5500A as for the previous test (see Figure 3-9).

2. Press  to open the INPUT A MEASUREMENTS menu, and choose:

**MEASURE on A: • DIODE**

3. Set the 5500A to **1 V DC**.

4. Observe the main reading and check to see if it is within 0.975 V and 1.025 V.

5. When you are finished, set the 5500A to Standby.