



America

CERTIFICATE

No. Z1 11 10 14544 076

Holder of Certificate: Fluke Corporation6920 Seaway Boulevard
Everett, WA 98203-9090
USA**Production
Facility(ies):**

14544, 56588, 47951

Certification Mark:**Product:****Multimeters
(True RMS Clamp Meter)****Model(s):****373****Parameters:**

Rated Input Voltage: 3.0 VDC (2 x 1.5V AA batteries IECLR6)
Measuring Ratings: 300V CAT IV / 600V CAT III
Protection Class: II
IP Protection: IPX0
Temperature Ambient: -10°C to +50°C

Tested according to:EN 61010-1:2001
EN 61010-2-032:2002

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

231-1108947-000

Date, 2011-10-13

Page 1 of 1

TUV SUD AMERICA INC • 10 Centennial Drive • Peabody, MA 01960 USA • www.TUVamerica.com



TUV®