

Manual Supplement



Manual Title: 381 Calibration
Part Number: Web Only
Print Date: December 2010
Revision/Date: 1, 6/12

Supplement Issue: **2**
Issue Date: 6/14
Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 59597, 62194

On page 4, Table 1, add:

	Conforms to relevant South Korean EMC standards.
	Conforms to CA/CSA-C22.2 No. 61010-1, second edition, including Amendment 1.

On page 7, replace the EMI, RFI, EMC, RF Specifications with:

EMI, RFI, EMC, RF..... EN 61326-1:2006, EN 61326-2-2:2006

ETSI EN 300 328 V1.7.1:2006

ETSI EN 300 489 V1.8.1:2008

FCC Part 15 Subpart C Sections 15.207, 15.209, 15.249

FCCID : T68-F381

RSS-210





IC: 6627A-F381



Applies to use in Korea only:

Class A Equipment (Industrial Broadcasting & Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.


Under Agency Approvals:

Change: Agency Approvals .. , , , 

To: Agency Approvals... , , , , 

Change #2

On page 14, replace Figure 7 with:



ghn14.eps

Figure 7. Current Test Simulated Voltage