

**Customer:** ROMAN MELNYK TEST ACCOUNT  
35 VANTAGE POINT DRIVE  
ROCHESTER, NY 14624

**PO Number:** abc123

---

**Certificate/SO Number: 1-D9D3D-20-1 Revision 3**

---

**Manufacturer:** Fluke Corporation  
**Model Number:** 87 V  
**Description:** Digital Multimeter  
**Serial Number:** ABC123  
**ID:** 54321

**As-Found:** In Tolerance  
**As-Left:** In Tolerance

**Issue Date:** Mar 10, 2021  
**Calibration Date:** Mar 10, 2021  
**Due Date:** Mar 10, 2022

**Calibrated To:** Manufacturer Specification  
**Calibration Procedure:** 1-AC08977-28


---

This calibration conforms to ISO/IEC 17025:2017, ANSI/NCSL Z540-1-1994, ISO 9001:2015, and ISO 10012:2003. When specified contractually, the requirements of ISO TS16949:2009, 10CFR21, 10CFR50 App. B, ASME NQA-1:2012, and ANSI/NCSL Z540.3-2006 (R2013) are also covered. Standards are traceable to the SI either through the National Institute of Standards & Technology (NIST), the National Research Council of Canada (NRC), other national measurement institutes (NMI) that are signatories to the CIPM Mutual Recognition Arrangement, natural physical constants, intrinsic standards, or ratio type measurements. The results of this report relate only to the item calibrated or tested. Reported data is valid at the time of calibration at the environmental conditions noted. This certificate shall not be reproduced except in full, without the written consent of Transcat Inc.


---

**Calibrated At:**  
35 Vantage Point Dr  
Rochester, NY 14624

**Facility Responsible:**  
35 Vantage Point Dr  
Rochester, NY 14624  
800-828-1470

**Calibrated By:**  
 **Electronically Signed By:**  
Kevin Kita

Kevin Kita      Mar 10, 2021  
Calibration Technician      22:52:26 -05:00

**Reviewed By:**  
 **Electronically Signed By:**  
Jared Payne for

Kevin OGrady      Mar 10, 2021  
Lab Manager      22:53:46 -05:00

**Unit Barcode:**



0900B332281

**Date Received:** January 10, 2020  
**Service Level :** R2

# CALIBRATED BY **TRANSCAT** CERTIFICATE OF CALIBRATION

**Customer:** ROMAN MELNYK TEST ACCOUNT  
35 VANTAGE POINT DRIVE  
ROCHESTER, NY 14624

**PO Number:** abc123

**Certificate/SO Number: 1-D9D3D-20-1 Revision 3**

---

## Traceable Standards

Asset	Manufacturer	Model Number	Description	Cal Date	Due Date	Traceability Number	Use
19-201	Fluke Corporation	5520A-SC1100	Multifunction Calibrator, w/Scope Option	19-Nov-20	30-Jun-21	5-&19-201-22-1	AF/AL

The use of the standard is defined as: AF - used for as-found readings, AL - used for as-left readings.

## Environmental Data

Temperature	Relative Humidity	Temp / RH Asset
68.90°F /20.50°C	32.10%	SOM-184