



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Transcat - Indianapolis**  
**2845 Tobey Drive**  
**Indianapolis, IN 46219**

Fulfils the requirements of

**ISO/IEC 17025:2017**

and national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

In the fields of

## **CALIBRATION and DIMENSIONAL MEASUREMENT**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 12 December 2027

Certificate Number: AC-2489.30



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

**ANSI/NCSL Z540-1-1994 (R2002)**

### Transcat - Indianapolis

2845 Tobey Drive  
Indianapolis, IN 46219  
Shane Smith shane.smith@transcat.com

### CALIBRATION AND DIMENSIONAL MEASUREMENT

ISO/IEC 17025 Accreditation Granted: **02 September 2025**

Certificate Number: **AC-2489.30** Certificate Expiry Date: **12 December 2027**

### CALIBRATION

#### Acoustics and Vibration

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Sound Level Measuring Devices <sup>1</sup>	94 dB 250 Hz 1 kHz 114 dB 250 Hz 1 kHz	0.4 dB 0.4 dB 0.4 dB 0.4 dB	Comparison to Sound Level Calibrator
Sound Level – Measure <sup>1</sup>	(35 to 130) dB 63 Hz to 8 kHz	1.8 dB	Comparison to Sound Level Meter

#### Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,2</sup>	4 pH 7 pH 10 pH	0.012 pH 0.012 pH 0.012 pH	Comparison to Accredited Solutions

### Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Conductivity Meters <sup>1,2</sup>	5 $\mu\text{S}/\text{cm}$ 10 $\mu\text{S}/\text{cm}$ 100 $\mu\text{S}/\text{cm}$ 1 000 $\mu\text{S}/\text{cm}$ 10 000 $\mu\text{S}/\text{cm}$ 100 000 $\mu\text{S}/\text{cm}$ 150 000 $\mu\text{S}/\text{cm}$	0.32 $\mu\text{S}/\text{cm}$ 0.32 $\mu\text{S}/\text{cm}$ 0.82 $\mu\text{S}/\text{cm}$ 3.3 $\mu\text{S}/\text{cm}$ 36 $\mu\text{S}/\text{cm}$ 300 $\mu\text{S}/\text{cm}$ 590 $\mu\text{S}/\text{cm}$	Comparison to Accredited Solutions
Refractometers <sup>1</sup>	(1.3 to 1.395) nD (0 to 30) % Brix	0.0006 nD 0.032 % Brix	Comparison to Accredited Solutions
Gas Monitors and Detectors <sup>1</sup>			
CO (Carbon Monoxide)	0.002 5 % CO 0.005 % CO 0.01 % CO 0.03 % CO	0.000 09 % CO 0.000 12 % CO 0.000 21 % CO 0.000 6 % CO	
CO <sub>2</sub> (Carbon Dioxide)	1 % CO <sub>2</sub> 3 % CO <sub>2</sub> 5 % CO <sub>2</sub> 15 % CO <sub>2</sub>	0.02 % CO <sub>2</sub> 0.092 % CO <sub>2</sub> 0.12 % CO <sub>2</sub> 0.31 % CO <sub>2</sub>	
H <sub>2</sub> S (Hydrogen Sulfide)	0.002 5 % H <sub>2</sub> S	0.000 15 % H <sub>2</sub> S	Comparison to Accredited Gases
C <sub>4</sub> H <sub>8</sub> (Isobutylene)	0.001 % C <sub>4</sub> H <sub>8</sub> 0.01 % C <sub>4</sub> H <sub>8</sub>	0.000 11 % C <sub>4</sub> H <sub>8</sub> 0.000 23 % C <sub>4</sub> H <sub>8</sub>	
LEL (Methane)	20 % LEL 50 % LEL	0.84 % LEL 1.2 % LEL	
O <sub>2</sub> (Oxygen)	18 % O <sub>2</sub> 20.9 % O <sub>2</sub>	0.37 % O <sub>2</sub> 0.42 % O <sub>2</sub>	
SO <sub>2</sub> (Sulfur Dioxide)	0.001 % SO <sub>2</sub>	0.000 022 % SO <sub>2</sub>	

### Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meter Simulation <sup>1</sup> (Electrical)	(0 to 15) pH	0.001 pH	Comparison to Fluke 5522A Multiproduct Calibrator

This Scope of Accreditation, version 008, was last updated on: 10 December 2025 and is valid only when accompanied by the Certificate.

Page 2 of 35

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Conductivity Meter Simulation <sup>1</sup> (Resistance)	(10 to 100) $\mu$ S (101 to 1000) $\mu$ S (1001 to 10 000) $\mu$ S (10 001 to 100 000) $\mu$ S (100 001 to 1 000 000) $\mu$ S	0.08 $\mu$ S 0.09 $\mu$ S 0.08 $\mu$ S 0.06 $\mu$ S 0.04 $\mu$ S	Comparison to Fluke 5522A Multiproduct Calibrator
AC Current – Source <sup>1</sup>	Up to 220 $\mu$ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.025 % of reading + 16 nA 0.016 % of reading + 10 nA 0.011 % of reading + 8 nA 0.028 % of reading + 12 nA 0.11 % of reading + 65 nA  0.025 % of reading + 40 nA 0.016 % of reading + 35 nA 0.011 % of reading + 35 nA 0.021 % of reading + 0.11 $\mu$ A 0.11 % of reading + 0.65 $\mu$ A  0.025 % of reading + 0.4 $\mu$ A 0.016 % of reading + 0.35 $\mu$ A 0.011 % of reading + 0.35 $\mu$ A 0.021 % of reading + 0.55 $\mu$ A 0.11 % of reading + 10 $\mu$ A  0.025 % of reading + 4 $\mu$ A 0.016 % of reading + 3.5 $\mu$ A 0.011 % of reading + 2.5 $\mu$ A 0.021 % of reading + 3.5 $\mu$ A 0.11 % of reading + 10 $\mu$ A  0.025 % of reading + 35 $\mu$ A 0.045 % of reading + 80 $\mu$ A 0.7 % of reading + 0.16 mA	Comparison to Fluke 5730A/03 Multiproduct Calibrator
AC Current – Source <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.048 % of reading + 0.17 mA 0.096 % of reading + 0.38 mA 0.36 % of reading + 0.75 mA	Comparison to Fluke 5730A/03 Multiproduct Calibrator, Fluke 5725A Amplifier

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1</sup>	(11 to 12) A (3 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (12 to 30.2) A (3 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz	0.03 % of reading + 0.78 mA 0.024 % of reading + 0.39 mA 0.03 % of reading + 0.63 mA 0.2 % of reading + 0.78 mA  0.078 % of reading + 7.8 mA 0.055 % of reading + 6.3 mA 0.39 % of reading + 6.3 mA	Comparison to Fluke 5560A Multiproduct Calibrator
AC Current – Source <sup>1</sup>	(10 to 40) Hz (30.2 to 72) A (72 to 100) A (40 to 65) Hz (30.2 to 72) A (72 to 100) A (65 to 300) Hz (30.2 to 72) A (72 to 100) A 300 Hz to 1 kHz (30.2 to 72) A (72 to 100) A	0.032 % of reading + 20 mA 0.028 % of reading + 33 mA  0.032 % of reading + 19 mA 0.028 % of reading + 20 mA  0.037 % of reading + 28 mA 0.041 % of reading + 28 mA  0.081 % of reading + 94 mA 0.086 % of reading + 94 mA	Comparison to Fluke 52120A Transconductance Amplifier, Fluke 5560A Multiproduct Calibrator
AC Clamp-on Ammeters (Toroidal Type) Transformer Type Sensor <sup>1</sup>	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz (1 000 to 1 500) A (45 to 440) Hz	0.24 % of reading + 19 mA 0.62 % of reading + 21 mA  0.24 % of reading + 70 mA 0.61 % of reading + 78 mA  0.51% of reading + 0.31 A	Comparison to Fluke 5522A Multiproduct Calibrator, Fluke 5560A Multiproduct Calibrator, Fluke 5500A/Coil 50-turn Coil
AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor <sup>1</sup>	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz (1 000 to 1 500) A (45 to 440) Hz	0.5 % of reading + 0.19 A 0.9 % of reading + 0.19 A  0.52 % of reading + 0.7 A 0.9 % of reading + 0.7 A  0.53% of reading + 0.31 A	Comparison to Fluke 5522A Multiproduct Calibrator, Fluke 5560A Multiproduct Calibrator, Fluke 5500A/Coil 50-turn Coil

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor <sup>1</sup>	(1.5 to 6) kA (10 to 300) Hz (1 to 2) kA (300 to 440) Hz (2 to 6) kA (300 to 440) Hz	0.6 % of reading 0.8 % of reading 0.66 % of reading	Comparison to Fluke 5522A Multiproduct Calibrator, Fluke 52120A Transconductance Amplifier, 3 kA and 6 kA Coil
AC Current – Measure <sup>1</sup>	Up to 200 $\mu$ A 1 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz (0.2 to 2) mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz (2 to 20) mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz (20 to 200) mA 1 Hz to 10 Hz 10 Hz to 10 kHz (10 to 30) kHz	0.031 % of reading + 20 nA 0.071 % of reading + 20 nA 0.4 % of reading + 20 nA 0.031 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A 0.071 % of reading + 0.2 $\mu$ A 0.4 % of reading + 0.2 $\mu$ A 0.031 % of reading + 2 $\mu$ A 0.03 % of reading + 2 $\mu$ A 0.071 % of reading + 2 $\mu$ A 0.4 % of reading + 2 $\mu$ A 0.031 % of reading + 20 $\mu$ A 0.029 % of reading + 20 $\mu$ A 0.063 % of reading + 20 $\mu$ A	Comparison to Fluke 8508A 8.5 Digit Multimeter
AC Current – Measure <sup>1</sup>	(0.2 to 2) A 10 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (2 to 20) A 10 Hz to 2 kHz (2 to 10) kHz	0.062 % of reading + 0.2 mA 0.074 % of reading + 0.2 mA 0.3 % of reading + 0.2 mA 0.082 % of reading + 2 mA 0.25 % of reading + 2 mA	Comparison to Fluke 8508A 8.5 Digit Multimeter
AC Voltage – Source <sup>1</sup>	Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.16 % of reading + 4 $\mu$ V 0.1 % of reading + 4 $\mu$ V 0.078 % of reading + 4 $\mu$ V 0.13 % of reading + 4 $\mu$ V 0.17 % of reading + 5 $\mu$ V 0.33 % of reading + 10 $\mu$ V 0.47 % of reading + 20 $\mu$ V 0.58 % of reading + 20 $\mu$ V	Comparison to Fluke 5730A/03 Multiproduct Calibrator

This Scope of Accreditation, version 008, was last updated on: 10 December 2025 and is valid only when accompanied by the Certificate.

Page 5 of 35

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz  (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz  (0.22 to 2.2) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz  (2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.042 % of reading + 4 $\mu$ V 0.03 % of reading + 4 $\mu$ V 0.014 % of reading + 4 $\mu$ V 0.03 % of reading + 4 $\mu$ V 0.058 % of reading + 5 $\mu$ V 0.12 % of reading + 10 $\mu$ V 0.16 % of reading + 20 $\mu$ V 0.27 % of reading + 20 $\mu$ V  0.028 % of reading + 12 $\mu$ V 0.011 % of reading + 7 $\mu$ V 0.009 % of reading + 7 $\mu$ V 0.021 % of reading + 7 $\mu$ V 0.047 % of reading + 17 $\mu$ V 0.091 % of reading + 20 $\mu$ V 0.14 % of reading + 25 $\mu$ V 0.28 % of reading + 45 $\mu$ V  0.027 % of reading + 40 $\mu$ V 0.01 % of reading + 15 $\mu$ V 0.005 % of reading + 8 $\mu$ V 0.008 % of reading + 10 $\mu$ V 0.012 % of reading + 30 $\mu$ V 0.043 % of reading + 80 $\mu$ V 0.01 % of reading + 0.2 mV 0.18 % of reading + 0.3 mV  0.028 % of reading + 0.4 mV 0.01 % of reading + 0.15 mV 0.005 % of reading + 50 $\mu$ V 0.008 % of reading + 0.1 mV 0.011 % of reading + 0.2 mV 0.03 % of reading + 0.6 mV 0.1 % of reading + 2 mV 0.17 % of reading + 3.2 mV	Comparison to Fluke 5730A/03 Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 o 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.028 % of reading + 4 mV 0.01 % of reading + 1.5 mV 0.006 % of reading + 0.6 mV 0.009 % of reading + 1 mV 0.016 % of reading + 2.5 mV 0.09 % of reading + 16 mV 0.44 % of reading + 40 mV 0.8 % of reading + 80 mV	Comparison to Fluke 5730A/03 Multiproduct Calibrator
AC Voltage – Source <sup>1</sup>	(220 to 1 100) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	0.011 % of reading + 4 mV 0.017 % of reading + 6 mV 0.061 % of reading + 11 mV	Comparison to Fluke 5730A/03 Multiproduct Calibrator, Fluke 5725A Amplifier
AC Voltage – Source <sup>1</sup> Extended Frequency Ranges	(220 to 750) V (30 to 50) kHz (50 to 100) kHz	0.061 % of reading + 11 mV 0.23 % of reading + 45 mV	Comparison to Fluke 5730A/03 Multiproduct Calibrator, Fluke 5725A Amplifier
AC Voltage – Measure <sup>1</sup>	Up to 200 mV (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz (0.1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (0.2 to 2) V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz (0.1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.021 % of reading + 14 $\mu$ V 0.017 % of reading + 4 $\mu$ V 0.014 % of reading + 4 $\mu$ V 0.013 % of reading + 2 $\mu$ V 0.016 % of reading + 4 $\mu$ V 0.04 % of reading + 8 $\mu$ V 0.09 % of reading + 20 $\mu$ V  0.019 % of reading + 0.12 mV 0.014 % of reading + 20 $\mu$ V 0.011 % of reading + 20 $\mu$ V 89 $\mu$ V/V of reading + 20 $\mu$ V 0.013 % of reading + 20 $\mu$ V 0.026 % of reading + 40 $\mu$ V 0.068 % of reading + 0.2 mV 0.37 % of reading + 2 mV 1.3 % of reading + 20 mV	Comparison to Fluke 8508A 8.5 Digit Multimeter

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(2 to 20) V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz (0.1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz  (20 to 200) V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz (0.1 to 2) kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz  (200 to 1 000) V (1 to 10) Hz (10 to 40) Hz (0.04 to 10) kHz (10 to 30) kHz (30 to 100) kHz	0.019 % of reading + 1.2 mV 0.014 % of reading + 0.2 mV 0.011 % of reading + 0.2 mV 90 $\mu$ V/V + 0.2 mV 0.013 % of reading + 0.2 mV 0.026 % of reading + 0.4 mV 0.068 % of reading + 2 mV 0.37 % of reading + 20 mV 2.2 % of reading + 0.2V  0.02 % of reading + 12 mV 0.014 % of reading + 2 mV 0.011 % of reading + 2 mV 90 $\mu$ V/V + 2 mV 0.013 % of reading + 2 mV 0.026 % of reading + 4 mV 0.068 % of reading + 20 mV  0.019 % of reading + 70 mV 0.015 % of reading + 20 mV 0.014 % of reading + 20 mV 0.027 % of reading + 40 mV 0.071 % of reading + 0.2 V	Comparison to Fluke 8508A 8.5 Digit Multimeter
AC High Voltage – Measure <sup>1</sup>	(0.7 to 9) kV 10 mHz to 10 Hz (10 to 30) Hz (30 to 50) Hz (50 to 70) Hz (70 to 100) Hz (100 to 200) Hz (200 to 450) Hz (450 to 600) Hz	0.15 % reading + 0.17 V 0.15 % reading + 0.17 V 0.14 % reading + 0.17 V 0.48 % reading + 0.17 V 0.87 % reading + 0.17 V	Comparison to Vitrek 4700 High Voltage Meter

## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC High Voltage – Measure <sup>1</sup>	(9 to 10) kV 10 mHz to 10 Hz (10 to 30) Hz (30 to 50) Hz (50 to 70) Hz (70 to 100) Hz (100 to 200) Hz (200 to 450) Hz (450 to 600) Hz	0.15 % reading + 0.69 V 0.15 % reading + 0.69 V 0.14 % reading + 0.69 V 0.48 % reading + 0.69 V 0.87 % reading + 0.69 V	Comparison to Vitrek 4700 High Voltage Meter
AC High Voltage – Measure <sup>1</sup>	(10 to 28) kV 60 Hz	5.2 % of reading + 1.7 V	Comparison to Keysight 34401A 6.5 Digit Multimeter, Fluke 80K-40 High Voltage Probe
Capacitance – Source <sup>1</sup> (Simulation)	Up to 1.2 nF (20 to 40) Hz (40 to 100) Hz 100 Hz to 10 kHz (10 to 12) kHz (12 to 14) kHz (1.2 to 3) nF (10 to 150) Hz 150 Hz to 5 kHz (5 to 6) kHz (6 to 8) kHz (3 to 12) nF 10 Hz to 5 kHz (5 to 6) kHz (6 to 8) kHz (12 to 30) nF (20 to 200) Hz 200 Hz to 1.3 kHz (1.3 to 2.7) kHz (2.7 to 3.7) kHz (30 to 120) nF 10 Hz to 1.3 kHz (1.3 to 2.7) kHz (2.7 to 3.7) kHz	0.87 % of reading + 1.6 pF 0.49 % of reading + 1.6 pF 0.094 % of reading + 1.6 pF 0.48 % of reading + 1.6 pF 0.87 % of reading + 1.6 pF  0.49 % of reading + 3.9 pF 0.09 % of reading + 3.9 pF 0.48 % of reading + 3.9 pF 0.87 % of reading + 3.9 pF  0.09 % of reading + 3.9 pF 0.48 % of reading + 3.9 pF 0.87 % of reading + 3.9 pF  0.49 % of reading + 24 pF 0.1 % of reading + 24 pF 0.49 % of reading + 24 pF 0.88 % of reading + 24 pF  0.1 % of reading + 24 pF 0.49 % of reading + 24 pF 0.88 % of reading + 24 pF	Comparison to Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source <sup>1</sup> (Simulation)	(0.12 to 1.2) $\mu$ F (2 to 310) Hz (310 to 800) Hz 800 Hz to 1.1 kHz (1.2 to 12) $\mu$ F 500 mHz to 110 Hz (110 to 250) Hz (250 to 350) Hz (12 to 120) $\mu$ F (100 to 500) mHz 500 mHz to 40 Hz (40 to 80) Hz (80 to 110) Hz (0.12 to 1.2) mF 100 mHz to 11 Hz (11 to 18) Hz (18 to 25) Hz (1.2 to 12) mF 30 mHz to 4 Hz (4 to 6) Hz (6 to 8) Hz (12 to 120) mF 10 mHz to 1.3 Hz (1.3 to 1.7) Hz (1.7 to 2.5) Hz	0.1 % of reading + 0.24 nF 0.49 % of reading + 0.24 nF 0.88 % of reading + 0.24 nF  0.1 % of reading + 2.4 nF 0.49 % of reading + 2.4 nF 0.88 % of reading + 2.4 nF  0.5 % of reading + 20 nF 0.12 % of reading + 20 nF 0.5 % of reading + 20 nF 0.89 % of reading + 20 nF  0.2 % of reading + 0.2 $\mu$ F 0.58 % of reading + 0.2 $\mu$ F 1 % of reading + 0.2 $\mu$ F  0.19 % of reading + 2.4 $\mu$ F 0.58 % of reading + 2.4 $\mu$ F 1 % of reading + 2.4 $\mu$ F  0.39 % of reading + 24 $\mu$ F 0.78 % of reading + 24 $\mu$ F 1.2 % of reading + 24 $\mu$ F	Comparison to Fluke 5560A Multiproduct Calibrator
Capacitance – Measure <sup>1</sup>	1 kHz (1 to 1 000) pF (1 to 1 000) nF (1 to 1 000) $\mu$ F	0.5 fF/pF + 23 fF 0.5 pF/nF + 0.88 pF 4.2 nF/ $\mu$ F	Comparison to Hameg LCR Meter
DC Current – Source <sup>1</sup>	(0 to 220) $\mu$ A (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	40 $\mu$ A/A + 6 nA 35 $\mu$ A/A + 7 nA 35 $\mu$ A/A + 40 nA 45 $\mu$ A/A + 0.7 $\mu$ A 80 $\mu$ A/A + 12 $\mu$ A	Comparison to Fluke 5730A/03 Multiproduct Calibrator
DC Current – Source <sup>1</sup>	(2.2 to 3.1) A (3.1 to 12) A (12 to 30.2) A	0.024 % of reading + 0.12 mA 0.024 % of reading + 0.2 mA 0.078 % of reading + 0.39 mA	Comparison to Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor <sup>1</sup>	(20 to 55) A (55 to 150) A (150 to 550) A (550 to 1 000) A (1 000 to 1 500) A (1 500 to 5 000) A	0.47 % of reading + 0.11 A 0.41 % of reading + 0.11 A 0.45 % of reading + 0.39 A 0.51 % of reading + 0.39 A 0.6 % of reading + 20 mA 0.58 % of reading	Comparison to Fluke 5522A Multiproduct Calibrator, Fluke 5560A Multiproduct Calibrator, Fluke 5500A/COIL, Fluke 55120A Transconductance Amplifier, 3 kA and 6 kA Coils
DC Current – Measure <sup>1</sup>	Up to 200 $\mu$ A (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A (2 to 20) A	13 $\mu$ A/A + 0.31 nA 13 $\mu$ A/A + 3.1 nA 14 $\mu$ A/A + 31 nA 47 $\mu$ A/A + 0.62 $\mu$ A 0.18 mA/A + 12 $\mu$ A 0.39 mA/A + 0.31 mA	Comparison to Fluke 8508A 8.5 Digit Multimeter
DC Current – Measure <sup>1</sup>	(0 to 500) A	0.2 mA/A + 0.16 A	Comparison to Fluke 8508A 8.5 Digit Multimeter, DC Current Shunt
DC Voltage – Source <sup>1</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	7.5 $\mu$ V/V + 0.4 $\mu$ V 5 $\mu$ V/V + 0.7 $\mu$ V 3.5 $\mu$ V/V + 2.5 $\mu$ V 3.5 $\mu$ V/V + 4 $\mu$ V 5 $\mu$ V/V + 40 $\mu$ V 6.5 $\mu$ V + 0.4 mV	Comparison to Fluke 5730A/03 Multiproduct Calibrator
DC Voltage – Source <sup>1</sup> (Fixed Point)	10 V	0.31 $\mu$ V/V	Comparison to Fluke 732A DC Voltage Reference Standard
DC Voltage – Measure <sup>1</sup>	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 000) V	5.2 $\mu$ V/V + 90 nV 3.6 $\mu$ V/V + 0.39 $\mu$ V 3.5 $\mu$ V/V + 3.9 $\mu$ V 5.5 $\mu$ V/V + 39 $\mu$ V 5.5 $\mu$ V/V + 0.47 mV	Comparison to Fluke 8508A 8.5 Digit Multimeter
DC High Voltage – Measure <sup>1</sup>	(1 to 10) kV	0.039 % of reading + 92 mV	Comparison to Vitrek 4700 High Voltage Meter

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC High Voltage – Measure <sup>1</sup>	(10 to 20) kV (20 to 35) kV (35 to 40) kV	2 % of reading 1 % of reading 2 % of reading	Comparison to Keysight 34401A 6.5 Digit Multimeter, Fluke 80K-40 High Voltage Probe
AC Power – Source <sup>1,2,4</sup> PF = 1	(3.3 to 9) mA  (9 to 33) mA  (33 to 90) mA  (90 to 330) mA  (330 to 900) mA  900 mA to 2.2 A  (2.2 to 4.5) A  (4.5 to 20.5) A	(10 to 65) Hz 110 $\mu$ W to 3 mW 3 mW to 9 W  (10 to 65) W 300 $\mu$ W to 10 mW 10 mW to 33 W  (10 to 65) Hz (1 to 30) mW 30 mW to 90 W  (10 to 65) Hz (3 to 100) mW 100 mW to 300 W  (10 to 65) Hz (11 to 300) mW 300 mW to 900 W  (10 to 65) Hz (30 to 720) mW 0.72 W to 2 kW  (10 to 65) Hz 80 mW to 1.4 W 1.4 W to 4.5 kW  (10 to 65) Hz 150 mW to 20kW	0.13 % of reading 0.077 % of reading  0.089 % of reading 0.077 % of reading  0.071 % of reading 0.057 % of reading  0.089 % of reading 0.078 % of reading  0.071 % of reading 0.081 % of reading  0.089 % of reading 0.079 % of reading  0.088 % of reading 0.18 % of reading  0.17 % of reading
			Comparison to Fluke 5520A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Power – Source <sup>1</sup>			
330 µW to 330 mA	11 µW to 1.1 mW (1.1 to 110) mW 110 mW to 110 W (110 to 330) W	0.024 % of reading 0.027 % of reading 0.024 % of reading 0.018 % of reading	
330 mA to 3 A	11 µW to 110 mW 110 mW to 990 W 990 W to 3 kW	0.044 % of reading 0.053 % of reading 0.01 % of reading	Comparison to Fluke 5520A Multiproduct Calibrator
(3 to 20.5) A	99 mW to 0.99 W 0.99 W to 6.8 kW (6.8 to 20.5) kW	0.088 % of reading 0.07 % of reading 0.04 % of reading	
Phase – Source <sup>1</sup>	(0 to 360)° (3 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.08° 0.19° 0.39° 1.9° 3.9° 7.8°	Comparison to Fluke 5560A Multiproduct Calibrator
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Pt 385, 100 Ω (-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C Pt 385, 200 Ω (-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.039 °C 0.039 °C 0.054 °C 0.07 °C 0.078 °C 0.093 °C 0.18 °C 0.031 °C 0.031 °C 0.031 °C 0.039 °C 0.093 °C 0.1 °C 0.11 °C 0.12 °C	Comparison to Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Pt 385, 500 $\Omega$ (-200 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.039 °C	
	(100 to 260) °C	0.047 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.062 °C	
	(400 to 600) °C	0.07 °C	
	(600 to 630) °C	0.085 °C	
	Pt 385, 1 k $\Omega$ (-200 to -80) °C	0.023 °C	
	(-80 to 0) °C	0.023 °C	
	(0 to 100) °C	0.031 °C	
	(100 to 260) °C	0.039 °C	
	(260 to 300) °C	0.047 °C	
	(300 to 400) °C	0.054 °C	
	(400 to 600) °C	0.054 °C	
	(600 to 630) °C	0.18 °C	
Pt 3916, 100 $\Omega$ (-200 to -190) °C		0.19 °C	Comparison to Fluke 5560A Multiproduct Calibrator
	(-190 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.047 °C	
	(100 to 260) °C	0.054 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.07 °C	
	(400 to 600) °C	0.078 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3926, 100 $\Omega$ (-200 to -80) °C	0.039 °C	
Ni 385, 120 $\Omega$ (-80 to 0) °C	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.054 °C	
	(100 to 300) °C	0.07 °C	
	(300 to 400) °C	0.078 °C	
	(400 to 630) °C	0.093 °C	
	Cu 427, 10 $\Omega$ (-80 to 260) °C	0.23 °C	

This Scope of Accreditation, version 008, was last updated on: 10 December 2025 and is valid only when accompanied by the Certificate.

Page 14 of 35

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Cu 428, 50 Ω (-180 to 200) °C Cu 428, 100 Ω (-180 to 40) °C (40 to 200) °C	0.31 °C 0.31 °C 0.51 °C	Comparison to Fluke 5560A Multiproduct Calibrator
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type B (600 to 800) °C (800 to 1 000) °C (1 000 to 1 550) °C (1 550 to 1 820) °C Type C (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 315) °C Type D (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 315) °C Type E (-250 to -150) °C (-150 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C Type G (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 315) °C	0.34 °C 0.26 °C 0.23 °C 0.26 °C 0.19 °C 0.16 °C 0.2 °C 0.35 °C 0.61 °C 0.19 °C 0.16 °C 0.2 °C 0.34 °C 0.61 °C 0.31 °C 0.11 °C 0.09 °C 0.12 °C 0.16 °C 0.39 °C 0.26 °C 0.2 °C 0.33 °C 0.6 °C	Comparison to Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	<p>Type J            (-210 to -100) °C            (-100 to -30) °C            (-30 to 150) °C            (150 to 760) °C            (760 to 1 200) °C</p> <p>Type K            (-200 to -100) °C            (-100 to -25) °C            (-25 to 120) °C            (120 to 1 000) °C            (1 000 to 1 372) °C</p> <p>Type L            (-200 to -100) °C            (-100 to 800) °C            (800 to 900) °C</p> <p>Type N            (-200 to -100) °C            (-100 to -25) °C            (-25 to 120) °C            (120 to 410) °C            (410 to 1 300) °C</p> <p>Type R            (0 to 250) °C            (250 to 400) °C            (400 to 1 000) °C            (1 000 to 1 767) °C</p> <p>Type S            (0 to 250) °C            (250 to 400) °C            (400 to 1 000) °C            (1 000 to 1 767) °C</p> <p>Type T            (-250 to -150) °C            (-150 to 0) °C            (0 to 120) °C            (120 to 400) °C</p>	0.19 °C 0.1 °C 0.09 °C 0.11 °C 0.16 °C  0.22 °C 0.1 °C 0.09 °C 0.16 °C 0.27 °C  0.24 °C 0.16 °C 0.09 °C  0.26 °C 0.12 °C 0.09 °C 0.09 °C 0.16 °C  0.4 °C 0.23 °C 0.21 °C 0.26 °C  0.33 °C 0.24 °C 0.25 °C 0.32 °C  0.47 °C 0.16 °C 0.1 °C 0.09 °C	Comparison to Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type BP (0 to 1 000) °C (1 000 to 2 000) °C (2 000 to 2 500) °C Type XK (-200 to 300) °C (300 to 800) °C	0.31 °C 0.47 °C 0.62 °C 0.16 °C 0.23 °C	Comparison to Fluke 5560A Multiproduct Calibrator
Magnetic Flux Density – Source <sup>1,5</sup> (Gauss Meters)	(0 to 35) G	0.26 % of reading	Comparison to Digital Multimeter, Helmholtz Coils
Resistance – Source <sup>1</sup> (Simulation)	(0 to 12) Ω (12 to 120) Ω (0.12 to 1.2) kΩ (1.2 to 12) kΩ (12 to 120) kΩ (0.12 to 1.2) MΩ (1.2 to 12) MΩ (12 to 120) MΩ (0.12 to 1.2) GΩ	0.002 % of reading + 0.78 mΩ 0.002 % of reading + 0.78 mΩ 0.002 % of reading + 1.6 mΩ 0.002 % of reading + 16 mΩ 0.002 % of reading + 0.16 Ω 0.002 % of reading + 1.6 Ω 0.002 8 % of reading + 24 Ω 0.034 % of reading + 2 kΩ 0.32 % of reading + 78 kΩ	Comparison to Fluke 5560A Multiproduct Calibrator (4-wire Configuration)
Resistance – Source <sup>1</sup> (Fixed Artifacts)	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	40 μΩ 95 μΩ/Ω 95 μΩ/Ω 23 μΩ/Ω 23 μΩ/Ω 10 μΩ/Ω 10 μΩ/Ω 6.5 μΩ/Ω 6.5 μΩ/Ω 6.5 μΩ/Ω 6.5 μΩ/Ω 8.5 μΩ/Ω 8.5 μΩ/Ω 13 μΩ/Ω 18 μΩ/Ω 40 μΩ/Ω 47 μΩ/Ω 0.011 % of reading	Comparison to Fluke 5730A/03 Multiproduct Calibrator
Resistance – Source <sup>1</sup> (Fixed Artifacts)	1 Ω 10 kΩ	9.3 μΩ/Ω 7.6 μΩ/Ω	Comparison to Fluke 742A Resistance Standards

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source <sup>1</sup> (Fixed Artifacts)	1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ 10 GΩ 100 GΩ	15 µΩ/Ω 15 µΩ/Ω 20 µΩ/Ω 25 µΩ/Ω 31 µΩ/Ω 0.013 % of reading 0.07 % of reading 0.25 % of reading 0.45 % of reading	Comparison to Standard Resistors
DC Resistance – Measure <sup>1</sup> Normal Mode	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	17 µΩ/Ω + 3.9 µΩ 9.5 µΩ/Ω + 14 µΩ 7.8 µΩ/Ω + 47 µΩ 8.1 µΩ/Ω + 0.47 mΩ 7.8 µΩ/Ω + 4.7 mΩ 8.6 µΩ/Ω + 47 mΩ 9.5 µΩ/Ω + 93 mΩ 27 µΩ/Ω + 9.3 Ω 0.12 mΩ/Ω + 0.93 kΩ 0.14 % of reading + 93 kΩ	Comparison to Fluke 8508A 8.5 Digit Multimeter
DC Resistance – Measure <sup>1</sup> Low Current Mode	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	17 µΩ/Ω + 3.9 µΩ 9.4 µΩ/Ω + 14 µΩ 8 µΩ/Ω + 0.14 mΩ 8.2 µΩ/Ω + 1.4 mΩ 7.9 µΩ/Ω + 14 mΩ 8.6 µΩ/Ω + 93 mΩ 21 µΩ/Ω + 0.93 Ω 88 µΩ/Ω + 93 Ω 0.14 % of reading + 93 kΩ 0.14 % of reading + 0.93 MΩ	Comparison to Fluke 8508A 8.5 Digit Multimeter
DC Resistance – Measure <sup>1</sup> High Voltage Mode	(2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ (2 to 20) GΩ	25 µΩ/Ω + 9.3 Ω 70 µΩ/Ω + 0.93 kΩ 0.19 mΩ/Ω + 93 kΩ 0.14 % of reading + 9.3 MΩ	Comparison to Fluke 8508A opt 001 8.5 Digit Multimeter
Inductance – Source <sup>1</sup> (Simulation)	Up to 120 µH (490 to 550) Hz 550 Hz to 1 kHz 1 kHz (1 to 13) kHz (13 to 17) kHz	0.93 % of reading + 0.16 µH 0.35 % of reading + 0.16 µH 0.16 % of reading + 0.16 µH 0.35 % of reading + 0.16 µH 0.93 % of reading + 0.16 µH	Comparison to Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Source <sup>1</sup> (Simulation)	(0.12 to 1.2) mH (260 to 330) Hz 330 Hz to 1 kHz 1 kHz (1 to 1.6) kHz (1.6 to 2.5) kHz (1.2 to 3.3) mH 500 mHz to 110 Hz 110 Hz (110 to 800) Hz (800 to 980) Hz (3.3 to 12) mH 500 mHz to 110 Hz 110 Hz 110 Hz to 1 kHz (1 to 1.4) kHz (12 to 83) mH 100 mHz to 100 Hz 100 Hz (100 to 180) Hz (180 to 230) Hz (83 to 120) mH 100 mHz to 100 Hz 100 Hz (100 to 320) Hz 320 Hz to 1 kHz (120 to 650) mH 50 mHz to 10 Hz 10 Hz (10 to 30) Hz (30 to 55) Hz (0.65 to 1.2) H 50 mHz to 10 Hz 10 Hz (10 to 100) Hz (100 to 170) Hz	0.87 % of reading + 0.78 $\mu$ H 0.29 % of reading + 0.78 $\mu$ H 0.094 % of reading + 0.78 $\mu$ H 0.29 % of reading + 0.78 $\mu$ H 0.87 % of reading + 0.78 $\mu$ H  0.3 % of reading + 7.8 $\mu$ H 0.094 % of reading + 7.8 $\mu$ H 0.29 % of reading + 7.8 $\mu$ H 0.87 % of reading + 7.8 $\mu$ H  0.29 % of reading + 7.8 $\mu$ H 0.093 % of reading + 7.8 $\mu$ H 0.29 % of reading + 7.8 $\mu$ H 0.87 % of reading + 7.8 $\mu$ H  0.29 % of reading + 78 $\mu$ H 0.093 % of reading + 78 $\mu$ H 0.29 % of reading + 78 $\mu$ H 0.87 % of reading + 78 $\mu$ H  0.29 % of reading + 78 $\mu$ H 0.093 % of reading + 78 $\mu$ H 0.3 % of reading + 78 $\mu$ H 0.87 % of reading + 78 $\mu$ H  0.32 % of reading + 0.78 mH 0.14 % of reading + 0.78 mH 0.32 % of reading + 0.78 mH 0.9 % of reading + 0.78 mH  0.32 % of reading + 0.78 mH 0.14 % of reading + 0.78 mH 0.32 % of reading + 0.78 mH 0.9 % of reading + 0.78 mH	Comparison to Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
Inductance – Source <sup>1</sup> (Simulation)	(1.2 to 5.5) H 10 mHz to 3 Hz 3 Hz (3 to 8) Hz (8 to 16) Hz (5.5 to 12) H 10 mHz to 3 Hz 3 Hz (3 to 19) Hz (19 to 37) Hz (12 to 30) H 5 mHz to 2 Hz 2 Hz (2 to 4) Hz (4 to 9) Hz (30 to 120) H 5 mHz to 2 Hz 2 Hz (2 to 7) Hz (7 to 14) Hz	0.36 % of reading + 7.8 mH 0.17 % of reading + 7.8 mH 0.36 % of reading + 7.8 mH 0.93 % of reading + 7.8 mH 0.36 % of reading + 7.8 mH 0.17 % of reading + 7.8 mH 0.36 % of reading + 7.8 mH 0.93 % of reading + 7.8 mH 0.39 % of reading + 78 mH 0.21 % of reading + 78 mH 0.39 % of reading + 78 mH 1 % of reading + 78 mH 0.39 % of reading + 78 mH 0.21 % of reading + 78 mH 0.39 % of reading + 78 mH 1 % of reading + 78 mH	Comparison to Fluke 5560A Multiproduct Calibrator	
Inductance – Source <sup>1</sup> (Variable Artifact)	1 kHz 1 $\mu$ H to 999 mH	0.5 % of reading	Comparison to Inductance Decade Box characterized with Hameg LCR Meter	
Inductance – Measure <sup>1</sup>	1 kHz 1 $\mu$ H to 100 H	0.5 % of reading	Comparison to Hameg LCR Meter	
Oscilloscopes <sup>1,2</sup>	Amplitude DC into 50 $\Omega$ load into 1 M $\Omega$ load Amplitude Square Wave into 50 $\Omega$ load into 1 M $\Omega$ load	(-6.6 to 6.6) V (-130 to 130) V 10 Hz to 10 kHz 1 mVp-p to 6.6 Vp-p 10 Hz to 1 kHz 1 mVp-p to 130 Vp-p (1 to 10) kHz 1 mVp-p to 130 Vp-p	0.22 % of reading + 31 $\mu$ V 0.12 % of reading + 31 $\mu$ V 0.22 % of reading + 31 $\mu$ V 0.078 % of reading + 31 $\mu$ V 0.19 % of reading + 31 $\mu$ V	Comparison to Fluke 55xxA/11 Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes <sup>1,2</sup>			
Timing – Generate into 50 Ω load	1 ns to 20 ms 50 ms 100 ms 200 ms 500 ms 1 s 2 s 5 s	0.000 22 % reading 0.005 9 % reading 0.009 8 % reading 0.018 % reading 0.041 % reading 0.08 % reading 0.16 % reading 0.39 % reading	
Rise Time – Generate into 50 Ω Load Rate: 1 kHz to 2 MHz Rate: 2 MHz to 10 MHz	5 mVp-p to 2.5 Vp-p (200 to 300) ps (250 to 350) ps	50 ps 50 ps	
Leveled Sine Wave – Generate into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz 100 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	1.8 % of reading + 0.23 mV 2.8 % of reading + 0.23 mV 3.2 % of reading + 0.23 mV 4 % of reading + 0.23 mV 5.5 % of reading + 0.23 mV	Comparison to Fluke 55xxA/11 Multiproduct Calibrator
Bandwidth/Flatness – Measure (50 kHz Reference) into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	1.4 % of reading + 78 µV 1.8 % of reading + 78 µV 3.2 % of reading + 78 µV 4 % of reading + 78 µV	

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes <sup>1,2</sup>			
Input Impedance – Measure into 50 Ω load into 1 MΩ load	(40 to 60) Ω (0.5 to 1.5) MΩ	0.082 % of reading 0.081 % of reading	
Input Capacitance – Measure	(5 to 50) pF	3.9 % of reading + 0.39 pF	
Wave Generator – Source Amplitude (Sine, Square, Triangle)	10 Hz to 10 kHz		Comparison to Fluke 55xxA/11 Multiproduct Calibrator
into 50 Ω load into 1 MΩ load	1.8 mVp-p to 2.5 Vp-p 1.8 mVp-p to 55 Vp-p	2.3 % of reading + 78 μV 2.3 % of reading + 78 μV	
Frequency	10 Hz to 10 kHz	0.001 9 % of reading + 12 mHz	

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks, Length Standards <sup>3</sup>			
Standard Size Length	(0.005 to 0.5) in (0.55 to 1) in 2 in 3 in 4 in 5 in 6 in 7 in 8 in 10 in 12 in 16 in 20 in	3.5 μin 4 μin 6.9 μin 9.3 μin 12 μin 15 μin 18 μin 21 μin 23 μin 29 μin 34 μin 45 μin 55 μin	Mechanical Comparison using Master Gage Blocks

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks, Length Standards <sup>3</sup> Non-standard Size Length	(0.001 to 0.005) in (0.005 to 1) in (1 to 6) in (6 to 20) in (20 to 40) in	5.8 $\mu$ in (5.9 + 1L) $\mu$ in (2.9 + 3.8L) $\mu$ in (1.9 + 4L) $\mu$ in (4.7 + 3.8L) $\mu$ in	Comparison to Universal Length Measuring Machine, Master Gage Blocks
Feeler Gages, <sup>1</sup> Shims	Up to 1 in Up to 25.4 mm	15 $\mu$ in 0.38 $\mu$ m	Comparison to Universal Length Measuring Machine, Gage Blocks
Height Masters <sup>1</sup>	Up to 24 in Up to 600 mm	24 $\mu$ in 0.61 $\mu$ m	Comparison to Indicator with Gage Amplifier, Gage Blocks, Surface Plate
Height Gages <sup>1,3,5</sup>	Up to 60 in Up to 1 500 mm	(8 + 3.5L) $\mu$ in (0.2 + 0.0035L) $\mu$ m	Comparison to Gage Blocks, Surface Plate
Micrometers <sup>1,3,5</sup>	Up to 20 in Up to 500 mm	(7 + 4.5L) $\mu$ in (0.2 + 0.005L) $\mu$ m	Comparison to Gage Blocks
Calipers <sup>1,3,5</sup>	Up to 80 in Up to 2 000 mm	(24 + 4.6L) $\mu$ in (0.6 + 0.005L) $\mu$ m	Comparison to Gage Blocks
LVDT's, Indicators <sup>1,3,5</sup>	Up to 12 in Up to 300 mm	(3 + 3.3L) $\mu$ in (0.08 + 0.005L) $\mu$ m	Comparison to Gage Blocks, Surface Plates
Measuring Tapes, Rulers <sup>1</sup>	Up to 5 ft (5 to 10) ft (10 to 48) ft (48 to 1 000) ft	0.0063 in 0.013 in 140 $\mu$ m/in 120 $\mu$ m/in	Comparison to Ruler Calibrator
Angle Blocks <sup>3</sup>	(0.25 to 90) $^{\circ}$	(13 + 0.07A)"	Comparison to Sine Bar, Gage Blocks, Surface Plate

## Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Angle Measuring Devices <sup>1, 3</sup> (Protractors, Inclinometers, Squares, Angle Gages, etc.)	0.005 6" to 5° (5 to 20)° (20 to 35)° (35 to 45)° (45 to 60)° (60 to 75)° (75 to 85)°	3.3" 6.5" 12" 17" 28" 59" 190"	Comparison to 5 in Sine Bar, Gage Blocks
Chamfer Gage <sup>1</sup>	(0.15 to 2) in (3.8 to 50) mm	290 $\mu$ in 7.5 $\mu$ m	Comparison to Reference Ring Gages
Cylindrical Square Squareness	Up to 24 in Up to 600 mm	25 $\mu$ in 0.6 $\mu$ m	Comparison to Indicator, Height Stand, Surface Plate
Electronic Levels, Inclination Levels <sup>3</sup>	Up to 1 000"	0.23"	Comparison to Sine Bar, Gage Blocks, Surface Plate
Spirit Levels, Clinometers <sup>5</sup>	Up to 4 ft Up to 1.2 m	50 $\mu$ in/ft 1.3 $\mu$ m/m	Sine Bar, Gage Blocks, Surface Plate
Sine Plates/Bars <sup>3</sup> Parallelism	Up to 15 in	78 $\mu$ in	Comparison to Surface Plate, Indicator with Gage Amplifier, Gage Blocks, Angle Blocks
Angle	(0 to 45)°	(10 + 0.6A)"	
Optical Flats/Parallels	Up to 4 in Up to 100 mm	6.4 $\mu$ in 0.16 $\mu$ m	Comparison to Master Optical Flat
Cylindrical Pins/Plugs <sup>3</sup>	(0.003 to 40) in	(8 + 0.8D) $\mu$ in	Comparison to Gage Blocks, Universal Length Measuring Machine
Cylindrical Rings/Bores <sup>3</sup>	(0.02 to 33) in	(4.8 + 0.9D) $\mu$ in	Comparison to Gage Blocks, Universal Length Measuring Machine

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
CMM Spheres, Gage Balls	(0.015 to 4) in (0.38 to 100) mm	12 $\mu$ in 0.3 $\mu$ m	Comparison to Gage Blocks, Universal Length Measuring Machine
Thread Wires	(0.003 to 0.825) in 76 $\mu$ m to 30 mm	10 $\mu$ in 0.25 $\mu$ m	Comparison to Master Thread Wires, Gage Blocks, Universal Length Measuring Machine
Thread Pitch Gages	Up to 6 in	430 $\mu$ in	Comparison to Optical Comparator
Thread Plug Gages <sup>3</sup> Pitch Diameter, (40 to 80) TPI	Up to 1.25 in (1.25 to 4.5) in (4.5 to 7) in	97 $\mu$ in 100 $\mu$ in 110 $\mu$ in	Comparisons to Universal Length Measuring Machine, Master Thread Wires, Gage Blocks
Major Diameter	Up to 1.75 in (1.75 to 7) in	59 $\mu$ in (56 + 2D) $\mu$ in	Universal Length Measuring Machine, Gage Blocks
Step Height	Up to 1 in	160 $\mu$ in	Gage Blocks, Test Stand
Thread Ring Gages <sup>3</sup> Pitch Diameter	Up to 20 in	(66 + 0.26D) $\mu$ in	Comparison to Universal Length Measuring Machine,
Minor Diameter	Up to 20 in	45 $\mu$ in	Thread Balls, Gage Blocks
Thread Ring Gages <sup>1,3</sup> Inner Pitch Diameter	Up to 2 in (2 to 12) in	(140 + 27.5D) $\mu$ in (160 + 18D) $\mu$ in	Comparison to Master Thread Plug Gages
Adjustable Thread Rings <sup>3</sup> Pitch Diameter	Up to 2 in	(140 + 2D) $\mu$ in	Tactile Fit to Class X Master Set Plugs
Parallels, Straight Edges	Up to 12 in	40 $\mu$ in	Comparison to Indicator, Height Stand, Surface Plate
PlaneKator	Up to 48 in	57 $\mu$ in	Comparison to Laser System, Surface Plate

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Profilometers, Surface Testers <sup>1</sup> (Primary)	$R_a = 120 \mu\text{in}$	$8.4 \mu\text{in}$	Comparison to Master Surface Patch
Profilometers, Surface Testers <sup>1</sup> (Secondary)	$R_a = (10 \text{ to } 200) \mu\text{in}$	$8 \% \text{ of reading} + 2.6 \mu\text{in}$	Comparison to Surface Patches
Roughness Standards/Patches (ISO Type C)	$R_a = (10 \text{ to } 200) \mu\text{in}$	$7 \% \text{ of reading} + 1.5 \mu\text{in}$	Comparison to Master Surface Patch, Profilometer
Coating Thickness Meters <sup>1</sup>	(0.94 to 20.24) mils	0.095 mils	Comparison to Coating Thickness Standards
V-Blocks			
Groove	Up to 4 in	64 $\mu\text{in}$	Comparison to Indicator, Height Stand, Surface Plate
Squareness	Up to 24 x 24 x 24 in	45 $\mu\text{in}$	
Bench Micrometers, Universal Length Measuring Machines <sup>1,3,5</sup> Standard Length	( $> 0.005$ to 1) in 2 in 3 in 4 in 5 in 6 in 7 in 8 in 10 in 12 in 16 in 20 in	( $2.5 + 0.5L$ ) $\mu\text{in}$ 4.2 $\mu\text{in}$ 4.9 $\mu\text{in}$ 6.4 $\mu\text{in}$ 8.1 $\mu\text{in}$ 8.9 $\mu\text{in}$ 9.7 $\mu\text{in}$ 11 $\mu\text{in}$ 12 $\mu\text{in}$ 14 $\mu\text{in}$ 17 $\mu\text{in}$ 20 $\mu\text{in}$	Comparison to Gage Blocks, Long Gage Blocks
Non-standard Length	(20 to 40) in	( $20 + 0.86L$ ) $\mu\text{in}$	
Coordinate Measuring Machines (CMM's) <sup>1,3</sup> Linear Displacement Accuracy (X, Y, Z)	Up to 72 in	( $8 + 3L$ ) $\mu\text{in}$	Comparisons to Gage Blocks, Optical Flats
Volumetric Repeatability	Up to 72 in	100 $\mu\text{in}$	Ball Bar
Probe Repeatability	Up to 72 in	25 $\mu\text{in}$	Sphere

This Scope of Accreditation, version 008, was last updated on: 10 December 2025 and is valid only when accompanied by the Certificate.

Page 26 of 35

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Comparators <sup>1,3</sup> X,Y Measuring Stage Travel	Up to 12 in	$(130 + 17L) \mu\text{in}$	Comparisons to Calibration Grids
	Squareness	$120 \mu\text{in}$	Calibration Grids
	Magnification	$(240 + 21L) \mu\text{in}$	Magnification Checker
Radius Gages	Up to 2 in	$460 \mu\text{in}$	Comparison to Optical Comparator
Granite Surface Plates <sup>1,3</sup> Overall Flatness	$(8.49 \text{ to } 299.25) DL$	$4 \sqrt{DL} \mu\text{in}$	In accordance with ASME B89.3.7 using Electronic Level System
	Up to 0.005 in	$40 \mu\text{in}$	Repeat-o-Meter
Torque Wheels, Torque Arms <sup>3</sup>	Up to 40 in	$(470 + 2L) \mu\text{in}$	Comparison to Surface Plate, Indicator with Gage Amplifier, Gage Blocks, Digital Outside Micrometer
Depth Micrometers <sup>1,3</sup>	$(0.01 \text{ to } 72) \text{ in}$	$(50 + 55.1L) \mu\text{in}$	Comparison to Gage Blocks

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air Velocity Measuring Devices (Anemometers, Thermal, Pitot, Vane-style and Similar Equipment)	$(30 \text{ to } 250) \text{ ft/min}$ $(250 \text{ to } 1\,500) \text{ ft/min}$ $(1\,500 \text{ to } 9\,000) \text{ ft/min}$	2.3 % of reading + 2.6 ft/min 2.3 % of reading + 6 ft/min 1.2 % of reading + 17 ft/min	Comparison to Wind Tunnel
Gas Mass Flow <sup>1</sup> (Flow Meters, Flow Controllers, Rotameters, and Leak Orifices)	$(1 \text{ to } 10) \text{ sccm}$ $(10 \text{ to } 1\,000) \text{ sccm}$	0.25 sccm 0.22 % of reading	Comparison to Fluke molBloc-L Calibration System
	$(1 \text{ to } 1\,000) \text{ slpm}$	0.1 % of reading	Comparison to Fluke molBloc-S Calibration System

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Liquid Flow Meters <sup>1</sup>	Up to 10 gpm	1.4 % of reading + 0.001 8 gpm	Calibration by Time and Mass
Hydrometers	(0.6 to 0.64) SG (0.64 to 1.67) SG (1.67 to 2) SG	0.002 5 SG 0.001 4 SG 0.005 SG	Comparison to Reference Hydrometer per ASTM E126.
Kinematic Viscosity Meters	< 10 mm <sup>2</sup> /s (11 to 100) mm <sup>2</sup> /s (101 to 1000) mm <sup>2</sup> /s (1001 to 10 000) mm <sup>2</sup> /s (10 001 to 100 000) mm <sup>2</sup> /s	0.25 % of reading 0.32 % of reading 0.37 % of reading 0.44 % of reading 0.51 % of reading	Comparison to Accredited Viscosity Standard, Temperature Indicator with Probe
Viscosity Cups Kinematic Viscosity @ 25 °C			Accredited Viscosity Standard per ASTM D4212
Zahn	(5 to 1 800) mm <sup>2</sup> /s	2.2 % of reading	
Shell	(2 to 1 300) mm <sup>2</sup> /s	2.2 % of reading	ASTM D4212
Ford	(2 to 1 400) mm <sup>2</sup> /s	2.2 % of reading	ASTM D1200
Piston Operated Volumetric Apparatus <sup>7</sup> (Pipettes, Syringes, Burettes, Liquid, Handlers, Dispensers)	(1 to 2 000) µL (2000 to 10 000) µL (10 000 to 100 000) µL	94 nL 0.22 µL 2 µL	Gravimetric Method using Electronic Balances and ASTM E617 Class 1 Weights.
Volume Measuring Equipment <sup>1</sup>	(0.001 to 4 000) mL	0.8 µL	Per ASTM E542 using Temperature Indicator with Probe, Electronic Balance, Distilled Water, Barometer
Load Cells, Force Testers, Force Gages <sup>1</sup> (Tension and Compression)	(25 to 250) cN (250 to 2 000) cN	5.2 cN 52 cN	Comparison to Correx Gram Gauge
	(0 to 500) lbf	0.088 % of reading	Comparison to NIST Class F Weights
	(100 to 5 000) lbf (5 000 to 10 000) lbf (10 000 to 25 000) lbf (25 000 to 50 000) lbf	0.008 % of reading + 1.7 lbf 0.013 % of reading + 1.7 lbf 0.013 % of reading + 4.2 lbf 0.013 % of reading + 8.3 lbf	Comparison to Tovey Engineering Calibration System

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Durometers <sup>1</sup> (Type A, B, C, CF, D, DO, E, M, O, OO, OOO-S, SA, ASKER C) Indenter Dimensions Extension Length Diameter Angle Radius Spring Force	Up to 0.2 in Up to 1 in Up to 35° Up to 1 in (0 to 100) Duro	160 $\mu$ in 170 $\mu$ in 0.1° 170 $\mu$ in 0.34 Duro	Direct Verification per ASTM D2240 using Optical Comparator Durometer Calibrator
Shore Durometer Calibrators Dimensional Measurements Mass	Up to 8 in Up to 4 kg	480 $\mu$ in 30 mg	Comparisons to Optical Comparator Single Substitution Method using Electronic Balance
Rockwell Hardness and Superficial Testers <sup>1</sup>	HRBW (80 to 100) HRBW (60 to 79) HRBW (40 to 59) HRBW HRC (60 to 65) HRC (35 to 55) HRC (20 to 30) HRC	0.48 HRBW 0.77 HRBW 1.1 HRBW 0.37 HRC 0.48 HRC 0.57 HRC	Indirect verification per ASTM E18 using hardness test blocks.
Mass Determination <sup>1</sup> (Variable)	1 mg to 5 g (5 to 30) g (30 to 200) g (200 to 500) g (500 to 2000) g (2 to 10) kg (10 to 30) kg (30 to 45) kg	21 $\mu$ g 24 $\mu$ g 0.21 mg 2.1 mg 21 mg 0.21 g 0.24 g 0.45 g	Comparison to Electronic Balances, OIML Class E1 Weights, ASTM E617 Class 1 Weights

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination (Fixed Points)	1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg	4.2 µg 4.2 µg 4.2 µg 4.2 µg 4.2 µg 4.3 µg 4.5 µg 4.5 µg 4.5 µg 8.6 µg 9.1 µg 11 µg 15 µg 19 µg 28 µg 18 µg 0.13 mg 0.91 mg 0.86 mg	Comparison to Electronic Balances, OIML Class E1 Weights
Pneumatic Pressure Measuring Equipment <sup>1</sup> (Absolute, Gauge)	(0.1 to 15) psi	0.003 4 % of reading + 0.000 5 psi	Comparison to Ruska 7010 Pressure Controller/Calibrator
	(0.1 to 1 000) psi	0.006 % of reading + 0.015 psi	Comparison to Ruska 7215I Pressure Controller/Calibrator
Pressure Measuring Equipment <sup>1</sup> (Absolute, Gauge, Negative and Positive)	(-104 to -7.5) kPa (-7.5 to -2.9) kPa (-2.9 to -0.75) kPa (-0.75 to 0.75) kPa (0.75 to 2.9) kPa	0.007 7 % of reading + 5.2 Pa 0.009 1 % of reading 0.009 % of reading + 35 mPa 0.007 5 % of reading + 51 mPa 0.009 % of reading + 35 mPa	Comparison to Fluke PPC4 Pressure Controller/Calibrator
Pressure Measuring Equipment <sup>1</sup> (Absolute, Gauge, Negative and Positive)	(2.9 to 7.5) kPa (7.5 to 26) kPa (26 to 63) kPa (63 to 700) kPa (700 to 7 000) kPa	0.009 1 % of reading 0.007 7 % of reading + 1.3 Pa 0.009 % of reading + 50 mPa 0.008 % of reading + 0.7 Pa 0.008 % of reading + 0.3 Pa	Comparison to Fluke PPC4 Pressure Controller/Calibrator
Pressure Measuring Equipment (Absolute, Gauge)	(7 000 to 82 737) kPa	0.005 % of reading	Comparison to Ruska 2400HL Hydraulic Piston Gauge

This Scope of Accreditation, version 008, was last updated on: 10 December 2025 and is valid only when accompanied by the Certificate.

Page 30 of 35

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales and Balances <sup>1,8</sup> Metric (SI)	Up to 500 mg (0.5 to 5) g (5 to 10) g (10 to 20) g (20 to 100) g (100 to 200) g (0.2 to 2) kg (2 to 20) kg	22 µg 29 µg 39 µg 54 µg 0.28 mg 0.43 mg 0.002 2% of reading 0.004 2% of reading	ASTM E617 Class 1 weights and NIST HB44 utilized in the calibration of the weighing system.
Scales and Balances <sup>1,8</sup> Avoirdupois	Up to 0.5 lb (0.5 to 1) lb (1 to 2 000) lb	0.026 % of reading 0.019 % of reading 0.013 % of reading	NIST Class F weights and NIST HB44 utilized in the calibration of the weighing system.
Scales and Balances <sup>1,8</sup> Metric (SI)	Up to 250 g (250 to 500) g (0.5 to 908) kg	0.026 % of reading 0.019 % of reading 0.013 % of reading	NIST Class F weights and NIST HB44 utilized in the calibration of the weighing system.
Torque Wrenches, Torque Watches <sup>1</sup>	(1 to 1 000) lbf·in (1 to 1 000) lbf·ft	0.52 % of reading 0.52 % of reading	Comparison to Torque Transducers, Torque Indicators, Manual Loader
Torque Transducers, Torque Meters <sup>1</sup>	(1 to 1 000) lbf·in (1 to 1 000) lbf·ft	0.09 % of reading 0.08 % of reading	Comparison to Torque Wheels, Torque Arms, NIST Class F Weights

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dew Point Measuring Equipment	(-25 to 68) °C	0.15 °C	Comparison to Thunder Scientific 2500 Two-pressure Humidity Generator
Dew Point – Measure <sup>1</sup>	(-15 to 20) °C	0.26 °C	Comparison to Chilled Mirror

## Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
Relative Humidity – Source	(-10 to 15) °C (10 to 75) % RH (75 to 95) %RH (15 to 35) °C (10 to 95) %RH (35 to 70) °C (10 to 50) %RH (50 to 75) %RH (75 to 95) %RH	0.5 %RH 0.65 %RH 0.5 %RH 0.5 %RH 0.7 %RH 0.85 %RH	Comparison to Thunder Scientific 2500 Two-Pressure Humidity Generator	
Relative Humidity – Measure <sup>1</sup>	(10 to 95) %RH	1.3 %RH	Direct measure using Chilled Mirror	
Temperature Measuring Equipment	0.01 °C	1.5 mK	Comparison to Triple Point of Water	
Temperature Measuring Equipment by Comparison <sup>2,3</sup>	NBPLN <sub>2</sub> Hg In Sn Zn Al	-195 °C -38.8 °C 156.6 °C 231.9 °C 419.5 °C 660.3 °C	4 mK 5 mK 5 mK 5 mK 4 mK 10 mK	Comparison to Fluke 1595A Super Thermometer, Fluke 5628 Secondary PRT
Temperature – Measure <sup>1</sup> (PRT, RTD, Thermistor, Thermocouple, Stirred Baths, Liquid Baths)	(-95 to 700) °C	10 mK	Comparison to SPRT and Indicator	
Infrared Thermometers <sup>1</sup>	(-15 to 0) °C (0 to 50) °C (50 to 100) °C (100 to 120) °C (120 to 200) °C (200 to 350) °C (350 to 500) °C	0.83 °C 0.66 °C 0.67 °C 0.72 °C 0.97 °C 1.6 °C 2.2 °C	Comparison to Blackbody Source (Flat Plate) $\lambda = (8 \text{ to } 14) \mu\text{m}$ , $\epsilon = (0.9 \text{ to } 1)$	

## Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Reference	10 MHz	3.7 pHz/Hz	Comparison to Fluke 910R GPS Frequency Standard

This Scope of Accreditation, version 008, was last updated on: 10 December 2025 and is valid only when accompanied by the Certificate.

Page 32 of 35

**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source <sup>1</sup>	1 µHz to 80 MHz	0.58 µHz/Hz	Comparison to Agilent 33250A Arbitrary Waveform Generator
Frequency – Source <sup>1</sup>	250 kHz to 3 GHz	0.58 µHz/Hz	Comparison to Agilent E4432B RF Signal Generator
Frequency – Measure <sup>1</sup>	1 Hz to 12.4 GHz	51 pHz/Hz	Comparison to Agilent 53132A Universal Frequency Counter
AC Duty Cycle – Source <sup>1,10</sup> Square Wave Freq: 10 mHz to 100 kHz	(1 to 10) % Duty Cycle 10 µs to 100 s (10 to 49) % Duty Cycle 10 µs to 100 s 50 % Duty Cycle 10 µs to 100 s (51 to 90) % Duty Cycle 10 µs to 100 s (90 to 99) % Duty Cycle 10 µs to 100 s	0.62 % of period + 78 ns 0.039 % of period + 78 ns 0.016 % of period + 78 ns 0.039 % of period + 78 ns 0.62 % of period + 78 ns	Comparison to Fluke 55xxA Multiproduct Calibrator
Stopwatches, Timers <sup>1</sup>	Up to 19.99 s/d	58 ms/d	Comparison to Helmut Klein TM-4500 Timometer
Tachometers – Optical Pickup <sup>1,3</sup>	Up to 100 000 rpm	0.000 23 % of reading + 0.001 1 rpm	Comparison to Agilent 33250A Arbitrary Waveform Generator, LED
Tachometers – Contact <sup>1,3</sup>	(1 to 10) rpm (10 to 1 000) rpm (1 000 to 8 000) rpm (8 000 to 20 000) rpm	0.012 % of reading + 0.085 rpm 0.012 % of reading + 0.12 rpm 0.012 % of reading + 0.53 rpm 0.012 % of reading + 5 rpm	Comparison to Tachometer Calibrator, Master Optical Tachometer
Rotation Speed – Measure <sup>1,3</sup> (Conveyor Belts, Line Speed, Centrifuges, and Mechanical Tachometers)	(6 to 8 300) rpm (8 300 to 99 999) rpm	1.7 rpm 2.6 rpm	Direct measure using Optical Tachometer
	(6 to 8 300) rpm (8 300 to 99 999) rpm	2.4 rpm 3 rpm	Direct measure using Mechanical Tachometer

## DIMENSIONAL MEASUREMENT

### 1 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 1D <sup>1</sup>	Up to 12 in	1 500 $\mu$ in	Digital Caliper utilized as the reference standard for 1-D Length Measurements.
Dimensional Measurement – 1D <sup>1</sup>	Up to 1 in	150 $\mu$ in	Digital Outside Micrometer utilized as the reference standard for 1-D Length Measurements.
Dimensional Measurement – 1D <sup>1</sup>	Up to 6 in	150 $\mu$ in	Digital Depth Micrometer utilized as the reference standard for 1-D Length Measurements.
Dimensional Measurement – 1D <sup>1</sup>	Up to 1 in	600 $\mu$ in	Gage Pins utilized as the reference standard for 1-D Length Measurements.

### 2 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 2D	Up to 6 in	(420 + 0.52L) $\mu$ in	Optical Comparator utilized as the reference standard for 2-D Length Measurements.
Dimensional Measurement – 2D	Up to 30 in	200 $\mu$ in	CMM utilized as the reference standard for 2-D Length Measurements.
Surface Finish <sup>1</sup>	$R_a = (10 \text{ to } 200) \mu\text{in}$	7 % of reading + 1.5 $\mu$ in	Profilmeter utilized as a reference standard for Dimensional Inspection.


### 3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurements – 3D <sup>3</sup> (Volumetric)	X-axis: Up to 18 in Y-axis: Up to 20 in Z-axis: Up to 16 in	(300 + 43L) $\mu$ in (300 + 44L) $\mu$ in (300 + 42L) $\mu$ in	CMM utilized as the reference standard for 3-D Length Measurements.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. The values represented here are nominal values. The certified values and associated uncertainty will be reported at the time of calibration.
3.  $L$  = length in inches or millimeters; rpm = revolutions per minute; PF = Power Factor;  $A$  = Angle in degrees ( $^{\circ}$ ); " = arcsecond;  $D$  = diameter in inches or millimeters;  $DL$  = diagonal length in inches; NBPLN<sub>2</sub> = Boiling Point of Liquid Nitrogen.
4. The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact the laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
5. At the time of Calibration,  $0.6R$  will be added to the Measurement Uncertainty, where  $R$  = resolution of the unit under calibration.
6.  $\Delta P$  measurement with density correction for standard condition normalization.
7. The contributions from "the best existing device" are not included in the CMC presented claim.
8. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
9. Volume calculations are based on independent linear measurements.
10. The CMC for AC Duty Cycle – Source is defined relative to the period of the sourced signal. The reported percent duty cycle measurement uncertainty will be determined and applied at the time of calibration.
11. Unless otherwise specified in the far-right column, the calibration procedure/method utilized by the laboratory was developed and validated internally.
12. The legal entity name for this client is Transcat, Inc.



Jason Stine, Vice President