



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat - Cleveland
9325 Progress Parkway, Suite B
Mentor, OH 44060

Fulfills the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 07 September 2025

Certificate Number: AC-1287



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Transcat - Cleveland

9325 Progress Parkway, Suite B

Mentor, OH 44060

Kyle Pillar 440-853-1900

Email: kyle.pillar@transcat.com

CALIBRATION

Valid to: September 7, 2025

Certificate Number: AC-1287

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Measure ^{1,3}	10 μ V to 100 mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 300) V (300 to 500) V (500 to 700) V (700 to 1 000) V	1.5 μ V 13 μ V 45 μ V 1.1 mV 8.7 mV 13 mV 18 mV 24 mV	HP 3458A Opt 002 8.5 Digit Multimeter
DC Voltage – Measure ^{1,3}	(1 to 10) kV	1.7 V/kV + 0.1 V	Ross VD15 Voltage Divider, HP 34401A 6.5 Digit Multimeter
	(10 to 60) kV	1.6 V/kV + 0.2 V	Ross VD60 Voltage Divider, HP34401A 6.5 Digit Multimeter
	(60 to 141) kV	1.8 V/kV + 8.8 V	Ross VMP200 Voltage Divider, Fluke 187 Digital Multimeter
DC Voltage – Source ^{1,3}	10 μ V to 330 mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 020) V	0.001 6 % of reading + 0.8 μ V 0.000 9 % of reading + 2 μ V 0.001 % of reading + 15 μ V 0.001 4 % of reading + 0.12 mV 0.001 5 % of reading + 1 mV	Fluke 5522A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure ^{1,3}	(0.1 to 100) μ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	4.3 nA 39 nA 0.39 μ A 5.6 μ A 0.15 mA	HP 3458A Opt 002 8.5 Digit Multimeter
DC Current – Measure ^{1,3}	(1 to 14) A (14 to 30) A	1.5 mA 2.1 mA	HP 3458A Opt 002 8.5 Digit Multimeter, IET DCCS-0.01 and DCCS-0.001 Shunts
DC Current – Source ^{1,3}	(0.1 to 330) μ A 330 μ A to 3.3 mA (3.3 to 33) mA (33 to 330) mA 330 mA to 1.1 A (1.1 to 3) A (3 to 11) A (11 to 20.5) A	0.012 % of reading + 16 nA 0.007 9 % of reading + 38 nA 0.007 9 % of reading + 0.2 μ A 0.007 9 % of reading + 2 μ A 0.017 % of reading + 28 μ A 0.03 % of reading + 37 μ A 0.04 % of reading + 0.4 mA 0.08 % of reading – 11 μ A	Fluke 5522A Multiproduct Calibrator
DC Current Clamp Meters ¹ Non-toroidal	(10 to < 16.5) A (16.5 to < 150) A (150 to 1 025) A	0.58 % of reading + 23 mA 0.58 % of reading + 0.16 A 0.58 % of reading + 0.58 A	Fluke 5522A Multiproduct Calibrator, Fluke 5500A/COIL 50-turn Current Coil
AC Voltage – Measure ^{1,3}	100 μ V to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 8) MHz (8 to 10) MHz	9.8 μ V 6.6 μ V 7.7 μ V 17 μ V 62 μ V 0.47 mV 0.16 mV 0.17 mV 15 μ V 13 μ V 21 μ V 40 μ V 97 μ V 0.37 mV 1.2 mV 1.9 mV 4.8 mV 18 mV	HP 3458A Opt 002 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ^{1,3}	(0.1 to 1) V		HP 3458A Opt 002 8.5 Digit Multimeter
	(1 to 40) Hz	0.15 mV	
	40 Hz to 1 kHz	0.12 mV	
	(1 to 20) kHz	0.21 mV	
	(20 to 50) kHz	0.39 mV	
	(50 to 100) kHz	0.97 mV	
	(100 to 300) kHz	3.7 mV	
	300 kHz to 1 MHz	12 mV	
	(1 to 2) MHz	19 mV	
	(2 to 8) MHz	48 mV	
	(8 to 10) MHz	0.18 V	
	(1 to 10) V		
	(1 to 40) Hz	2.5 mV	
	40 Hz to 1 kHz	1.2 mV	
	(1 to 20) kHz	2.1 mV	
	(20 to 50) kHz	3.9 mV	
	(50 to 100) kHz	9.7 mV	
	(100 to 300) kHz	37 mV	
	300 kHz to 1 MHz	0.12 V	
	(1 to 2) MHz	0.18 V	
	(2 to 8) MHz	0.48 V	
	(8 to 10) MHz	1.8 V	
AC Voltage – Measure ^{1,3}	(10 to 100) V		Ross VD15 Voltage Divider, HP 34401A 6.5 Digit Multimeter
	40 Hz to 20 kHz	28 mV	
	(20 to 50) kHz	45 mV	
	(50 to 100) kHz	0.15 V	
	(100 to 1 000) V		
	40 Hz to 1 kHz	0.046 % of reading + 40 mV	
	(1 to 20) kHz	0.073 % of reading + 0.8 mV	
	(1 to 10) kV 60 Hz	7.2 V/kV + 2.9 V	
	(10 to 42) kV 60Hz	6.2 V/kV + 7.9 V	Ross VMP200 Voltage Divider, Fluke 187 Digital Multimeter
	(42 to 106) kV 60 Hz	16 V/kV + 6 V	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ^{1,3}	100 μ V to 33 mV (10 to 45) Hz	0.064 % of reading + 5 μ V	Fluke 5522A Multiproduct Calibrator
	45 Hz to 10 kHz (10 to 20) kHz	0.012 % of reading + 5 μ V	
	(20 to 50) kHz	0.015 % of reading + 5 μ V	
	(50 to 100) kHz	0.08 % of reading + 5 μ V	
	(100 to 500) kHz	0.27 % of reading + 10 μ V	
	(33 to 330) mV (10 to 45) Hz	0.64 % of reading + 39 μ V	
	45 Hz to 10 kHz (10 to 20) kHz	0.023 % of reading + 8 μ V	
	(20 to 50) kHz	0.012 % of reading + 7 μ V	
	(50 to 100) kHz	0.013 % of reading + 8 μ V	
	(100 to 500) kHz	0.027 % of reading + 7 μ V	
	330 mV to 3.3 V (10 to 45) Hz	0.062 % of reading + 25 μ V	
	45 Hz to 10 kHz (10 to 20) kHz	0.16 % of reading + 58 μ V	
	(20 to 50) kHz	0.024 % of reading + 41 μ V	
	(50 to 100) kHz	0.012 % of reading + 48 μ V	
	(100 to 500) kHz	0.015 % of reading + 46 μ V	
	(3.3 to 33) V (10 to 45) Hz	0.024 % of reading + 41 μ V	
	45 Hz to 10 kHz (10 to 20) kHz	0.055 % of reading + 0.11 mV	
	(20 to 50) kHz	0.19 % of reading + 0.47 mV	
	(50 to 100) kHz	0.024 % of reading + 0.51 mV	
	(33 to 330) V (10 to 20) kHz	0.012 % of reading + 0.48 mV	
	(20 to 50) kHz	0.02 % of reading + 0.47 mV	
	(50 to 100) kHz	0.028 % of reading + 0.5 mV	
	45 Hz to 1 kHz (1 to 10) kHz	0.072 % of reading + 1.3 mV	
	(10 to 20) kHz	0.015 % of reading + 1.5 mV	
	(20 to 50) kHz	0.016 % of reading + 4.8 mV	
	(50 to 100) kHz	0.02 % of reading + 5.5 mV	
	330 V to 1.02 kV (1 to 5) kHz	0.024 % of reading + 5.1 mV	
	(5 to 10) kHz	0.16 % of reading + 40 mV	
		0.024 % of reading + 9 mV	
		0.021 % of reading + 4 mV	
		0.024 % of reading + 9 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ^{1,3}	(29 to 330) μ A		Fluke 5522A Multiproduct Calibrator
	(10 to 20) Hz	0.16 % of reading + 90 nA	
	(20 to 45) Hz	0.12 % of reading + 90 nA	
	45 Hz to 1 kHz	0.1 % of reading + 90 nA	
	(1 to 5) kHz	0.24 % of reading + 0.12 μ A	
	(5 to 10) kHz	0.62 % of reading + 0.17 μ A	
	(10 to 30) kHz	1.3 % of reading + 0.32 μ A	
	(0.33 to 3.3) mA		
	(10 to 20) Hz	0.16 % of reading + 0.12 μ A	
	(20 to 45) Hz	0.1 % of reading + 0.11 μ A	
	45 Hz to 1 kHz	0.08 % of reading + 0.11 μ A	
	(1 to 5) kHz	0.16 % of reading + 0.2 μ A	
	(5 to 10) kHz	0.39 % of reading + 0.3 μ A	
	(10 to 30) kHz	0.78 % of reading + 0.5 μ A	
	(3.3 to 33) mA		
	(10 to 20) Hz	0.14 % of reading + 1.5 μ A	
	(20 to 45) Hz	0.07 % of reading + 1.6 μ A	
	45 Hz to 1 kHz	0.032 % of reading + 1.6 μ A	
	(1 to 5) kHz	0.062 % of reading + 1.6 μ A	
	(5 to 10) kHz	0.16 % of reading + 2.3 μ A	
	(10 to 30) kHz	0.33 % of reading + 3 μ A	
	(33 to 330) mA		
	(10 to 20) Hz	0.14 % of reading + 15 μ A	
	(20 to 45) Hz	0.071 % of reading + 16 μ A	
	45 Hz to 1 kHz	0.032 % of reading + 16 μ A	
	(1 to 5) kHz	0.08 % of reading + 40 μ A	
	(5 to 10) kHz	0.16 % of reading + 78 μ A	
	(10 to 30) kHz	0.32 % of reading + 0.16 mA	
	(0.33 to 1.1) A		
	(10 to 45) Hz	0.15 % of reading + 50 μ A	
	45 Hz to 1 kHz	0.04 % of reading + 76 μ A	
	(1 to 5) kHz	0.41 % of reading + 0.84 mA	
	(5 to 10) kHz	2 % of reading + 4.3 mA	
	(1.1 to 3) A		
	(10 to 45) Hz	0.14 % of reading + 0.14 mA	
	45 Hz to 1 kHz	0.048 % of reading + 67 μ A	
	(1 to 5) kHz	0.48 % of reading + 0.7 mA	
	(5 to 10) kHz	1.9 % of reading + 5.4 mA	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ^{1,3}	(3 to 11) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz (11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.05 % of reading + 1.6 mA 0.09 % of reading + 1.3 mA 2.4 % of reading + 1.1 mA 0.09 % of reading + 4 mA 0.12 % of reading + 4 mA 2.3 % of reading + 22 mA	Fluke 5522A Multiproduct Calibrator
AC Current Clamp Meters ¹ Non-toroidal	(10 to < 16.5) A (45 to 65) Hz (65 to 440) Hz (16.5 to < 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 025) A (45 to 65) Hz (65 to 440) Hz	0.65 % of reading + 35 mA 1.2 % of reading + 35 mA 0.84 % of reading + 3 mA 1.2 % of reading + 0.3 A 0.65 % of reading + 1 A 1.2 % of reading + 1.1 A	Fluke 5522A Multiproduct Calibrator, Fluke 5500A/COIL 50-turn Current Coil
Toroidal	(10 to < 16.5) A (45 to 65) Hz (65 to 440) Hz (16.5 to < 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 025) A (45 to 65) Hz (65 to 440) Hz	0.38 % of reading + 3.2 mA 0.93 % of reading + 3.8 mA 0.38 % of reading + 27 mA 0.94 % of reading + 32 mA 0.39 % of reading + 88 mA 0.94 % of reading + 0.11 A	
AC Current – Measure ^{1,3}	(10 to 100) μ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (0.1 to 1) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.5 μ A 0.21 μ A 0.11 μ A 4.9 μ A 1.8 μ A 0.74 μ A 0.41 μ A	HP 3458A Opt 002 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ^{1,3}	(1 to 10) mA		HP 3458A Opt 002 8.5 Digit Multimeter
	(10 to 20) Hz	49 μ A	
	(20 to 45) Hz	18 μ A	
	(45 to 100) Hz	7.4 μ A	
	100 Hz to 5 kHz	4.1 μ A	
	(10 to 100) mA		
	(10 to 20) Hz	0.49 mA	
	(20 to 45) Hz	0.18 mA	
	(45 to 100) Hz	73 μ A	
	100 Hz to 5 kHz	40 μ A	
	(0.1 to 1) A		
	(10 to 20) Hz	4.9 mA	
	(20 to 45) Hz	1.9 mA	
	(45 to 100) Hz	1 mA	
	100 Hz to 5 kHz	1.3 mA	
AC Current – Measure ^{1,3}	(1 to 3) A		HP 3458A Opt 002 8.5 Digit Multimeter, Precision Measurements 9810 Current Shunt
	60 Hz	0.8 mA	
	400 Hz	1.2 mA	
	1 kHz	1.4 mA	
	10 kHz	1.1 mA	
	20 kHz	1.4 mA	
	50 kHz	7.2 mA	
	100 kHz	8.2 mA	
	(3 to 10) A		
	60 Hz	5.5 mA	
	400 Hz	6.3 mA	
	1 kHz	6.7 mA	
	10 kHz	9.4 mA	
	20 kHz	9.7 mA	
	50 kHz	29 mA	
	100 kHz	48 mA	
	(10 to 30) A		
	60 Hz	7.5 mA	
	400 Hz	12 mA	
	1 kHz	14 mA	
	10 kHz	11 mA	
	20 kHz	14 mA	
	50 kHz	72 mA	
	100 kHz	82 mA	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ^{1,3}	(30 to 50) A 60 Hz 400 Hz 1 kHz 10 kHz 20 kHz 50 kHz 100 kHz	18 mA 19 mA 20 mA 28 mA 31 mA 0.13 A 0.17 A	HP 3458A Opt 002 8.5 Digit Multimeter, Precision Measurements 9830 Current Shunt
Inductance – Source ¹ (Fixed Artifacts)	1 kHz 50 μ H 1 mH 5 H	0.32 μ H 2 μ H 14 mH	GenRad Standard Inductors
Inductance – Measure ¹	1 kHz 100 μ H to 1 mH (1 to 10) mH (10 to 100) mH 100 mH to 1 H (1 to 10) H	0.2 nH 2.4 μ H 2.4 μ H 0.24 mH 2.4 mH	GenRad 1689 RLC Bridge
Capacitance – Measure ¹	1 kHz Up to 1 nF (1 to 10) nF (10 to 100) nF (0.1 to 1) μ F (1 to 1.111) μ F	0.2 pF 2.4 pF 24 pF 0.24 nF 0.29 nF	GenRad 1689 RLC Bridge
Capacitance – Source ^{1,3} (Simulation) 10 Hz to 10 kHz 10 Hz to 3 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz (10 to 600) Hz (10 to 300) Hz (10 to 150) Hz (10 to 120) Hz (10 to 80) Hz (10 to 50) Hz (10 to 20) Hz (0 to 6) Hz (0 to 2) Hz (0 to 0.6) Hz (0 to 0.2) Hz	220 pF to 1.1 nF (1.1 to 3.3) nF (3.3 to 11) nF (11 to 110) nF (110 to 330) nF 330 nF to 1.1 μ F (1.1 to 3.3) μ F (3.3 to 11) μ F (11 to 33) μ F (33 to 110) μ F (110 to 330) μ F 330 μ F to 1.1 mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	0.41 % of reading + 8 pF 0.41 % of reading + 8 pF 0.19 % of reading + 9 pF 0.18 % of reading + 88 pF 0.2 % of reading + 0.24 nF 0.19 % of reading + 0.9 nF 0.2 % of reading + 2.4 nF 0.19 % of reading + 9 nF 0.33 % of reading + 22 nF 0.36 % of reading + 80 nF 0.36 % of reading + 0.24 μ F 0.36 % of reading + 0.8 μ F 0.36 % of reading + 2.4 μ F 0.36 % of reading + 8 μ F 0.61 % of reading + 21 μ F 0.97 % of reading + 44 μ F	Fluke 5522A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance – Source ^{1,3}	(0.1 to 11) Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) k Ω (1.1 to 3.3) k Ω (3.3 to 11) k Ω (11 to 33) k Ω (33 to 110) k Ω (110 to 330) k Ω (0.33 to 1.1) M Ω (1.1 to 3.3) M Ω (3.3 to 11) M Ω (11 to 33) M Ω (33 to 110) M Ω (110 to 330) M Ω (0.33 to 1.1) G Ω	0.003 9 % of reading + 0.8 m Ω 0.002 4 % of reading + 1.3 m Ω 0.002 1 % of reading + 1.2 m Ω 0.002 2 % of reading + 1.6 m Ω 0.002 2 % of reading + 1.4 m Ω 0.002 2 % of reading + 16 m Ω 0.001 6 % of reading + 37 m Ω 0.002 2 % of reading + 0.16 Ω 0.002 3 % of reading + 0.14 Ω 0.002 5 % of reading + 1.7 Ω 0.002 5 % of reading + 1.5 Ω 0.004 8 % of reading + 22 Ω 0.011 % of reading + 28 Ω 0.02 % of reading + 2 k Ω 0.038 % of reading + 3.1 k Ω 0.24 % of reading + 80 k Ω 1.3 % of reading + 0.11 M Ω	Fluke 5522A Multiproduct Calibrator
DC Resistance – Measure ^{1,3}	Up to 10 Ω (10 to 100) Ω 100 Ω to 1 k Ω (1 to 10) k Ω (10 to 100) k Ω 100 k Ω to 1 M Ω (1 to 10) M Ω (10 to 100) M Ω	0.36 m Ω 2.7 m Ω 19 m Ω 0.19 Ω 2 Ω 30 Ω 0.85 k Ω 64 k Ω	HP 3458A Opt 002 8.5 Digit Multimeter
AC Resistance – Measure ^{1,3}	60 Hz 8 m Ω 23 m Ω	1.7 $\mu\Omega$ 4.2 $\mu\Omega$	Precision Measurements 9810/9830 Shunts
Oscilloscopes ^{1,2,3} Amplitude – DC into 50 Ω load into 1 M Ω load Amplitude – Square Wave into 50 Ω load into 1 M Ω load	(-6.6 to 6.6) V (-130 to 130) V 1 mVp-p to 6.6 Vp-p 10 Hz to 10 kHz 1 mVp-p to 130 Vp-p 10 Hz to 10 kHz	0.65 % of reading + 40 μ V 0.6 % of reading + 40 μ V 0.65 % of reading + 40 μ V 0.61 % of reading + 40 μ V	Fluke 5522A/SC1100 Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,2,3} Leveled Sine Wave Amplitude (relative to 50 kHz)	5 mVp-p to 5.5 Vp-p 50 kHz 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	2 % of reading + 0.3 mV 3.5 % of reading + 0.3 mV 4 % of reading + 0.3 mV 6 % of reading + 0.3 mV 7 % of reading + 0.3 mV	Fluke 5522A/SC1100 Multiproduct Calibrator
Flatness	50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 600 MHz to 1.1 GHz	1.6 % of reading + 0.1 mV 2.1 % of reading + 0.1 mV 4 % of reading + 0.1 mV 5 % of reading + 0.1 mV	
Time Marker into 50 Ω load	2 ns to 20 ms 50 ms 0.1 s 0.2 s 0.5 ms 1 s 2 s 5 s	0.000 21 % of reading 0.006 % of reading 0.009 8 % of reading 0.018 % of reading 0.042 % of reading 0.08 % of reading 0.16 % of reading 0.4 % of reading	
Rise Time	1 kHz to 10 MHz (200 to 350) ps	61 ps	
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ^{1,3}	Type B (600 to 800) °C (800 to 1 000) °C (1 000 to 1 550) °C (1 550 to 1 820) °C Type C (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 316) °C	0.35 °C 0.27 °C 0.24 °C 0.26 °C 0.24 °C 0.21 °C 0.24 °C 0.39 °C 0.66 °C	Fluke 5522A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ^{1,3}	Type E		Fluke 5522A Multiproduct Calibrator
	(-250 to -100) °C	0.39 °C	
	(-100 to -25) °C	0.13 °C	
	(-25 to 350) °C	0.11 °C	
	(350 to 650) °C	0.13 °C	
	(650 to 1 000) °C	0.17 °C	
	Type J		
	(-210 to -100) °C	0.21 °C	
	(-100 to -30) °C	0.13 °C	
	(-30 to 150) °C	0.11 °C	
	(150 to 760) °C	0.14 °C	
	(760 to 1 200) °C	0.18 °C	
	Type K		
	(-200 to -100) °C	0.26 °C	
	(-100 to -25) °C	0.14 °C	
	(-25 to 120) °C	0.13 °C	
	(120 to 1 000) °C	0.21 °C	
	(1 000 to 1 372) °C	0.31 °C	
	Type L		
	(-200 to -100) °C	0.29 °C	
	(-100 to 800) °C	0.21 °C	
	(800 to 900) °C	0.14 °C	
	Type N		
	(-200 to -100) °C	0.31 °C	
	(-100 to -25) °C	0.18 °C	
	(-25 to 120) °C	0.15 °C	
	(120 to 410) °C	0.14 °C	
	(410 to 1 300) °C	0.21 °C	
	Type R		
	(0 to 250) °C	0.45 °C	
	(250 to 400) °C	0.28 °C	
	(400 to 1 000) °C	0.26 °C	
	(1 000 to 1 767) °C	0.31 °C	
	Type S		
	(0 to 250) °C	0.37 °C	
	(250 to 1 000) °C	0.28 °C	
	(1 000 to 1 400) °C	0.29 °C	
	(1 400 to 1 767) °C	0.36 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ^{1,3}	Type T (-250 to -150) °C	0.49 °C	Fluke 5522A Multiproduct Calibrator
	(-150 to 0) °C	0.19 °C	
	(0 to 120) °C	0.13 °C	
	(120 to 400) °C	0.11 °C	
	Type U (-200 to 0) °C	0.44 °C	
	(0 to 600) °C	0.21 °C	
Phase Angle – Source ^{1,3}	Up to 360 °		Fluke 5522A Multiproduct Calibrator
	(10 to 65) Hz	0.08 °	
	(65 to 500) Hz	0.2 °	
	500 Hz to 1 kHz	0.4 °	
	(1 to 5) kHz	2 °	
	(5 to 10) kHz	4 °	
	(10 to 30) kHz	7.8 °	

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure ^{1,3,4} Up to 26.5 GHz	(-70 to -30) dBm		Agilent 4418B Power Meter, Agilent 8842A Power Sensor, Agilent 8485D Power Sensor
	50 MHz to 8 GHz	0.15 dB	
	(8 to 20) GHz	0.2 dB	
	(20 to 26.5) GHz	0.22 dB	
	(-30 to +10) dBm		
	100 kHz to 2.5 GHz	0.04 dB	
	(2.5 to 4.2) GHz	0.09 dB	
	(4.2 to 13) GHz	0.12 dB	
	(13 to 18) GHz	0.14 dB	
	(18 to 26.5) GHz	0.16 dB	
RF Power – Measure ^{1,3,4}	(10 to 20) dBm		Agilent 4418B Power Meter, Agilent 8482A/8485A Power Sensors
	100 kHz to 2.5 GHz	0.16 dB	
	(2.5 to 4.2) GHz	0.16 dB	
	(4.2 to 13) GHz	0.18 dB	
	(13 to 18) GHz	0.19 dB	
	(18 to 26.5) GHz	0.21 dB	
Amplitude Modulation – Measure ^{1,3}	(5 to 99) % Depth		HP 8902A Measuring Receiver
	150 kHz to 10 MHz	2.4 % Depth	
	10 MHz to 1.3 GHz	1.2 % Depth	

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation – Measure ^{1,3}	(5 to 99) % Deviation 150 kHz to 10 MHz 10 MHz to 1.3 GHz	2.4 % of reading 1.2 % of reading	HP 8902A Measuring Receiver
Phase Modulation – Measure ^{1,3}	(5 to 99) % Deviation 150 kHz to 10 MHz 10 MHz to 1.3 GHz	3.6 % of reading 3.6 % of reading	HP 8902A Measuring Receiver
Tuned RF Level Attenuation – Measure ^{1,3,4}	2.5 MHz to 1.3 GHz (-10 to 0) dBm (-20 to -10) dBm (-30 to -20) dBm (-40 to -30) dBm (-50 to -40) dBm (-60 to -50) dBm (-70 to -60) dBm (-80 to -70) dBm (-90 to -80) dBm (-100 to -90) dBm (-120 to -100) dBm	0.04 dB 0.05 dB 0.07 dB 0.08 dB 0.1 dB 0.11 dB 0.13 dB 0.15 dB 0.16 dB 0.18 dB 0.19 dB	HP 8902A Measuring Receiver, HP 11722A Power Sensor
Harmonics – Measure ^{1,3}	DC Coupled (-80 to 0) dB 30 Hz to 6.5GHz	2.1 dB	HP 8561E Spectrum Analyzer
	AC Coupled (-80 to 0) dB 100 kHz to 6.5 GHz	2.4 dB	
AM Distortion – Measure ^{1,3}	(-80 to 0) dB 20 Hz to 20 kHz (20 to 100) kHz	1.2 dB 2.4 dB	HP 8903B Audio Analyzer
FM Distortion – Measure ^{1,3}	(-80 to 0) dB 20Hz to 20 kHz (20 to 100) kHz	1.2 dB 2.4 dB	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length Standards	Up to 4 in (4 to 10) in (10 to 20) in	(9.6 + 2.4L) μ in (29 + 3.2L) μ in (29 + 3.1L) μ in	Universal Length Measuring Machine, Grade 0 Gage Block Set

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers ^{1,2} Outside Diameter 0.000 5 in Resolution Parallelism, ID, Depth, Step 0.000 5 in Resolution	(0.01 to 20) in (20 to 40) in (40 to 48) in (0.01 to 1) in	(286 + 1.1L) μin (266 + 2.4L) μin (256 + 3L) μin 300 μin	Gage Blocks, Gage Block Accessory Set, Surface Plate, Master Gage Pins
Inside Micrometers ^{1,2}	Up to 4 in (5 to 20) in (24 to 40) in (44 to 48) in	(53 + 1.9L) μin (53 + 3.4L) μin (59 + 4.2L) μin (66 + 4.5L) μin	Gage Blocks, Gage Block Accessory Set (ID)
Outside Micrometers ^{1,2}	Up to 4 in (4 to 20) in (20 to 40) in	(55 + 1.8L) μin (74 + 3L) μin (266 + 2.4L) μin	Gage Blocks
Depth Gages ^{1,2} (Micrometer/Indicator)	Up to 4 in (5 to 20) in (24 to 40) in (44 to 48) in	(73 + 4L) μin (87 + 5.5L) μin (69 + 4.1L) μin (94 + 4L) μin	Gage Blocks, Surface Plate
Dial/Digital Indicators ^{1,2} 0.000 05 in Resolution 0.000 1 in Resolution	(0.01 to 2) in (0.01 to 4) in	82 μin (94 + 1.4L) μin	Gage Blocks, Surface Plate
Height Gages ^{1,2,3}	Up to 20 in (20 to 40) in (40 to 48) in	(70 + 3.6L) μin (66 + 4.8L) μin (76 + 5L) μin	Gage Blocks, Surface Plate
Laser Micrometers ¹	0.25 in 0.5 in 0.75 in 1 in	25 μin 25 μin 25 μin 30 μin	XX Master Pin Gages
Rules, Scales, Tape Measures	(0.01 to 48) in	0.018 in	Gage Blocks, Surface Plate

Mass and Mass-Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Gages, Pressure Transducers ^{1,3}	(0.2 to 50) psia (0.2 to 50) psig	0.001 2 % of reading	Ruska 2465A Piston Gauge

Mass and Mass-Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Gages, Pressure Transducers ^{1,3}	(29 to 300) psia	0.14 % of reading	Comparison to Druck DPI-145 Pressure Indicator
Pressure Gages, Pressure Transducers ^{1,3}	(50 to 1 300) psig (1 300 to 10 000) psig (10 000 to 30 000) psig	0.042 % of reading + 0.002 psi 0.042 % of reading + 0.012 psi 0.06 % of reading	Mansfield and Green TD-40000N Deadweight Tester
Vacuum ^{1,3}	(-15 to 50) psi	0.001 2 % of reading	Ruska 2465A Piston Gauge
Volumetric Flow ^{1,3}	(5 to 500) sccm 500 sccm to 50 slpm	0.24 % of reading 0.24 % of reading	Comparison to DryCal ML-800-10 and DryCal ML-800-44 Flow Cells
Volumetric Flow ^{1,3}	(0.7 to 7) scfm (7 to 25) scfm (25 to 90) scfm	0.1 % of reading + 0.025 scfm 0.1 % of reading + 0.14 scfm 0.1 % of reading + 0.36 scfm	Comparison to Cox 16-064, Cox 16-121, and Cox 16-228 Sonic Nozzles
Pipettes ³	(10 to 20) µl (20 to 50) µl (50 to 100) µl (100 to 200) µl (200 to 500) µl (500 to 1 000) µl	65 nl 66 nl 69 nl 80 nl 0.14 µl 0.24 µl	A&D 4212B-101 Balance and Software
Torque Transducers ^{1,3}	(0.5 to 10) lbf·in	0.03 % of reading + 0.000 1 lbf·in	NIST Class F Weights, 4 in Torque Wheel
	(10 to 1 920) lbf·in	0.018 % of reading + 0.000 1 lbf·in	NIST Class F Weights, 10 in Torque Wheel
Torque Transducers ^{1,3}	(160 to 1 000) lbf·ft	0.24 % of reading + 0.01 lbf·ft	NIST Class F Weights, 4 ft Torque Arm
Torque Tools ^{1,3}	(10 to 96) ozf·in	0.6 % of reading	AIMCO UET-0100 Torque Tester
	(6 to 96) lbf·in	0.6 lbf·in	HIOS HP-100 Torque Tester
Torque Tools ^{1,3}	(8 to 295) lbf·ft	0.6 % of reading + 0.03 lbf·ft	Norbar Pro Test 400 (43219)
	(295 to 1 100) lbf·ft	1.2 % of reading	Norbar Pro Test 1500ER (43189)
Force ^{1,3} (Compression & Tension)	Up to 50 lbf	0.006 lbf	NIST Class F Weights

Mass and Mass-Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force ^{1,3} (Compression & Tension)	(50 to 1 000) lbf	0.31 lbf	1 000 lbf Interface Load Cell
	(1 000 to 10 000) lbf	2.5 lbf	10 000 lbf Interface Load Cell
	(10 000 to 50 000) lbf	13 lbf	50 000 lbf Interface Load Cell
Force ^{1,3} (Compression & Tension)	Up to 20 lb (20 to 50) lb (50 to 100) lb (100 to 200) lb (200 to 400) lb (400 to 600) lb	1.5 g 3.8 g 7.5 g 15 g 30 g 45 g	NIST Class F Weights
Laboratory Balances ^{1,5}	Up to 2 mg (2 to 10) mg (10 to 20) mg (20 to 100) mg (100 to 200) mg (200 to 500) mg 500 mg to 1 g	29 µg 35 µg 41 µg 58 µg 70 µg 93 µg 0.12 mg	ASTM E617 Class 3 Weights and NIST HB 44 utilized in the calibration of the weighing system.
Laboratory Balances ^{1,5}	(1 to 5) g (5 to 10) g (10 to 20) g (20 to 50) g (50 to 100) g (100 to 200) g (200 to 500) g	22 µg 30 µg 44 µg 83 µg 0.16 mg 0.3 mg 0.72 mg	Ultra-Class Weights and NIST HB 44 utilized in the calibration of the weighing system.
Scales ^{1,5}	500 g to 1 kg (1 to 2) kg (2 to 5) kg (5 to 10) kg (10 to 25) kg (25 to 40) kg	2.9 mg 3.1 mg 15 mg 33 mg 74 mg 82 mg	ASTM E617 Class 1 Weights and NIST HB 44 utilized in the calibration of the weighing system.
Scales ^{1,5}	Up to 20 lb (20 to 50) lb (50 to 100) lb (100 to 200) lb (200 to 400) lb (400 to 600) lb	1.1 g 2.7 g 5.4 g 11 g 22 g 32 g	NIST Class F Weights and NIST HB 44 utilized in the calibration of the weighing system.

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Measure ^{1,3}	Up to 90 %RH (90 to 95) %RH	1.3 %RH 2.2 %RH	Vaisala HMP-363 Transmitter
Humidity – Source ^{1,3}	11.3 %RH 32.9 %RH 75.4 %RH 96.7 %RH	1.3 %RH 1.3 %RH 1.3 %RH 2.2 %RH	Vaisala HMP-363 Transmitter, Saturated Salt Solutions
Temperature – Measure ^{1,3}	(-196 to 0) °C (0 to 420) °C	35 mK 0.005 % of reading + 29 mK	Fluke 5628 SPRT, HP 3458A Opt 002 8.5 Digit Multimeter
Temperature – Source ^{1,3} (Temperature Measuring Instruments)	(-40 to 0) °C (0 to 200) °C (200 to 400) °C	36 mK 0.016 % of reading + 35 mK 0.006 % of reading + 60 mK	Additel 878-160 Drywell Calibrator, Omega DBCL400 Drywell Calibrator, Fluke 5628 SPRT, HP 3458A 8.5 Digit Multimeter


Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source ^{1,3}	1 µHz to 250 kHz	70 pHz/Hz + 0.58 µHz	Agilent 33220A Signal Generator Locked to Datum LPRO Rubidium Freq Std
Frequency – Source ^{1,3}	250 kHz to 3 GHz	70 pHz/Hz + 5.8 mHz	HP ESG-D3000A Signal Generator locked to Datum LPRO Rubidium Freq Std
Frequency – Source ^{1,3}	10 Hz to 2 MHz	0.003 % of reading + 3 mHz	Fluke 5522A Multiproduct Calibrator
Frequency – Measure ^{1,3}	100 mHz to 225 MHz	76 pHz/Hz	Agilent 53131A Counter locked to Datum LPRO Rubidium Freq Standard
Frequency – Measure ^{1,3}	225 MHz to 26.5 GHz	128 pHz/Hz	HP 5348A Counter locked to Datum LPRO Rubidium Freq Standard
Stopwatches/Timers	Up to 24 h	61 ms/d	NIST 960-12 Time Base Method utilizing Timometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = length in inches, t = time in seconds.
3. Uncertainties do not include contributors from a “best available” unit under test.
4. The uncertainty does not include the mismatch uncertainty. This will be determined and reported at time of calibration. The reported uncertainty will be higher than listed.
5. The uncertainties for scales and balances are highly dependent upon the resolution of the unit under test. The uncertainties presented here do not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
6. This location is part of a Multi-site Organization whose legal entity name is Transcat, Inc.
7. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1287.



Jason Stine, Vice President

