



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat – Philadelphia
100 Dobbs Lane, Suite 108-110
Cherry Hill, NJ 08034

Fulfills the requirements of

ISO/IEC 17025:2017

and the national standards

ANSI/NCSL Z540-1-1994 (R2002) AND
ANSI/NCSL Z540.3-2006 (R2013)

In the fields of

CALIBRATION AND DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 07 September 2025
Certificate Number: AC-2489.03



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

ANSI/NCSL Z540.3-2006 (R2013)

Transcat – Philadelphia

100 Dobbs Lane, Suite 108-110

Cherry Hill, NJ 08034

Alex Katyan alex.katyan@transcat.com

CALIBRATION AND DIMENSIONAL MEASUREMENT

Valid to: September 7, 2025

Certificate Number: AC-2489.03

CALIBRATION

Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH – Measuring Equipment ¹	4 pH 7 pH 10 pH	0.011 pH 0.011 pH 0.012 pH	Accredited Reference Material; Direct Measure
Conductivity Meters – Measuring Equipment	5 µS/cm 10 µS/cm 100 µS/cm 1000 µS/cm 10 000 µS/cm 100 000 µS/cm 150 000 µS/cm	0.35 µS/cm 0.35 µS/cm 0.84 µS/cm 3.5 µS/cm 38 µS/cm 310 µS/cm 610 µS/cm	Accredited Reference Material; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source ¹	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	7.5 µV/V + 0.4 µV 5 µV/V + 0.7 µV 3.5 µV/V + 2.5 µV 3.5 µV/V + 4 µV 5 µV/V + 40 µV 6.5 µV/V + 0.4 mV	Fluke 5730A Multiproduct Calibrator; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Measure ¹	Up to 100 mV 100 mV to 1 V (1 to 10) V (10 to 100) V (100 to 500) V (500 to 800) V (800 to 1 000) V	8.3 $\mu\text{V/V} + 0.58 \mu\text{V}$ 5.3 $\mu\text{V/V} + 0.58 \mu\text{V}$ 5.3 $\mu\text{V/V} + 0.58 \mu\text{V}$ 7.7 $\mu\text{V/V} + 35 \mu\text{V}$ 15 $\mu\text{V/V} + 0.12 \text{ mV}$ 18 $\mu\text{V/V} + 0.12 \text{ mV}$ 21 $\mu\text{V/V} + 0.12 \text{ mV}$	Agilent 3458A Opt.002 8.5 Digit Multimeter; Direct Measure
DC High Voltage – Measure ¹	(1 to 10) kV (10 to 35) kV (35 to 70) kV (70 to 100) kV	0.04 % of reading + 92 mV 0.064 % of reading + 0.66 V 0.088 % of reading + 0.81 V 0.17 % of reading + 0.92 V	Vitretek 4700A High Voltage Meter; Vitretek HVL-35, HVL-70, HVL-100 High Voltage Probes
AC Voltage – Source ¹	Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.024 % of reading + 4 μV 0.009 % of reading + 4 μV 0.008 % of reading + 4 μV 0.02 % of reading + 4 μV 0.05 % of reading + 5 μV 0.11 % of reading + 10 μV 0.14 % of reading + 20 μV 0.27 % of reading + 20 μV 0.024 % of reading + 4 μV 0.009 % of reading + 4 μV 0.008 % of reading + 4 μV 0.02 % of reading + 4 μV 0.05 % of reading + 5 μV 0.11 % of reading + 10 μV 0.14 % of reading + 20 μV 0.27 % of reading + 20 μV 0.024 % of reading + 12 μV 0.009 % of reading + 7 μV 0.005 7 % of reading + 7 μV 0.012 % of reading + 7 μV 0.031 % of reading + 17 μV 0.066 % of reading + 20 μV 0.14 % of reading + 25 μV 0.27 % of reading + 45 μV	Fluke 5730A Multiproduct Calibrator; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(0.22 to 2.2) V		Fluke 5730A Multiproduct Calibrator; Direct Measure
	(10 to 20) Hz	0.024 % of reading + 40 μ V	
	(20 to 40) Hz	0.009 % of reading + 15 μ V	
	40 Hz to 20 kHz	0.004 2 % of reading + 8 μ V	
	(20 to 50) kHz	0.006 7 % of reading + 10 μ V	
	(50 to 100) kHz	0.008 5 % of reading + 30 μ V	
	(100 to 300) kHz	0.034 % of reading + 80 μ V	
	(300 to 500) kHz	0.1 % of reading + 0.2 mV	
	500 kHz to 1 MHz	0.17 % of reading + 0.3 mV	
	(2.2 to 22) V		
	(10 to 20) Hz	0.024 % of reading + 0.4 mV	
	(20 to 40) Hz	0.009 % of reading + 0.15 mV	
	40 Hz to 20 kHz	0.004 2 % of reading + 50 μ V	
	(20 to 50) kHz	0.006 7 % of reading + 0.1 mV	
	(50 to 100) kHz	0.008 3 % of reading + 0.2 mV	
	(100 to 300) kHz	0.034 % of reading + 0.6 mV	
	(300 to 500) kHz	0.1 % of reading + 2 mV	
	500 kHz to 1 MHz	0.17 % of reading + 3.2 mV	
	(22 to 220) V		
	(10 to 20) Hz	0.024 % of reading + 4 mV	
	(20 to 40) Hz	0.009 % of reading + 1.5 mV	
AC Voltage – Source ¹	40 Hz to 20 kHz	0.005 2 % of reading + 0.6 mV	Fluke 5730A Multiproduct Calibrator, Fluke 5725A Amplifier; Direct Measure
	(20 to 50) kHz	0.008 % of reading + 1 mV	
	(50 to 100) kHz	0.015 % of reading + 2.5 mV	
	(100 to 300) kHz	0.09 % of reading + 16 mV	
	(300 to 500) kHz	0.44 % of reading + 40 mV	
	500 kHz to 1 MHz	0.8 % of reading + 80 mV	
	(220 to 250) V		
	(15 to 50) Hz	0.03 % of reading + 16 mV	
	(250 to 1 100) V		
	50 Hz to 1 kHz	0.007 % of reading + 3.5 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	Up to 10 mV		Agilent 3458A Opt.002 8.5 Digit Multimeter; Direct Measure
	(1 to 40) Hz	0.04 % of reading + 3.5 μ V	
	40 Hz to 1 kHz	0.03 % of reading + 1.2 μ V	
	(1 to 20) kHz	0.04 % of reading + 1.2 μ V	
	(20 to 50) kHz	0.15 % of reading + 1.2 μ V	
	(50 to 100) kHz	0.59 % of reading + 1.2 μ V	
	(100 to 300) kHz	4.6 % of reading + 2.3 μ V	
	300 kHz to 1 MHz	1.5 % of reading + 5.8 μ V	
	1 MHz to 4 MHz	8.1 % of reading + 8.1 μ V	
	(10 to 100) mV		
	(1 to 40) Hz	0.013 % of reading + 4.6 μ V	
	40 Hz to 1 kHz	0.009 7 % of reading + 2.3 μ V	
	(1 to 20) kHz	0.017 % of reading + 2.3 μ V	
	(20 to 50) kHz	0.038 % of reading + 2.3 μ V	
	(50 to 100) kHz	0.093 % of reading + 2.3 μ V	
	(100 to 300) kHz	0.36 % of reading + 12 μ V	
	300 kHz to 1 MHz	1.2 % of reading + 12 μ V	
	(1 to 2) MHz	1.8 % of reading + 12 μ V	
	(2 to 4) MHz	4.7 % of reading + 81 μ V	
	(4 to 8) MHz	4.7 % of reading + 92 μ V	
	(8 to 10) MHz	17 % of reading + 0.12 mV	
	(0.1 to 1) V		
	(1 to 40) Hz	0.008 8 % of reading + 46 μ V	
	40 Hz to 1 kHz	0.008 3 % of reading + 23 μ V	
	(1 to 20) kHz	0.017 % of reading + 23 μ V	
	(20 to 50) kHz	0.036 % of reading + 23 μ V	
	(50 to 100) kHz	0.093 % of reading + 23 μ V	
	(100 to 300) kHz	0.35 % of reading + 0.12 mV	
	300 kHz to 1 MHz	1.2 % of reading + 0.12 mV	
	(1 to 2) MHz	1.8 % of reading + 0.12 mV	
	(2 to 4) MHz	4.6 % of reading + 0.81 mV	
	(4 to 8) MHz	4.6 % of reading + 0.92 mV	
	(8 to 10) MHz	17 % of reading + 1.2 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	(1 to 10) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.009 5 % of reading + 0.46 mV 0.023 % of reading + 0.23 mV 0.017 % of reading + 0.23 mV 0.036 % of reading + 0.23 mV 0.093 % of reading + 0.23 mV 0.35 % of reading + 1.2 mV 1.2 % of reading + 1.2 mV 1.8 % of reading + 1.2 mV 4.6 % of reading + 8.1 mV 4.6 % of reading + 9.2 mV 17 % of reading + 12 mV 0.024 % of reading + 4.6 mV 0.024 % of reading + 2.3 mV 0.024 % of reading + 2.3 mV 0.041 % of reading + 2.3 mV 0.14 % of reading + 2.3 mV 0.46 % of reading + 12 mV 1.7 % of reading + 12 mV 0.048 % of reading + 46 mV 0.048 % of reading + 23 mV 0.071 % of reading + 23 mV 0.19 % of reading + 23 mV 0.35 % of reading + 23 mV	Agilent 3458A Opt.002 8.5 Digit Multimeter; Direct Measure
	(1 to 10) kV (10 to 200) Hz (200 to 450) Hz (450 to 600) Hz (10 to 35) kV (30 to 200) Hz (200 to 450) Hz (450 to 600) Hz (35 to 70) kV (30 to 100) Hz (100 to 450) Hz (450 to 600) Hz (70 to 100) kV (30 to 100) Hz (100 to 450) Hz (450 to 600) Hz	0.14 % of reading + 0.17 V 0.46 % of reading + 0.17 V 0.86 % of reading + 0.17 V 0.11 % of reading + 0.81 V 0.7 % of reading + 0.81 V 1.5 % of reading + 0.81 V 0.14 % of reading + 1 V 0.7 % of reading + 1 V 2.9 % of reading + 1 V 0.21 % of reading + 1.3 V 1.2 % of reading + 1.3 V 1.7 % of reading + 1.3 V	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source ¹ (Simulation)	10 Hz to 10 kHz 190 pF to 1.1 nF	0.39 % of reading + 7.8 pF	Fluke 5522A Multiproduct Calibrator; Direct Measure
	10 Hz to 3 kHz (1.1 to 3.3) nF	0.39 % of reading + 7.8 pF	
	10 Hz to 1 kHz (3.3 to 11) nF	0.21 % of reading + 7.8 pF	
	(11 to 110) nF	0.21 % of reading + 78 pF	
	10 Hz to 1 kHz (110 to 330) nF	0.21 % of reading + 0.23 nF	
	(10 to 600) Hz 330 nF to 1.1 μF	0.21 % of reading + 0.78 nF	
	(10 to 300) Hz (1.1 to 3.3) μF	0.21 % of reading + 2.3 nF	
	(10 to 150) Hz (3.3 to 11) μF	0.21 % of reading + 7.8 nF	
	(10 to 120) Hz (11 to 33) μF	0.32 % of reading + 23 nF	
	(10 to 80) Hz (33 to 110) μF	0.35 % of reading + 78 nF	
	DC to 50 Hz (110 to 330) μF	0.35 % of reading + 0.23 μF	
	DC to 20 Hz 330 μF to 1.1 mF	0.35 % of reading + 0.78 μF	
	DC to 6 Hz (1.1 to 3.3) mF	0.35 % of reading + 2.3 μF	
	DC to 2 Hz (3.3 to 11) mF	0.35 % of reading + 7.8 μF	
	DC to 0.6 Hz (11 to 33) mF	0.58 % of reading + 23 μF	
	DC to 0.2 Hz (33 to 110) mF	0.85 % of reading + 78 μF	
Capacitance – Measure ¹	100 Hz to 1 kHz Up to 10 pF	0.5 % of reading + 50 fF	GR 1689-M Precision Impedance Meter; Direct Measure
	(10 to 100) pF	0.059 % of reading + 50 fF	
	100 pF to 1 μF (1 to 100) μF	0.027 % of reading + 50 fF	
	100 μF to 1 mF	0.037 % of reading + 50 fF	
		0.24 % of reading	
DC Current – Source ¹	(0.2 to 220) μA	40 μA/A + 6 nA	Fluke 5730A Multiproduct Calibrator; Direct Measure
	(0.22 to 2.2) mA	35 μA/A + 7 nA	
	(2.2 to 22) mA	35 μA/A + 40 nA	
	(22 to 220) mA	45 μA/A + 0.7 μA	
	(0.22 to 2.2) A	80 μA/A + 12 μA	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Source ¹	Up to 2 A (2.2 to 11) A (2 to 20) A	0.036 % of reading + 0.48 mA 0.012 % of reading + 0.16 mA 0.012 % of reading + 1.6 mA	Fluke 5730A Multiproduct Calibrator, Fluke 5725A Amplifier; Direct Measure
DC Current – Source ¹	(20 to 120) A	0.012 % of reading + 9.6 mA	Fluke 5730A Multiproduct Calibrator, Fluke 52120A Current Amplifier; Direct Measure
DC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹	(20 to 54.999 5) A (55 to 149.999 5) A (150 to 549.995) A (550 to 1 025) A	0.47 % of reading + 0.11 A 0.41 % of reading + 0.11 A 0.45 % of reading + 0.39 A 0.51 % of reading + 0.39 A	Fluke 5520A Multiproduct Calibrator, 50-turn Coil; Direct Measure
DC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹	(1 to 5) kA	0.58 % of reading	Fluke 5522A Multiproduct Calibrator, Fluke 52120A Amplifier, 3 kA Coil, 6 kA Coil; Direct Measure
DC Current – Source/Measure ¹	Up to 100 μ A 100 μ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A	33 μ A/A + 0.92 nA 29 μ A/A + 5.8 nA 29 μ A/A + 58 nA 46 μ A/A + 0.58 μ A 0.013 % of reading + 12 μ A	Current Source, Characterized with Agilent 3458A Opt.002 8.5 Digit Multimeter; Direct Measure
DC Current – Source/Measure ¹	(10 to 100) A	0.15 mA/A + 3 mA	Ohms Labs CS-100 Current Shunt, Agilent 3458A Opt. 002 8.5 Digit Multimeter; Direct Measure
DC Current – Measure ¹	(1 to 3) A (3 to 10) A	0.096 % of reading + 0.47 mA 0.12 % of reading + 0.62 mA	Fluke 8846A 6.5 Digit Multimeter; Direct Measure
DC Current – Measure ¹	(100 to 1 500) A	0.37 % of reading + 0.17 A	Empro WT-1500-50 Current Shunt, Agilent 3458A Opt. 002 8.5 Digit Multimeter; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	Up to 220 μ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.025 % of reading + 16 nA 0.016 % of reading + 10 nA 0.011 % of reading + 8 nA 0.028 % of reading + 12 nA 0.11 % of reading + 65 nA	Fluke 5730A Multiproduct Calibrator; Direct Measure
AC Current – Source ¹	(0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.25 % of reading + 40 nA 0.016 % of reading + 35 nA 0.011 % of reading + 35 nA 0.02 % of reading + 0.11 μ A 0.11 % of reading + 0.65 μ A 0.025 % of reading + 0.4 μ A 0.016 % of reading + 0.35 μ A 0.011 % of reading + 0.35 μ A 0.020 % of reading + 0.55 μ A 0.11 % of reading + 5 μ A	Fluke 5730A Multiproduct Calibrator; Direct Measure
AC Current – Source ¹	(2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.025 % of reading + 0.4 μ A 0.016 % of reading + 0.35 μ A 0.011 % of reading + 0.35 μ A 0.02 % of reading + 0.55 μ A 0.11 % of reading + 5 μ A 0.025 % of reading + 4 μ A 0.016 % of reading + 3.5 μ A 0.011 % of reading + 2.5 μ A 0.02 % of reading + 3.5 μ A 0.11 % of reading + 10 μ A	Fluke 5730A Multiproduct Calibrator, Fluke 5725A Amplifier; Direct Measure
AC Current – Source ¹	(0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 11) A (40 to 100) Hz (1 to 5) kHz (5 to 10) kHz	0.025 % of reading + 35 μ A 0.045 % of reading + 80 μ A 0.7 % of reading + 0.16 μ A 0.046 % of reading + 0.17 mA 0.095 % of reading + 0.38 mA 0.36 % of reading + 0.75 mA	Fluke 5730A Multiproduct Calibrator, Fluke 5725A Amplifier; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	Up to 2 A (10 to 850) Hz 850 Hz to 6 kHz (6 to 10) kHz (2 to 20) A (10 to 850) Hz 850 Hz to 6 kHz (6 to 10) kHz	0.009 % of reading + 40 μ A 0.04 % of reading + 80 μ A 1.6 % of reading + 62 mA 0.009 % of reading + 0.4 mA 0.04 % of reading + 0.8 mA 2.3 % of reading + 94 mA	Fluke 5730A Multiproduct Calibrator, Fluke 52120A Current Amplifier; Direct Measure
AC Current – Source ¹	(20 to 120) A (10 to 850) Hz 850 Hz to 6 kHz (6 to 10) kHz	0.009 % of reading + 2.4 mA 0.04 % of reading + 4.8 mA 3.1 % of reading + 0.7 A	Fluke 5730A Multiproduct Calibrator, Fluke 52120A Current Amplifier; Direct Measure
AC Current – Source ¹ Extended Frequency Ranges	(29 to 330) μ A (10 to 30) kHz (0.33 to 3.3) mA (10 to 30) kHz (3.3 to 33) mA (10 to 30) kHz (33 to 330) mA (10 to 30) kHz	1.2 % of reading + 0.31 μ A 0.78 % of reading + 0.47 μ A 0.031 % of reading + 3.1 μ A 0.31 % of reading + 0.16 mA	Fluke 5522A Multiproduct Calibrator; Direct Measure
AC Clamp-on Ammeters (Toroidal Type) Transformer Type Sensor ¹	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 025) A (45 to 65) Hz (65 to 440) Hz	0.24% of reading + 19 mA 0.62% of reading + 21 mA 0.24% of reading + 70 mA 0.61% of reading + 78 mA	Fluke 5520A Multiproduct Calibrator, 5500A/COIL 50-turn Coil; Direct Measure
AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.5 % of reading + 0.19 A 0.9 % of reading + 0.19 A 0.52 % of reading + 0.7 A 0.9 % of reading + 0.7 A	Fluke 5520A Multiproduct Calibrator, Fluke 5500A/COIL 50-turn Coil; Direct Measure
AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹	(10 to 300) Hz (1 to 6) kA (300 to 440) Hz (1 to 2) kA (2 to 6) kA	0.6 % of reading 0.8 % of reading 0.66 % of reading	Fluke 5522A Multiproduct Calibrator, Fluke 52120A Amplifier, 3 kA Coil, 6 kA Coil; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ¹	Up to 100 μ A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 1 kHz (0.1 to 1) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.46 % of reading + 35 nA 0.17 % of reading + 35 nA 0.072 % of reading + 35 nA 0.072 % of reading + 35 nA 0.46 % of reading + 0.23 μ A 0.17 % of reading + 0.23 μ A 0.071 % of reading + 0.23 μ A 0.038 % of reading + 0.23 μ A	Agilent 3458A Opt.002 8.5 Digit Multimeter; Direct Measure
AC Current – Measure ¹	(1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (0.1 to 1) A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.46 % of reading + 2.3 μ A 0.17 % of reading + 2.3 μ A 0.071 % of reading + 2.3 μ A 0.038 % of reading + 2.3 μ A 0.48 % of reading + 23 μ A 0.17 % of reading + 23 μ A 0.071 % of reading + 23 μ A 0.037 % of reading + 23 μ A 0.46 % of reading + 0.23 mA 0.19 % of reading + 0.23 mA 0.097 % of reading + 0.23 mA 0.12 % of reading + 0.23 mA	Agilent 3458A Opt.002 8.5 Digit Multimeter; Direct Measure
AC Current – Measure	(0.2 to 2) A 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (2 to 20) A 10 Hz to 2 kHz (2 to 10) kHz (20 to 30) A 10 Hz to 2 kHz (2 to 10) kHz	0.3 mA/A + 0.1 mA 0.56 mA/A + 0.1 mA 0.8 mA/A + 0.1 mA 0.84 mA/A + 0.5 mA 0.86 mA/A + 0.5 mA 0.84 mA/A + 12 mA 1.2 mA/A + 12 mA	Fluke 8588A 8.5 Digit Multimeter; Direct Measure
AC Current – Measure ¹	(10 to 100) A 50/60 Hz 400 Hz 1 kHz	0.023 % of reading + 5 mA 0.11 % of reading + 5 mA 0.2 % of reading + 5 mA	Ohms Labs CS-100 Current Shunt, Agilent 3458A Opt. 002 8.5 Digit Multimeter; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance – Source ¹ (Fixed Artifacts)	1 mΩ 10 mΩ 100 mΩ 1 Ω 100 Ω	59 μΩ/Ω 58 μΩ/Ω 58 μΩ/Ω 58 μΩ/Ω 1.3 μΩ/Ω	Standard Resistors; Direct Measure
DC Resistance – Source ¹ (Variable Artifact)	(1 to 10) GΩ (10 to 100) GΩ 100 GΩ to 1 TΩ	0.59 % of reading + 1.2 μΩ/Ω/V 1.2 % of reading + 2.3 μΩ/Ω/V 1.2 % of reading + 5.8 μΩ/Ω/V	IET HRRS-B-7-100k-5kV Decade Resistor; Direct Measure (V is the DUT Voltage)
DC Resistance – Source ¹ (Simulation)	(1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ	0.005 9 % of reading + 23 Ω 0.011 % of reading + 39 Ω 0.021 % of reading + 1.9 kΩ 0.041 % of reading + 2.3 kΩ 0.26 % of reading + 78 kΩ	Fluke 5522A Multiproduct Calibrator; Direct Measure
DC Resistance – Source/Measure ¹	250 μΩ to 4 mΩ (4 to 40) mΩ (40 to 400) mΩ 400 mΩ to 4 Ω 4 Ω to 400 kΩ	85 μΩ/Ω 25 μΩ/Ω 20 μΩ/Ω 16 μΩ/Ω 5 μΩ/Ω	Fluke 1594A Bridge in Ratio Mode, Characterized Resistors; Direct Measure
DC Resistance – Source/Measure ¹	Up to 10 Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1 GΩ	18 μΩ/Ω + 58 μΩ 15 μΩ/Ω + 0.58 mΩ 13 μΩ/Ω + 0.58 mΩ 12 μΩ/Ω + 5.8 mΩ 13 μΩ/Ω + 58 mΩ 21 μΩ/Ω + 2.3 Ω 62 μΩ/Ω + 0.12 kΩ 0.059 % of reading + 1.2 kΩ 0.82 % of reading + 12 kΩ	Agilent 3458A Opt. 002 8.5 Digit Multimeter, Decade Resistors; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance – Source (Fixed-Point Simulation)	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	40 $\mu\Omega$ 95 $\mu\Omega/\Omega$ 95 $\mu\Omega/\Omega$ 23 $\mu\Omega/\Omega$ 23 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 6.5 $\mu\Omega/\Omega$ 6.5 $\mu\Omega/\Omega$ 6.5 $\mu\Omega/\Omega$ 6.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 13 $\mu\Omega/\Omega$ 18 $\mu\Omega/\Omega$ 40 $\mu\Omega/\Omega$ 47 $\mu\Omega/\Omega$ 0.1 m Ω/Ω	Fluke 5730A Multiproduct Calibrator; Direct Measure
AC Resistance – Measure ^{1,10}	10 Ω to 100 k Ω (12 to 29) Hz (30 to 999) Hz 1 kHz 10 Ω to 100 k Ω 10 kHz 50 kHz	0.11 % of reading 0.06 % of reading 0.03 % of reading 0.06 % of reading 0.21 % of reading	GenRad 1689M Precision Impedance Meter; Direct Measure
Inductance – Source ¹ (Fixed Artifacts)	1 kHz 1 mH 10 mH 100 mH 1 H	0.13 % of reading 0.13 % of reading 0.13 % of reading 0.13 % of reading	Standard Inductors; Direct Measure
Inductance – Measure ¹	100 Hz to 1 kHz (1 to 10) mH 10 mH to 10 H	0.041 % of reading + 0.1 μH 0.035 % of reading + 1.4 μH	GenRad 1689M Precision Impedance Meter; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹	Type B		Ectron 1140A Thermocouple Calibrator/Simulator; Direct Measure
	(250 to 350) °C	1.2 °C	
	(350 to 445) °C	0.9 °C	
	(445 to 580) °C	0.71 °C	
	(580 to 750) °C	0.55 °C	
	(750 to 1 000) °C	0.45 °C	
	(1 000 to 1 820) °C	0.35 °C	
	Type C		
	(0 to 250) °C	0.24 °C	
	(250 to 1 000) °C	0.19 °C	
	(1 000 to 1 500) °C	0.21 °C	
	(1 500 to 1 800) °C	0.24 °C	
	(1 800 to 2 000) °C	0.27 °C	
	(2 000 to 2 250) °C	0.33 °C	
	(2 250 to 2 315) °C	0.37 °C	
	Type E		
	(-270 to -245) °C	1.6 °C	
	(-245 to -195) °C	0.24 °C	
	(-195 to -155) °C	0.12 °C	
	(-155 to -90) °C	0.09 °C	
	(-90 to 0) °C	0.08 °C	
	(0 to 15) °C	0.08 °C	
	(15 to 890) °C	0.06 °C	
	(890 to 1 000) °C	0.07 °C	
	Type J		
	(-210 to -180) °C	0.13 °C	
	(-180 to -120) °C	0.11 °C	
	(-120 to -50) °C	0.09 °C	
	(-50 to 990) °C	0.08 °C	
	(990 to 1 200) °C	0.08 °C	
	Type K		
	(-270 to -255) °C	2.5 °C	
	(-255 to -195) °C	0.85 °C	
	(-195 to -115) °C	0.16 °C	
	(-115 to -55) °C	0.12 °C	
	(-55 to 1 000) °C	0.09 °C	
	(1 000 to 1 372) °C	0.1 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹	Type N		Ectron 1140A Thermocouple Calibrator/Simulator; Direct Measure
	(-270 to -260) °C	5.4 °C	
	(-260 to -200) °C	1.5 °C	
	(-200 to -140) °C	0.29 °C	
	(-140 to -70) °C	0.18 °C	
	(-70 to 25) °C	0.14 °C	
	(-25 to 160) °C	0.12 °C	
	(160 to 1 300) °C	0.11 °C	
	Type R		
	(-50 to -30) °C	0.8 °C	
	(-30 to 45) °C	0.69 °C	
	(45 to 160) °C	0.49 °C	
	(160 to 380) °C	0.35 °C	
	(380 to 775) °C	0.3 °C	
	(775 to 1 768) °C	0.26 °C	
	Type S		
	(-50 to -30) °C	0.76 °C	
	(-30 to 45) °C	0.68 °C	
	(45 to 105) °C	0.49 °C	
	(105 to 310) °C	0.31 °C	
	(310 to 615) °C	0.35 °C	
	(615 to 1 768) °C	0.31 °C	
	Type T		
	(-270 to -255) °C	1.9 °C	
	(-255 to -240) °C	0.6 °C	
	(-240 to -210) °C	0.36 °C	
	(-210 to -150) °C	0.22 °C	
	(-150 to -40) °C	0.15 °C	
	(-40 to 100) °C	0.09 °C	
	(100 to 400) °C	0.08 °C	
DC Power – Source ¹			Fluke 5520A Multiproduct Calibrator; Direct Measure
(0.33 to 330) mA	(11 to 330) W	0.018 % of reading	
(0.33 to 3) A	11 W to 3 kW	0.017 % of reading	
(3 to 20.5) A	99 mW to 20.9 kW	0.054 % of reading	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Source ^{1,2} PF = 1 3.3 mA to 3 A 3.3 mA to 20.5 A 33 mA to 3 A 33 mA to 20.5 A (3 to 20.5) A	(10 to 45) Hz 0.11 mW to 99 W (45 to 65) Hz 0.11 mW to 20.9 kW (65 to 500) Hz 11 mW to 3.06 kW 500 Hz to 1 kHz 11 mW to 20.9 kW (65 to 500) Hz 9.9 W to 20.9 kW	0.18 % of reading 0.14 % of reading 0.16 % of reading 0.17 % of reading 0.16 % of reading	Fluke 5520A Multiproduct Calibrator; Direct Measure
Phase – Source ¹	Up to 180° (10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 20) kHz	0.092° 0.2° 0.39° 1.9° 3.9° 7.8°	Fluke 5520A Multiproduct Calibrator; Direct Measure
Oscilloscopes ^{1,3} Amplitude – DC into 50 Ω load into 1 MΩ load Amplitude – Square Wave Rate: 10 Hz to 10 kHz into 50 Ω load into 1 MΩ load Rate: 10 Hz to 100 kHz into 50 Ω load into 1 MΩ load	(-5 to 5) V (-200 to 200) V 40 μVp-p to 1 mVp-p 1 mVp-p to 5 Vp-p 40 μVp-p to 1 mVp-p 1 mVp-p to 5 Vp-p 1 mVp-p to 200 Vp-p	0.023 % of reading + 19 μV 0.023 % of reading + 19 μV 0.78 % of reading + 7.8 μV 0.078 % of reading + 7.8 μV 0.78 % of reading + 7.8 μV 0.16 % of reading + 7.8 μV 0.78 % of reading + 7.8 μV	Fluke 9500B/3200 3.2 GHz High-Performance Oscilloscope Calibrator; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,3} Time Markers 100 mVp-p to 1 Vp-p into 50 Ω load Square Wave Sine Wave Pulse Triangle Wave Rise Time into 50 Ω load Rate: 10 Hz to 2 MHz Rate: 10 Hz to 1 MHz	9.009 1 ns to 83 μ s 83 μ s to 55s 450.5 ps to 9.009 ns 900.91 ns to 83 μ s 83 μ s to 55s 900.91 ns to 83 μ s 83 μ s to 55s 5 mVp-p to 3 Vp-p 500 ps (nominal) 150 ps (nominal) 25 mVp-p to 2 Vp-p 70 ps (nominal) 425 mVp-p to 575 mVp-p 25 ps (nominal)	0.19 μ s/s 2.3 μ s/s 0.19 μ s/s 0.19 μ s/s 2.3 μ s/s 0.19 μ s/s 2.3 μ s/s 290 ps 35 ps 24 ps 6.7 ps	Fluke 9500B/3200 3.2 GHz High-Performance Oscilloscope Calibrator; Direct Measure
Oscilloscopes ^{1,3} Leveled Sine Wave 50 kHz Reference into 50 Ω load Input Impedance Measure Input Capacitance Measure	5 mVp-p to 5 Vp-p 50 kHz to 10 MHz (10 to 40) Ω (40 to 90) Ω (90 to 150) Ω (50 to 800) k Ω (0.8 to 1.2) M Ω (1.2 to 12) M Ω (1 to 35) pF (35 to 95) pF	1.2 % of reading 0.39 % of reading 0.083 % of reading 0.39 % of reading 0.39 % of reading 0.083 % of reading 0.39 % of reading 1.6 % of reading + 0.19 pF 2.3 % of reading + 0.19 pF	Fluke 9500B/3200 3.2 GHz High-Performance Oscilloscope Calibrator, Fluke 9530 3.2 GHz Active Head, Fluke 9550 Active Head w/ 25 ps Capability, Fluke 9560 6 GHz Active Head, Tektronix 067-1330-000 Calibration Fixture; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Bandwidth Flatness Measure ¹ into VSWR (1.2:1) (wrt Reference Frequency)	5 mVp-p to 5 Vp-p 100 MHz to 300 MHz (300 to 550) MHz 5 mVp-p to 3 Vp-p 550 MHz to 1.1 GHz (1.1 to 2.5) GHz 5 mVp-p to 2 Vp-p (2.5 to 3.2) GHz	1.6 % of reading 1.9 % of reading 2.7 % of reading 3.1 % of reading 3.1 % of reading	Fluke 9500B/3200 3.2 GHz High-Performance Oscilloscope Calibrator, Fluke 9530 3.2 GHz Active Head; Direct Measure
Total Harmonic Distortion – Measure	(-80 to 0) dB 20 Hz to 20 kHz (20 to 100) kHz	1.1 dB 2 dB	Agilent 8903A Audio Analyzer; Direct Measure
Total Harmonic Distortion – Measure (0.3 to 100) % 0.1 %	< 30 V 10 Hz to 1 MHz (1 to 3) MHz > 30 V 10 Hz to 300 kHz (300 to 500) kHz 500 kHz to 3 MHz < 30 V (10 to 20) Hz (20 to 30) Hz 30 Hz to 300 kHz (300 to 500) kHz 500 kHz to 1.2 MHz	3 % of reading 6 % of reading 3 % of reading 6 % of reading 12 % of reading 12 % of reading 6 % of reading 3 % of reading 6 % of reading 12 % of reading	HP 334A Distortion Analyzer; Direct Measure
Total Harmonic Distortion – Measure 0.1 %	> 30 V (20 to 30) Hz 30 Hz to 300 kHz (300 to 500) kHz 500 kHz to 1.2 MHz	12 % of reading 3 % of reading 6 % of reading 12 % of reading	HP 334A Distortion Analyzer; Direct Measure
Rise Time – Measure	≥ 350 ps	33 ps	Tektronix TDS874D Oscilloscope; Direct Measure

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rise Time – Measure	≥ 17.5 ps	12 ps	Agilent 83484A Dual Channel 50 GHz Electrical Module, Agilent 86100C Wideband Oscilloscope Mainframe; Direct Measure
Sine Wave Flatness	(0.1 to 3.9) V 20 Hz to 1 MHz (1 to 10) MHz (10 to 30) MHz (30 to 50) MHz (50 to 80) MHz (80 to 100) MHz	0.067 % of reading 0.078 % of reading 0.11 % of reading 0.28 % of reading 0.49 % of reading 0.61 % of reading	Thermal Converters, HP 3458A 8.5 Digit Multimeter; Direct Measure

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Power Meter Reference	50 MHz 1 mW Reference	0.43 % of reading	HP 478A Coaxial Thermistor Mount, HP 432A Power Meter; Direct Measure
S11/S22 Reflection Coefficients Magnitude – Measure ^{1,7} (Linear)	(10 to 700) MHz ≤ 0.25 lin (> 0.25 to 0.5) lin (> 0.5 to ≤ 0.7) lin (> 0.7 to ≤ 1) lin 700 MHz to 24 GHz ≤ 0.25 lin (> 0.25 to 0.5) lin (> 0.5 to ≤ 0.7) lin (> 0.7 to ≤ 1) lin	0.012 0.013 0.014 0.016 0.006 2 0.006 6 0.076 0.097	R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
S11/S22 Reflection Coefficients Magnitude – Measure ^{1,7} (Linear)	(24 to 33) GHz ≤ 0.25 lin (> 0.25 to 0.5) lin (> 0.5 to ≤ 0.7) lin (> 0.7 to ≤ 1) lin (33 to 50) GHz ≤ 0.25 lin (> 0.25 to 0.5) lin (> 0.5 to ≤ 0.7) lin (> 0.7 to ≤ 1) lin	0.019 0.019 0.022 0.023 0.019 0.019 0.022 0.023	R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure
S11/S22 Reflection Coefficients Phase – Measure ¹	(10 to 700) MHz ≤ 0.25 lin (> 0.25 to 0.5) lin (> 0.5 to ≤ 0.7) lin (> 0.7 to ≤ 1) lin	4° 2° 2° 1°	R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure
S11/S22 Reflection Coefficients Phase – Measure ¹ (Linear)	700 MHz to 24 GHz ≤ 0.25 lin (> 0.25 to 0.5) lin (> 0.5 to ≤ 0.7) lin (> 0.7 to ≤ 1) lin (24 to 50) GHz ≤ 0.25 lin (> 0.25 to 0.5) lin (> 0.5 to ≤ 0.7) lin (> 0.7 to ≤ 1) lin	4° 1° 1° 1° 11° 3° 2° 2°	R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure
S21/S12 Transmission Coefficients Magnitude – Measure ¹ (dB)	(10 to 700) MHz (-90 to ≤ -80) dB (-80 to ≤ -70) dB (-70 to ≤ -60) dB (-60 to ≤ -50) dB (-50 to ≤ -40) dB (-40 to ≤ -30) dB (-30 to ≤ -20) dB (-20 to ≤ -10) dB (-10 to ≤ 0) dB	0.56 dB 0.21 dB 0.13 dB 0.12 dB 0.12 dB 0.12 dB 0.12 dB 0.12 dB 0.12 dB	R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
S21/S12 Transmission Coefficients Magnitude – Measure ¹ (dB)	700 MHz to 24 GHz		R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure
	(-80 to ≤ -70) dB	0.55 dB	
	(-70 to ≤ -60) dB	0.19 dB	
	(-60 to ≤ -50) dB	0.08 dB	
	(-50 to ≤ -40) dB	0.06 dB	
	(-40 to ≤ -30) dB	0.05 dB	
	(-30 to ≤ -20) dB	0.05 dB	
	(-20 to ≤ -10) dB	0.05 dB	
	(-10 to ≤ 0) dB	0.06 dB	
	(24 to 33) GHz		
	(-80 to ≤ -70) dB	0.56 dB	
	(-70 to ≤ -60) dB	0.21 dB	
	(-60 to ≤ -50) dB	0.11 dB	
	(-50 to ≤ -40) dB	0.1 dB	
	(-40 to ≤ -30) dB	0.1 dB	
	(-30 to ≤ -20) dB	0.1 dB	
	(-20 to ≤ -10) dB	0.1 dB	
	(-10 to ≤ 0) dB	0.1 dB	
	(33 to 50) GHz		
	(-80 to ≤ -70) dB	0.56 dB	
	(-70 to ≤ -60) dB	0.21 dB	
	(-60 to ≤ -50) dB	0.11 dB	
	(-50 to ≤ -40) dB	0.1 dB	
	(-40 to ≤ -30) dB	0.1 dB	
	(-30 to ≤ -20) dB	0.1 dB	
	(-20 to ≤ -10) dB	0.1 dB	
	(-10 to ≤ 0) dB	0.1 dB	
S21/S12 Transmission Coefficients Phase – Measure ¹ (dB)	(10 to 700) MHz		R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure
	(-90 to ≤ -40) dB	180°	
	(-40 to ≤ -30) dB	52°	
	(-30 to ≤ -20) dB	14°	
	(-20 to ≤ -10) dB	4.5°	
	(-10 to ≤ 0) dB	1°	
	700 MHz to 24 GHz		
	(-80 to ≤ -40) dB	180°	
	(-40 to ≤ -30) dB	23°	
	(-30 to ≤ -20) dB	7.2°	
	(-20 to ≤ -10) dB	2.6°	
	(-10 to ≤ 0) dB	0.6°	

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
S21/S12 Transmission Coefficients Phase – Measure ¹ (dB)	(24 to 33) GHz (-80 to ≤ -40) dB (-40 to ≤ -30) dB (-30 to ≤ -20) dB (-20 to ≤ -10) dB (-10 to ≤ 0) dB (33 to 50) GHz (-80 to ≤ -40) dB (-40 to ≤ -30) dB (-30 to ≤ -20) dB (-20 to ≤ -10) dB (-10 to ≤ 0) dB	180° 52° 14° 4.5° 1° 180° 52° 14° 4.5° 1°	R&S® ZVA50 Vector Network Analyzer, Calibration Kits; Direct Measure
Absolute RF Power – Measure ⁸	8 kHz to 18 GHz (-30 to 23) dBm	0.18 dB	R&S® NRP18A Power Sensor, NRX Power Meter; Direct Measure
Absolute RF Power – Measure ⁸	(-35 to 20) dBm DC to 100 MHz > 100 MHz to 2.4 GHz (> 2.4 to 12.4) GHz (> 12.4 to 18) GHz (>18 to 26.5) GHz (>26.5 to 40) GHz (>40 to 50) GHz	0.08 dB 0.08 dB 0.09 dB 0.1 dB 0.11 dB 0.13 dB 0.17 dB	R&S® NRP18A R&S® NRP50T Power Sensors, NRX Power Meter; Direct Measure
Relative RF Power – Measure ⁸	(-30 to 20) dBm DC to 50 GHz	0.04 dB	R&S® NRP18T R&S® NRP50T Power Sensors, NRX Power Meter; Direct Measure
Amplitude Modulation – AM Depth Measure (Absolute) Rate: 10 Hz to 150 kHz	100 kHz to 50 GHz Up to 100 % Depth	0.5 % of reading	R&S® FSMR Measuring Receiver (B2, B4, B24); Direct Measure
Amplitude Modulation – AM Depth Measure (Flatness reference to 1 kHz) Rate: 10 Hz to 150 kHz	100 kHz to 50 GHz Up to 100 % Depth	0.3 % of reading	R&S® FSMR Measuring Receiver (B2, B4, B24); Direct Measure

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation – FM Deviation Measure Rate: 10 Hz to 5 MHz	100 kHz to 50 GHz Dev $\leq$ 5 MHz	1.2 % of reading	R&S® FSMR Measuring Receiver (B2, B4, B24); Direct Measure
Phase Modulation – Deviation Measure Rate: 10 Hz to 5 MHz	100 kHz to 50 GHz Dev $\leq$ 10 000 rad	1.2 % of reading	R&S® FSMR Measuring Receiver (B2, B4, B24); Direct Measure
Single-sideband Phase Noise – Source/Measure ⁸	1 MHz to 50 GHz 10 mHz $\leq$ Offset $<$ 1 MHz 1 MHz $\leq$ Offset $\leq$ 30 MHz Offset $>$ 30 MHz	1.8 dB 2.3 dB 3.5 dB	R&S® FSWP50 Phase Noise Analyzer (B1, B24, B320, K70); Direct Measure
Adjacent Channel Leakage Ratio (ACLR) ⁸ Signal Noise $>$ 16 dB	DC to 50 GHz (-90 to -70) dB (-70 to 0) dB	0.18 dB 0.14 dB	R&S® FSWP50 Phase Noise Analyzer; Direct Measure
Error Vector Magnitude (EVM) Rate: Up to 320 MHz	100 kHz to 6 GHz FSK/ASK/PSK/APSK/ MSK/16QAM/64QAM/ 128QAM/256QAM/ 1024QAM	2.3 % of reading	R&S® FSWP50 Phase Noise Analyzer; Direct Measure
Relative Tuned RF Power – Measure ¹	100 kHz to 22 GHz (-120 to -110) dBm (-110 to -100) dBm (-100 to -90) dBm (-90 to -80) dBm (-80 to -70) dBm (-70 to -60) dBm (-60 to -50) dBm (-50 to -40) dBm (-40 to -30) dBm (-30 to -20) dBm (-20 to -10) dBm (-10 to 0) dBm	0.67 dB 0.22 dB 0.13 dB 0.12 dB 0.11 dB 0.11 dB 0.11 dB 0.11 dB 0.088 dB 0.086 dB 0.084 dB 0.083 dB	R&S® FSMR Measuring Receiver; Direct Measure

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Tuned RF Power – Measure ¹	(22 to 40) GHz		R&S® FSMR Measuring Receiver; Direct Measure
	(-120 to -110) dBm	0.68 dB	
	(-110 to -100) dBm	0.23 dB	
	(-100 to -90) dBm	0.14 dB	
	(-90 to -80) dBm	0.13 dB	
	(-80 to -70) dBm	0.13 dB	
	(-70 to -60) dBm	0.12 dB	
	(-60 to -50) dBm	0.12 dB	
	(-50 to -40) dBm	0.12 dB	
	(-40 to -30) dBm	0.092 dB	
	(-30 to -20) dBm	0.09 dB	
	(-20 to -10) dBm	0.088 dB	
	(-10 to 0) dBm	0.086 dB	
	(40 to 50) GHz		
	(-120 to -110) dBm	0.69 dB	
	(-110 to -100) dBm	0.26 dB	
	(-100 to -90) dBm	0.19 dB	
	(-90 to -80) dBm	0.18 dB	
	(-80 to -70) dBm	0.17 dB	
	(-70 to -60) dBm	0.17 dB	
	(-60 to -50) dBm	0.16 dB	
	(-50 to -40) dBm	0.16 dB	
	(-40 to -30) dBm	0.11 dB	
	(-30 to -20) dBm	0.11 dB	
	(-20 to -10) dBm	0.1 dB	
	(-10 to 0) dBm	0.1 dB	
Absolute Tuned RF Power – Measure ¹	100 kHz to 22 GHz		R&S® FSMR Measuring Receiver; R&S® NRP18T R&S® NRP50T Power Sensors, NRX Power Meter; Direct Measure
	(-120 to -110) dBm	0.68 dB	
	(-110 to -100) dBm	0.25 dB	
	(-100 to -90) dBm	0.17 dB	
	(-90 to -80) dBm	0.16 dB	
	(-80 to -70) dBm	0.16 dB	
	(-70 to -60) dBm	0.16 dB	
	(-60 to -50) dBm	0.16 dB	
	(-50 to -40) dBm	0.16 dB	
	(-40 to -30) dBm	0.14 dB	
	(-30 to -20) dBm	0.14 dB	
	(-20 to -10) dBm	0.14 dB	
	(-10 to 0) dBm	0.14 dB	

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Absolute Tuned RF Power – Measure ¹	(22 to 40) GHz		R&S® FSMR Measuring Receiver; R&S® NRP18T R&S® NRP50T Power Sensors, NRX Power Meter; Direct Measure
	(-120 to -110) dBm	0.69 dB	
	(-110 to -100) dBm	0.26 dB	
	(-100 to -90) dBm	0.19 dB	
	(-90 to -80) dBm	0.18 dB	
	(-80 to -70) dBm	0.18 dB	
	(-70 to -60) dBm	0.18 dB	
	(-60 to -50) dBm	0.18 dB	
	(-50 to -40) dBm	0.18 dB	
	(-40 to -30) dBm	0.16 dB	
	(-30 to -20) dBm	0.16 dB	
	(-20 to -10) dBm	0.16 dB	
	(-10 to 0) dBm	0.16 dB	
	(40 to 50) GHz		
	(-120 to -110) dBm	0.71 dB	
	(-110 to -100) dBm	0.31 dB	
	(-100 to -90) dBm	0.25 dB	
	(-90 to -80) dBm	0.25 dB	
	(-80 to -70) dBm	0.24 dB	
	(-70 to -60) dBm	0.24 dB	
	(-60 to -50) dBm	0.23 dB	
	(-50 to -40) dBm	0.23 dB	
	(-40 to -30) dBm	0.2 dB	
	(-30 to -20) dBm	0.2 dB	
	(-20 to -10) dBm	0.2 dB	
	(-10 to 0) dBm	0.2 dB	
Absolute Tuned RF Power – Measure	2.5 MHz to 26.5 GHz		HP 8902A Opt. 050 Measuring Receiver; HP 11722A, HP 11792A, HP 11793A Power Sensors; Direct Measure
	(-127 to -120) dB	0.26 dB	
	(-120 to -110) dB	0.26 dB	
	(-110 to -100) dB	0.26 dB	
	(-100 to -90) dB	0.26 dB	
	(-90 to -80) dB	0.26 dB	
	(-80 to -70) dB	0.25 dB	
	(-70 to -60) dB	0.25 dB	
	(-60 to -50) dB	0.25 dB	
	(-50 to -40) dB	0.25 dB	
	(-40 to -30) dB	0.14 dB	
	(-30 to -20) dB	0.14 dB	
	(-20 to -10) dB	0.14 dB	
	(-10 to -0) dB	0.14 dB	

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Tuned RF Power – Measure	2.5 MHz to 26.5 GHz		HP 8902A Opt. 050 Measuring Receiver; HP 11722A, HP 11792A, HP 11793A Power Sensors; Direct Measure
	(-127 to -120) dB	0.23 dB	
	(-120 to -110) dB	0.23 dB	
	(-110 to -100) dB	0.23 dB	
	(-100 to -90) dB	0.23 dB	
	(-90 to -80) dB	0.22 dB	
	(-80 to -70) dB	0.084 dB	
	(-70 to -60) dB	0.081 dB	
	(-60 to -50) dB	0.074 dB	
	(-50 to -40) dB	0.071 dB	
	(-40 to -30) dB	0.068 dB	
	(-30 to -20) dB	0.064 dB	
	(-20 to -10) dB	0.06 dB	
	(-10 to -0) dB	0.056 dB	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Angle Measuring Devices, Protractors, Inclinometers, Squares, Angle Gages ⁴	(0.005 6 to 5)°	3.2"	Direct Measure; 5 in Sine Bar, Gage Blocks, Surface Plate
	(5 to 20)°	6.1"	
	(20 to 35)°	11"	
	(35 to 45)°	15"	
	(45 to 60)°	25"	
	(60 to 75)°	54"	
	(75 to 85)°	166"	
	90°	1.1"	Master Square, Surface Plate;
Calipers, Micrometers ^{1,4} Travel (Outside, Inside, Depth, Step)	Up to 1 in (1 to 9) in	(16 + 1L) μin (11 + 4L) μin	Direct Measure; B89.1 Grade 00 Gage Blocks
	(4 to 15) in (15 to 40) in	(13 + 4.5L) μin (15 + 4.6L) μin	B89.1 Grade 0 Gage Blocks
Anvil Flatness	Up to 1 in diameter	4.4 μin	Optical Flats
Anvil Parallelism	Up to 1 in diameter	8.2 μin	Optical Parallels

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dial/Digital Indicators	Up to 1 in (1 to 6) in	17 μ in (12 + 4L) μ in	B89.1 Grade 00 Gage Blocks; Direct Measure
Length – Single Axis ⁴ Outside Dimension	Up to 1 in (1 to 7) in (7 to 12) in	(6 + 1L) μ in (4.3 + 3.5L) μ in (1 + 4L) μ in	Universal Length Measuring Machine; Direct Measure
Length – Single Axis ⁴ Inside Dimension	(0.04 to 0.125) in (0.125 to 0.25) in (0.25 to 1) in (1 to 2.5) in (2.5 to 10) in	17 μ in 17 μ in 11 μ in (14 + 2L) μ in (18 + 3L) μ in	Universal Length Measuring Machine; Direct Measure
Height Gages ^{1,4}	(0.05 to 1) in (1 to 5) in (5 to 40) in	(13 + 1L) μ in (9 + 4L) μ in (10 + 5L) μ in	Gage Blocks, Surface Plate; Direct Measure
Parallelism, Flatness, Straightness	Up to 36 in	45 μ in	Gage Amplifier, Surface Plate; Direct Measure
Squareness	Up to 18 in	9.4 μ in/in	Gage Amplifier, Granite Master Square; Direct Measure
Cylindrical Plug Gages ⁴ Outside Diameter	Up to 1 in (1 to 7) in	12 μ in (9 + 3L) μ in	Universal Length Measuring Machine; Direct Measure
Pin Gages (Outside Diameter)	0.01 to 1 in	31 μ in	Non-contact Method using Laser Micrometer
Cylindrical Ring Gages ⁴ Inside Diameter	Up to 2.5 in (2.5 to 10) in (10 to 14) in	11 μ in (18 + 3L) μ in (38 + 3L) μ in	Universal Length Measuring Machine; Direct Measure
Thread Plug Gages ⁴ Pitch Diameter	Up to 1 in (1 to 7) in (7 to 12) in	79 μ in 80 μ in 83 μ in	Direct Measure; Universal Length Measuring Machine, Thread Wires
Major Diameter	Up to 1 in (1 to 7) in	13 μ in (10 + 3L) μ in	Universal Length Measuring Machine

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Ring Gages ⁴ Inner Pitch Diameter	Up to 1 in (1 to 4) in (4 to 7) in	79 μ in 80 μ in 83 μ in	Comparison to Master Thread Setting Plug and using the Measurement Uncertainty from that Master.
Thread Wires (2 to 120) TPI	(0.008 to 0.5) in	12 μ in	Universal Length Measuring Machine; Direct Measure
Measuring Tapes, Rulers ^{1,4}	Up to 40 in	(600 + 8L) μ in	Comparison to Glass Rule
Measuring Tapes, Rulers ⁴	Up to 1 ft (1 to 3) ft (3 to 1 000) ft	(463 + 2L) μ in (410 + 6L) μ in (18L) μ in	Single Axis Vision System; Direct Measure
Optical Comparators ^{1,4} X, Y Length	Up to 12 in	(90 + 5L) μ in	Direct Measure; Calibration Grids
Magnification	10X to 50X	200 μ in	Magnification Checker
Perpendicularity of X-Y axis	Up to 6 in	58 μ in/in	Precision Square

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gas Flow Devices	(2 to 200) sccm (0.2 to 40) slpm (40 to 80) slpm (80 to 100) slpm	0.33 % of reading 0.23 % of reading 0.32 % of reading 0.57 % of reading	Comparison to Fluke molbloc Laminar Flow Element Gas Flow Calibration System
Gas Flow Devices	(100 to 300) slpm (300 to 1 200) slpm	0.6 % of reading 0.81 % of reading	Comparison to Fluke molbloc Sonic Nozzle Gas Flow Calibration System
Force Gages (Tension and Compression)	Up to 5 lbf (5 to 10) lbf (10 to 20) lbf (20 to 30) lbf (30 to 100) lbf	0.001 1 lbf 0.002 lbf 0.005 8 lbf 0.009 3 lbf 0.061 lbf	Characterized NIST Class F Weights; Direct Measure

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force Gages (Tension and Compression)	Up to 5 lbf (5 to 10) lbf (10 to 20) lbf (20 to 30) lbf (30 to 100) lbf	0.001 2 lbf 0.002 3 lbf 0.006 2 lbf 0.009 9 lbf 0.063 lbf	NIST Class F Weights; Direct Measure
Mass Determination Metric (SI)	1 mg 2 mg 3 mg 5 mg 10 mg 20 mg 30 mg 50 mg 100 mg 200 mg 300 mg 500 mg 1 g 2 g 3 g 5 g 10 g 20 g 30 g 50 g 100 g 200 g 300 g 500 g 1 kg 2 kg 3 kg 5 kg 10 kg 20 kg 25 kg	1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 1.9 µg 4.2 µg 4.4 µg 6 µg 5.2 µg 7.6 µg 11.1 µg 13.6 µg 16.5 µg 33 µg 70 µg 74 µg 0.16 mg 0.33 mg 0.73 mg 1 mg 1.6 mg 3.6 mg 7.2 mg 9.9 mg	Double Substitution Method; Reference Weights, Electronic Balances

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination Metric (SI)	1 mg	4.5 µg	Single Substitution Method; Reference Weights, Electronic Balances
	2 mg	4.5 µg	
	3 mg	4.5 µg	
	5 mg	4.5 µg	
	10 mg	4.5 µg	
	20 mg	4.5 µg	
	30 mg	4.5 µg	
	50 mg	4.5 µg	
	100 mg	4.5 µg	
	200 mg	4.5 µg	
	300 mg	4.5 µg	
	500 mg	4.5 µg	
	1 g	11.8 µg	
	2 g	12 µg	
	3 g	13.4 µg	
	5 g	12.4 µg	
	10 g	18.5 µg	
	20 g	27 µg	
	30 g	29 µg	
	50 g	43 µg	
	100 g	62 µg	
	200 g	0.18 mg	
	300 g	0.25 mg	
	500 g	0.42 mg	
	1 kg	0.9 mg	
	2 kg	1.9 mg	
	3 kg	2.9 mg	
	5 kg	4.4 mg	
	10 kg	6.9 mg	
	20 kg	19 mg	
	25 kg	25 mg	

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination Avoirdupois	0.0312 5 oz 0.062 5 oz 0.125 oz 0.25 oz 0.5 oz 1 oz 2 oz 4 oz 8 oz 1 lb 2 lb 5 lb 7.5 lb 10 lb 15 lb 20 lb 25 lb 50 lb	7.7 µg 14 µg 15 µg 18 µg 26 µg 34 µg 45 µg 65 µg 0.18 mg 0.26 mg 0.5 mg 1.9 mg 2.9 mg 3.1 mg 4.5 mg 5.4 mg 6.8 mg 19 mg	Single Substitution Method; Reference Weights, Electronic Balances
Rockwell Hardness Testers ¹	HRC Scale (23 to 27) HRC (43 to 47) HRC (63 to 67) HRC HRBw Scale (48 to 52) HRBw (68 to 72) HRBw (93 to 97) HRBw	0.87 HRC 0.66 HRC 0.49 HRC 1.24 HRBw 1.2 HRBw 1.21 HRBw	Indirect verification per ASTM E18 using Hardness Test Blocks.
Durometers Spring Force Types A, B, E, O Type D, C, DO Indicator Geometry Length Angle Radius	Up to 100 Duro Up to 100 Duro Up to 1 in Up to 180° Up to 1 in	0.31 Duro 0.16 Duro 210 µin 0.008° 260 µin	Direct Verification per ASTM D2240 using Duro Calibrator Optical Comparator
Torque Wrenches, Torque Drivers, Torque Indicators ¹	(3 to 80) ozf·in 15 ozf·in to 600 lbf·ft 5 lbf·in to 800 lbf·ft	1.7 % of reading 0.5 % of reading 1 % of reading	Torque Calibrators; Direct Measure

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Torque Angle ¹	45° 90° 135° 180° 360°	0.35° 0.35° 0.35° 0.35° 0.35°	Comparison to Torque Angle Fixture
Torque Calibration Equipment	5 ozf·in to 2.5 lbf·in (2.5 to 50) lbf·in	0.09 % of reading 0.05 % of reading	Torque Wheel, Master Weights; Direct Measure
Torque Calibration Equipment	50 lbf·in to 250 lbf·ft	0.06 % of reading	Torque Butterfly, Master Weights; Direct Measure
Balances and Scales ^{1,5} Metric (SI)	Up to 500 mg 500 mg to 2 g (2 to 5) g (5 to 10) g (10 to 20) g (20 to 50) g (50 to 100) g (100 to 200) g (200 to 500) g 500 g to 1 kg (1 to 2) kg (2 to 5) kg (5 to 10) kg (10 to 20) kg (20 to 30) kg (30 to 40) kg (40 to 50) kg (50 to 60) kg (60 to 70) kg (70 to 80) kg (80 to 100) kg	2.5 µg 6.2 µg 6.7 µg 11 µg 16 µg 27 µg 58 µg 0.11 mg 0.27 mg 0.54 mg 1.8 mg 3 mg 6.6 mg 14 mg 20 mg 27 mg 25 mg 26 mg 28 mg 29 mg 30 mg	Characterized ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system.
Balances and Scales ^{1,5} Metric (SI)	Up to 5 g (5 to 10) g (10 to 20) g (20 to 30) g (30 to 50) g (50 to 100) g	32 µg 44 µg 59 µg 88 µg 0.15 mg 0.29 mg	Characterized ASTM E617 Class 2 weights and internal calibration procedure utilized for the calibration of the weighing system.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Balances and Scales ^{1,5} Metric (SI)	(100 to 200) g (200 to 300) g (300 to 500) g (500 to 1 000) g (1 to 2) kg (2 to 3) kg (3 to 5) kg (5 to 6) kg (6 to 7) kg (7 to 8) kg (8 to 9) kg (9 to 10) kg	0.58 mg 0.89 mg 1.5 mg 3 mg 6.2 mg 9 mg 15 mg 18 mg 21 mg 24 mg 27 mg 30 mg	Characterized ASTM E617 Class 2 weights and internal calibration procedure utilized for the calibration of the weighing system.
Balances and Scales ^{1,5} Metric (SI)	Up to 250 g (250 to 500) g 500 g to 750 kg	0.023 % of reading 0.017 % of reading 0.012 % of reading	NIST Class F weights and internal calibration procedure utilized for the calibration of the weighing system.
Balances and Scales ^{1,5} Avoirdupois	Up to 0.5 lb (0.5 to 1) lb (1 to 1 700) lb	0.023 % of reading 0.017 % of reading 0.012 % of reading	NIST Class F weights and internal calibration procedure utilized for the calibration of the weighing system.
Volumetric Devices	1 mL to 5 L	0.2 % of reading + 0.2 mL	Gravimetric method utilizing Balances.
Pneumatic Absolute Pressure	Up to 30 psia (30 to 300) psia (300 to 1 000) psia	0.002 6 psi 0.008 8 % of reading 0.01 % of reading	Comparison to Fluke RPM4 Reference Pressure Monitor with Pressure Source
Pneumatic Gauge Pressure	(-60 to -22) inH ₂ O (-22 to 22) inH ₂ O (22 to 60) inH ₂ O (60 to 72) inH ₂ O (72 to 804) inH ₂ O	0.009% of reading + 150 µinH ₂ O 0.002 2 inH ₂ O 0.009% of reading + 150 µinH ₂ O 0.006 7 inH ₂ O 0.009% of reading + 150 µinH ₂ O	Comparison to Fluke PPC4 Pressure Controller/Calibrator
Pneumatic Gauge Pressure	(-15 to 30) psig (30 to 300) psig (300 to 1000) psig	0.002 2 psi 0.007 5 % of reading 0.01 % of reading	Comparison to Fluke RPM4 Reference Pressure Monitor with Pressure Source.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Hydraulic Pressure ¹	(50 to 15 000) psia	0.011 % of reading + 0.002 6 psi	Comparison to Ametek T-150 Deadweight Tester, Fluke RPM4 Reference Pressure Monitor
Hydraulic Pressure ¹	(50 to 15 000) psig	0.011 % of reading	Ametek T-150 Deadweight Tester; Direct Measure

Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Illuminance At 2 856 k	(30 to 10 764) lx (10 764 to 21 258) lx (21 258 to 32 300) lx	1.1 % of reading 1.7 % of reading 2.1 % of reading	Comparison to Standard Lamp

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Measuring Equipment	(-10 to 15) °C (10 to 75) %RH (75 to 95) %RH (15 to 35) °C (10 to 95) %RH (35 to 70) °C (10 to 50) %RH (50 to 70) %RH (70 to 95) %RH	0.5 %RH 0.65 %RH 0.5 %RH 0.5 %RH 0.7 %RH 0.85 %RH	Humidity Generator; Direct Measure
Humidity – Measure ¹	(10 to 30) °C (10 to 90) %RH	1.3 %RH	Comparison to Vaisala HMI41/HMP46 Temp/Humidity Indicator/Probe
Temperature – Source (Thermocouple Probes, Digital Thermometers, etc.)	(-195 to 0) °C (0 to 150) °C (150 to 420) °C (420 to 650) °C	0.014 °C 0.001 % of reading + 0.024 °C 0.001 % of reading + 0.036 °C 0.001 % of reading + 0.078 °C	Comparison to SPRT, Hart Black Stack, Metrology Wells

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Source (Thermocouple Probes, Digital Thermometers, etc.)	(600 to 1 000) °C (1 000 to 1 200) °C	3 °C 4.1 °C	Comparison to Hart 5650 Type S Thermocouple Probe, Digital Multimeter
Temperature – Measure ¹ (Room Measurements, Heat Sources, etc.)	(-195 to 0) °C (0 to 420) °C (420 to 650) °C	0.012 °C 0.001 % of reading + 0.02 °C 0.001 % of reading + 0.028 °C	Comparison to SPRT, Hart Black Stack
Temperature – Measure ¹ (Heat Sources)	(600 to 1 000) °C (1 000 to 1450) °C	0.94 °C 2.9 °C	Comparison to Hart 5650 Type S Thermocouple Probe, Digital Multimeter
SPRT/PRT/RTD Calibration ⁶	-195 °C	3.4 mK	NBPLN ₂ , SPRT, Super Thermometer; Comparison Method
SPRT/PRT/RTD Calibration	-78 °C -38 °C 0 °C	3.2 mK 2.9 mK 2.9 mK	Precision Bath, SPRT, Super Thermometer; Comparison Method
SPRT/PRT/RTD Calibration	0.01 °C	1.5 mK	Triple Point of Water Cell; Direct Comparison Method
SPRT/PRT/RTD Calibration	100 °C 156 °C 231 °C 300 °C 420 °C	3.6 mK 4.6 mK 6 mK 5.8 mK 8.4 mK	Precision Bath, SPRT, Super Thermometer; Comparison Method
SPRT/PRT/RTD Calibration	(-100 to 0) °C (0 to 420) °C	8.9 mK 0.003 % of reading + 5.9 mK	Precision Bath, SPRT, Super Thermometer; Comparison Method
Infrared Temperature Measuring Equipment	(-15 to 0) °C (0 to 50) °C (50 to 100) °C (100 to 120) °C (120 to 200) °C (200 to 350) °C (350 to 500) °C	0.8 °C 0.65 °C 0.7 °C 0.76 °C 0.95 °C 1.6 °C 2.1 °C	Blackbody Sources (flat plate) $\epsilon = (0.1 \text{ to } 1)$, $\lambda = (8 \text{ to } 14) \mu\text{m}$; Direct Measure

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Reference	10 MHz	3.7 pHz/Hz	Comparison to Fluke 910R GPS Frequency Standard
Frequency – Measure ¹	(1 to 10) kHz 10 kHz to 10 MHz (10 to 225) MHz	19 pHz/Hz + 4.5 μ Hz 19 pHz/Hz + 18 μ Hz 19 pHz/Hz + 0.64 mHz	Comparison to Agilent 53132A Universal Counter, Fluke 910R GPS Frequency Standard
Frequency – Source ¹	1 Hz to 50 MHz	58 nHz/Hz	Agilent 33250A Function/Arbitrary Function Generator, Fluke 910R GPS Frequency Standard; Direct Measure
Period – Measure ¹	(1 to 100) s	45 μ s	Comparison to Agilent 53132A Universal Counter, Fluke 910R GPS Frequency Standard
Period – Source ¹	(1 to 100) s	58 ns/s	Agilent 33250A Function/Arbitrary Function Generator, Fluke 910R GPS Frequency Standard; Direct Measure
Stopwatches, Timers ¹	Up to 599 s/mon	58 ms/day	Vibrograf 4500 Timometer; Time Base Method
AC Duty Cycle – Source ¹ Square-wave < 3.3 Vp-p Freq: 0.1 Hz to 100 kHz	(1 to 10) % Duty Cycle 10 μ s to 100 s (10 to 49) % Duty Cycle 10 μ s to 100 s 50 % Duty Cycle 10 μ s to 100 s (51 to 90) % Duty Cycle 10 μ s to 100 s (91 to 99) % Duty Cycle 10 μ s to 100 s	0.62 % of reading + 78 ns 0.039 % of reading + 78 ns 0.001 6 % of reading + 78 ns 0.039 % of reading + 78 ns 0.62 % of reading + 78 ns	Fluke 5522A Multiproduct Calibrator; Direct Measure

DIMENSIONAL MEASUREMENT

1 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 1D	X-Axis		Optical Comparator utilized as Reference for Length Measurement Inspection; Customer Drawings, Specifications
	Up to 1 in	210 μ in	
	(1 to 3) in	370 μ in	
	(3 to 6) in	480 μ in	
	Y-Axis		
	Up to 2 in	360 μ in	
	(2 to 3) in	410 μ in	
	(3 to 5) in	560 μ in	


2 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 2D	Up to 180°	0.008°	Optical Comparator utilized as Reference for Angle Measurement Inspection; Customer Drawings, Specifications
Dimensional Measurement – 2D	Up to 1 in	260 μ in	Optical Comparator utilized as Reference for Radius Measurement Inspection; Customer Drawings, Specifications
	(1 to 3) in	450 μ in	
	(3 to 6) in	590 μ in	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact the laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
3. The stated uncertainty for rise time is the laboratory's ability to source a fast rise pulse that is approximately 500 ps, 125 ps, and 25 ps. In the typical application of measuring rise time of an oscilloscope, this value is one of the contributing factors, but other factors are derived from the DUT. The known source rise time is mathematically removed from the total measured rise time measured on the DUT.
4. L = length in inches; DL = diagonal length in inches; " = arc-second.
5. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
6. NBPLN₂ = Boiling Point of Liquid Nitrogen.
7. The Uncertainty for this measurand is a Unitless measure.
8. Mismatch due to the effects of the device-under-test (DUT) and instruments is not included in the Scope CMC but will be included in the Measurement Uncertainty (MU) on the calibration certificate.
9. Accuracies are based on non-constant voltage setting at (1 to 1.26) V, Slow Speed, and a Quality Factor of < 0.05 and a lab temperature of $23\text{ }^{\circ}\text{C} \pm 7\text{ }^{\circ}\text{C}$. Measurements taken at resistances below $10\text{ }\Omega$ and other voltages, frequencies and speeds require recalculation using the IET accuracy tool and will result in greater uncertainties.
10. Unless otherwise specified in the far-right column above, the laboratory utilizes internally written calibration procedures in the process of calibrating the parameters listed in this document.
11. The legal entity for this location is Transcat, Inc.
12. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.03.



Jason Stine, Vice President

