



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat - Portland
14036 SW Milton Court
Tigard, OR 97224

Fulfils the requirements of

ISO/IEC 17025:2017

and national standards

ANSI/NCSL Z540-1-1994 (R2002) AND
ANSI/NCSL Z540.3-2006 (R2013)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.

The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 07 September 2025
Certificate Number: AC-2489.01



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)
ANSI/NCSL Z540.3 (R2013)

Transcat - Portland
14036 SW Milton Court
Tigard, OR 97224
Michael Gettle 503-598-8700

CALIBRATION

Valid to: **September 7, 2025**

Certificate Number: **AC-2489.01**

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|---|---|
| Sine Wave Flatness ¹ | Up to 3 V (10 to 100) Hz 100 Hz to 300 kHz 300 kHz to 10 MHz (10 to 20) MHz (20 to 30) MHz (30 to 50) MHz (50 to 70) MHz (70 to 80) MHz (80 to 100) MHz | 0.07 % of reading 0.071 % of reading 0.13 % of reading 0.21 % of reading 0.22 % of reading 0.48 % of reading 0.75 % of reading 0.89 % of reading 1 % of reading | Comparison to Thermal Voltage Converters |
| DC Current – Source/Measure ¹ | Up to 100 μ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A | 33 μ A/A + 0.92 nA 29 μ A/A + 5.8 nA 29 μ A/A + 58 nA 46 μ A/A + 0.58 μ A 0.013 % of reading + 12 μ A | Comparison to Agilent 3458A Opt 02 8.5 Digit Multimeter, Current Source |
| DC Current – Source/Measure ¹ | (1 to 10) A (10 to 100) A (100 to 300) A | 0.013 % of reading 0.048 % of reading 0.062 % of reading | Comparison to Guildline 9211 DC Current Shunt, Current Source |
| DC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor ¹ | (20 to 150) A (150 to 1 000) A | 0.5 % of reading + 0.14 A 0.52 % of reading + 0.5 A | Comparison to Fluke 5520A Multiproduct Calibrator, Wavetek Coil |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|--|---|
| DC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor ¹ | (1 000 to 5 000) A | 0.58 % of reading | Comparison to Fluke 52120A Transconductance Amplifier, Fluke 5520A Multiproduct Calibrator, 3 kA or 6 kA Coil |
| AC Current – Measure ¹ | Up to 100 μ A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 1 kHz (0.1 to 1) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (0.1 to 1) A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz | 0.46 % of reading + 35 nA 0.17 % of reading + 35 nA 0.072 % of reading + 35 nA 0.072 % of reading + 35 nA 0.46 % of reading + 0.23 μ A 0.17 % of reading + 0.23 μ A 0.07 % of reading + 0.23 μ A 0.038 % of reading + 0.23 μ A 0.46 % of reading + 2.3 μ A 0.17 % of reading + 2.3 μ A 0.071 % of reading + 2.3 μ A 0.038 % of reading + 2.3 μ A 0.48 % of reading + 23 μ A 0.17 % of reading + 23 μ A 0.071 % of reading + 23 μ A 0.037 % of reading + 23 μ A 0.46 % of reading + 0.23 mA 0.19 % of reading + 0.23 mA 0.097 % of reading + 0.23 mA 0.12 % of reading + 0.23 mA | Comparison to Agilent 3458A Opt 02 8.5 Digit Multimeter |
| AC Current – Measure ¹ | (1 to 3) A 10 Hz to 5 kHz (3 to 10) A 10 Hz to 1 kHz | 0.17 % of reading + 1.8 mA 0.18 % of reading + 6 mA | Comparison to Fluke 8846A 6.5 Digit Multimeter |
| AC Current – Measure ¹ | (10 to 100) A 10 Hz to 1 kHz | 0.12 % of reading | Comparison to Ohms Labs Current Shunt, Digital Multimeter |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------------|---|---|--|
| AC Current – Source ¹ | Up to 220 μ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz | 0.031 % of reading + 16 nA 0.019 % of reading + 10 nA 0.015 % of reading + 8 nA 0.03 % of reading + 12 nA 0.11 % of reading + 65 nA 0.03 % of reading + 40 nA 0.018 % of reading + 35 nA 0.013 % of reading + 35 nA 0.021 % of reading + 0.11 μ A 0.11 % of reading + 0.65 μ A 0.039 % of reading + 0.4 μ A 0.019 % of reading + 0.35 μ A 0.014 % of reading + 0.35 μ A 0.021 % of reading + 0.55 μ A 0.11 % of reading + 5 μ A 0.033 % of reading + 4 μ A 0.018 % of reading + 3.5 μ A 0.014 % of reading + 2.5 μ A 0.021 % of reading + 3.5 μ A 0.11 % of reading + 10 μ A 0.027 % of reading + 35 μ A 0.046 % of reading + 80 μ A 0.7 % of reading + 0.16 mA | Comparison to Fluke 5720A Multiproduct Calibrator |
| AC Current – Source ¹ | (2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz | 0.048 % of reading + 0.17 mA 0.096 % of reading + 0.38 mA 0.36 % of reading + 0.75 mA | Comparison to Fluke 5720A Multiproduct Calibrator, Fluke 5725A Amplifier |
| AC Current – Source ¹ | (11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz | 0.09 % of reading + 3.9 mA 0.12 % of reading + 3.9 mA 2.3 % of reading + 3.9 mA | Comparison to Fluke 5520 A Multiproduct Calibrator |
| AC Current – Source ¹ | (20 to 100) A 10 Hz to 1 kHz | 0.12 % of reading | Comparison to Ohms Labs Current Shunt, Current Source |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|---|---|
| AC Current – Source ¹ | Up to 10 A 50 Hz to 1 kHz 1 kHz (10 to 100) A (50 to 100) Hz (100 to 999) Hz | 0.05 % of reading + 1.3 mA 0.12 % of reading + 1.3 mA 0.04 % of reading + 2.3 mA 0.42 % of reading + 2.3 mA | Comparison to Ohms Labs CS-100 Precision Shunt w/ Agilent 3458A Opt 02 Multimeter and Source |
| AC Current – Source ¹ Extended Frequency Ranges | (29 to 330) μ A (10 to 30) kHz (0.33 to 3.3) mA (10 to 30) kHz (3.3 to 33) mA (10 to 30) kHz (33 to 330) mA (10 to 30) kHz | 1.2 % of reading + 3 μ A 0.78 % of reading + 0.5 μ A 0.31 % of reading + 3 μ A 0.31 % of reading + 0.16 mA | Comparison to Fluke 5520A Multiproduct Calibrator |
| AC Clamp-on Ammeters (Toroidal Type) Transformer Type Sensor ¹ | (20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz | 0.34 % of reading + 35 mA 0.95 % of reading + 66 mA 0.38 % of reading + 0.17 A 1.2 % of reading + 0.29 A | Comparison to Fluke 5520A Calibrator, Wavetek Coil |
| AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹ | (20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz | 0.66 % of reading + 0.26 A 1.2 % of reading + 0.29 A 0.68 % of reading + 1 A 1.4 % of reading + 1.2 A | Comparison to Fluke 5520A Calibrator, Wavetek Coil |
| AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹ | (1 000 to 6 000) A (10 to 300) Hz (300 to 440) Hz | 0.77 % of reading 0.77 % of reading | Comparison to Fluke 52120A Transconductance Amplifier, Fluke 5520A Multiproduct Calibrator, 3 kA or 6 kA Coil |
| DC Resistance – Source/Measure ¹ | Up to 10 Ω (10 to 100) Ω (0.1 to 1) k Ω (1 to 10) k Ω (10 to 100) k Ω | 18 $\mu\Omega/\Omega$ + 58 $\mu\Omega$ 15 $\mu\Omega/\Omega$ + 0.58 m Ω 13 $\mu\Omega/\Omega$ + 0.58 m Ω 12 $\mu\Omega/\Omega$ + 5.8 m Ω 13 $\mu\Omega/\Omega$ + 58 m Ω | Comparison to Agilent 3458A Opt 02 8.5 Digit Multimeter, Decade Resistor |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|---|---|
| DC Resistance – Source/Measure ¹ | (0.1 to 1) MΩ (1 to 10) MΩ (10 to 100) MΩ (0.1 to 1) GΩ | 21 µΩ/Ω + 2.3 Ω 62 µΩ/Ω + 120 Ω 0.059 % of reading + 1.2 kΩ 0.82 % of reading + 12 kΩ | Comparison to Agilent 3458A Opt 02 8.5 Digit Multimeter, Decade Resistor |
| DC Resistance – Source ¹ (Fixed) | 0.33 mΩ 1 mΩ 10 mΩ 100 mΩ | 0.047 % of reading 0.037 % of reading 0.013 % of reading 0.012 % of reading | Comparison to DC Current Shunts |
| DC Resistance – Source (Variable) | (10 to 100) MΩ (0.1 to 1) GΩ (1 to 10) GΩ (10 to 100) GΩ (100 to 900) GΩ 1 TΩ | 0.036 % of reading 0.13 % of reading 0.25 % of reading 0.59 % of reading 0.77 % of reading 1.6 % of reading | Comparison to Decade Resistor |
| DC Voltage – Source/Measure ¹ | (0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 500) V (500 to 800) V (800 to 1 000) V | 8.3 µV/V + 0.58 µV 5.3 µV/V + 0.58 µV 5.3 µV/V + 0.58 µV 7.7 µV/V + 35 µV 15 µV/V + 0.12 mV 18 µV/V + 0.12 mV 21 µV/V + 0.12 mV | Comparison to Agilent 3458A Opt 02 8.5 Digit Multimeter, Fluke 5700A-EP Multiproduct Calibrator |
| DC High Voltage – Measure ¹ | (1 to 10) kV (10 to 20) kV (20 to 70) kV (70 to 100) kV | 0.04 % of reading + 92 mV 0.09 % of reading + 2.4 V 0.09 % of reading + 2.4 V 0.17 % of reading + 2.5 V | Comparison to Vitrek 4700 Digital HV Meter, Associated High Voltage Probes |
| DC Voltage – Source ¹ | Up to 0.22 V (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V | 8.6 µV/V + 0.4 µV 5.1 µV/V + 0.7 µV 4 µV/V + 2.5 µV 3.9 µV/V + 4 µV 6.2 µV/V + 40 µV 7.6 µV/V + 0.4 mV | Comparison to Fluke 5720A Multiproduct Calibrator, Fluke 5725A Amplifier |
| AC Voltage – Measure ¹ | Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz | 0.04 % of reading + 3.5 µV 0.03 % of reading + 1.2 µV 0.04 % of reading + 1.2 µV 0.15 % of reading + 1.2 µV 0.59 % of reading + 1.2 µV 4.6 % of reading + 2.3 µV 1.5 % of reading + 5.8 µV 8.1 % of reading + 8.1 µV | Comparison to Agilent 3458A Opt 02 8.5 Digit Multimeter |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-----------------------------------|--|--|---|
| AC Voltage – Measure ¹ | (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (0.1 to 1) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (1 to 10) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz | 0.013 % of reading + 4.6 μ V 0.009 7 % of reading + 2.3 μ V 0.017 % of reading + 2.3 μ V 0.038 % of reading + 2.3 μ V 0.093 % of reading + 2.3 μ V 0.36 % of reading + 12 μ V 1.2 % of reading + 12 μ V 1.8 % of reading + 12 μ V 4.7 % of reading + 81 μ V 4.7 % of reading + 92 μ V 17 % of reading + 0.12 mV 0.008 8 % of reading + 46 μ V 0.008 3 % of reading + 23 μ V 0.017 % of reading + 23 μ V 0.036 % of reading + 23 μ V 0.093 % of reading + 23 μ V 0.35 % of reading + 0.12 mV 1.2 % of reading + 0.12 mV 1.8 % of reading + 0.12 mV 4.6 % of reading + 0.81 mV 4.6 % of reading + 0.92 mV 17 % of reading + 1.2 mV 0.009 5 % of reading + 0.46 mV 0.023 % of reading + 0.23 mV 0.017 % of reading + 0.23 mV 0.036 % of reading + 0.23 mV 0.093 % of reading + 0.23 mV 0.35 % of reading + 1.2 mV 1.2 % of reading + 1.2 mV 1.8 % of reading + 1.2 mV 4.6 % of reading + 8.1 mV 4.6 % of reading + 9.2 mV 17 % of reading + 12 mV | Comparison to Agilent 3458A Opt 02 8.5 Digit Multimeter |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|---|--|
| AC Voltage – Measure ¹ | (10 to 100) V (1 to 40) Hz 40Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz | 0.024 % of reading + 4.6 mV 0.024 % of reading + 2.3 mV 0.024 % of reading + 2.3 mV 0.041 % of reading + 2.3 mV 0.14 % of reading + 2.3 mV 0.46 % of reading + 12 mV 1.7 % of reading + 12 mV 0.048 % of reading + 46 mV 0.048 % of reading + 23 mV 0.071 % of reading + 23 mV 0.19 % of reading + 23 mV 0.35 % of reading + 23 mV | Comparison to Agilent 3458A Opt 2 8.5 Digit Multimeter |
| AC High Voltage – Measure ¹ | (0.7 to 10) kV (20 to 100) Hz (100 to 400) Hz | 0.14 % of reading + 0.37 V 0.48 % of reading + 0.17 V | Comparison to Vitrek 4700 Digital HV Meter |
| AC High Voltage – Measure ¹ | (10 to 30) kV (30 to 70) Hz (70 to 200) Hz (200 to 450) Hz (30 to 50) kV (30 to 70) Hz (70 to 200) Hz (200 to 450) Hz (50 to 70) kV (30 to 70) Hz (70 to 200) Hz | 0.11 % of reading + 2.4 V 0.7 % of reading + 2.4 V 1.4 % of reading + 2.4 V 0.13 % of reading + 2.5 V 0.7 % of reading + 2.5 V 2.9 % of reading + 2.5 V 0.16 % of reading + 2.6 V 1.2 % of reading + 2.6 V | Comparison to Vitrek 4700 Digital HV Meter; Vitrek HVL-35, HVL-70, HVL-100 High Voltage Probes |
| AC Voltage – Source ¹ | Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz | 0.16 % of reading + 4 μ V 0.1 % of reading + 4 μ V 0.078 % of reading + 4 μ V 0.13 % of reading + 4 μ V 0.17 % of reading + 5 μ V 0.33 % of reading + 10 μ V 0.47 % of reading + 20 μ V 0.58 % of reading + 20 μ V | Comparison to Fluke 5720A Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------------|--|--|---|
| AC Voltage – Source ¹ | (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 o 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 o 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (0.22 to 2.2) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 o 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 o 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz | 0.042 % of reading + 4 µV 0.03 % of reading + 4 µV 0.014 % of reading + 4 µV 0.03 % of reading + 4 µV 0.058 % of reading + 5 µV 0.12 % of reading + 10 µV 0.16 % of reading + 20 µV 0.27 % of reading + 20 µV 0.028 % of reading + 12 µV 0.011 % of reading + 7 µV 0.0085 % of reading + 7 µV 0.021 % of reading + 7 µV 0.047 % of reading + 17 µV 0.091 % of reading + 20 µV 0.14 % of reading + 25 µV 0.28 % of reading + 45 µV 0.027 % of reading + 40 µV 0.01 % of reading + 15 µV 0.0048 % of reading + 8 µV 0.008 % of reading + 10 µV 0.012 % of reading + 30 µV 0.043 % of reading + 80 µV 0.1 % of reading + 0.2 mV 0.18 % of reading + 0.3 mV 0.028 % of reading + 0.4 mV 0.01 % of reading + 0.15 mV 0.0049 % of reading + 50 µV 0.0083 % of reading + 0.1 mV 0.01 % of reading + 0.2 mV 0.03 % of reading + 0.6 mV 0.1 % of reading + 2 mV 0.17 % of reading + 3.2 mV | Comparison to Fluke 5720A Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|--|--|
| AC Voltage – Source ¹ | (22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 o 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz | 0.028 % of reading + 4 mV 0.01 % of reading + 1.5 mV 0.005 6 % of reading + 0.6 mV 0.009 3 % of reading + 1 mV 0.016 % of reading + 2.5 mV 0.09 % of reading + 16 mV 0.44 % of reading + 40 mV 0.8 % of reading + 80 mV | Comparison to Fluke 5720A Multiproduct Calibrator |
| AC Voltage – Source ¹ | (220 to 750) V (30 to 50) kHz (50 to 100) kHz (220 to 1 100) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz | 0.061 % of reading + 11 mV 0.23 % of reading + 45 mV 0.011 % of reading + 4 mV 0.017 % of reading + 6 mV 0.061 % of reading + 11 mV | Comparison to Fluke 5720A Multiproduct Calibrator, Fluke 5725A Amplifier |
| Capacitance – Source ¹ (Simulation) | (0.19 to 3.3) nF (3.3 to 11) nF (11 to 110) nF (110 to 330) nF (0.33 to 1.1) μ F (1.1 to 3.3) μ F (3.3 to 11) μ F (11 to 33) μ F (33 to 110) μ F (110 to 330) μ F (0.33 to 1.1) mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF | 0.39 % of reading + 7.8 pF 0.21 % of reading + 7.8 pF 0.21 % of reading + 78 pF 0.21 % of reading + 0.23 nF 0.2 % of reading + 0.78 nF 0.2 % of reading + 2.3 nF 0.2 % of reading + 7.8 nF 0.31 % of reading + 23 nF 0.35 % of reading + 78 nF 0.35 % of reading + 0.23 μ F 0.35 % of reading + 0.78 μ F 0.35 % of reading + 2.3 μ F 0.35 % of reading + 7.8 μ F 0.58 % of reading + 23 μ F 0.86 % of reading + 78 μ F | Comparison to Fluke 5520A Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|---|--|
| Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹ | <p>Type B</p> <p>(250 to 350) °C (350 to 445) °C (445 to 580) °C (580 to 750) °C (750 to 1 000) °C (1 000 to 1 820) °C</p> <p>Type C</p> <p>(0 to 250) °C (250 to 1 000) °C (1 000 to 1 500) °C (1 500 to 1 800) °C (1 800 to 2 000) °C (2 000 to 2 250) °C (2 250 to 2 315) °C</p> <p>Type E</p> <p>(-270 to -245) °C (-245 to -195) °C (-195 to -155) °C (-155 to -90) °C (-90 to 0) °C (0 to 15) °C (15 to 890) °C (890 to 1 000) °C</p> <p>Type J</p> <p>(-210 to -180) °C (-180 to -120) °C (-120 to -50) °C (-50 to 990) °C (990 to 1 200) °C</p> <p>Type K</p> <p>(-270 to -255) °C (-255 to -195) °C (-195 to -115) °C (-115 to -55) °C (-55 to 1 000) °C (1 000 to 1 372) °C</p> | <p>1.2 °C 0.9 °C 0.71 °C 0.55 °C 0.45 °C 0.35 °C</p> <p>0.24 °C 0.19 °C 0.21 °C 0.24 °C 0.27 °C 0.33 °C 0.37 °C</p> <p>1.6 °C 0.24 °C 0.12 °C 0.095 °C 0.08 °C 0.076 °C 0.064 °C 0.074 °C</p> <p>0.15 °C 0.12 °C 0.093 °C 0.08 °C 0.094 °C</p> <p>2.5 °C 0.85 °C 0.16 °C 0.12 °C 0.087 °C 0.096 °C</p> | Comparison to Ectron 1140A Thermocouple Calibrator/Simulator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|---|--|
| Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹ | Type N (-270 to -260) °C (-260 to -200) °C (-200 to -140) °C (-140 to -70) °C (-70 to 25) °C (-25 to 160) °C (160 to 1 300) °C Type R (-50 to -30) °C (-30 to 45) °C (45 to 160) °C (160 to 380) °C (380 to 775) °C (775 to 1 768) °C Type S (-50 to -30) °C (-30 to 45) °C (45 to 105) °C (105 to 310) °C (310 to 615) °C (615 to 1 768) °C Type T (-270 to -255) °C (-255 to -240) °C (-240 to -210) °C (-210 to -150) °C (-150 to -40) °C (-40 to 100) °C (100 to 400) °C | 5.4 °C 1.5 °C 0.29 °C 0.18 °C 0.14 °C 0.12 °C 0.11 °C 0.8 °C 0.69 °C 0.49 °C 0.35 °C 0.3 °C 0.26 °C 0.76 °C 0.68 °C 0.49 °C 0.41 °C 0.35 °C 0.31 °C 1.9 °C 0.6 °C 0.36 °C 0.22 °C 0.15 °C 0.095 °C 0.08 °C | Comparison to Ectron 1140A Thermocouple Calibrator/Simulator |
| Scope Voltage – Source ¹ Amplitude DC into 50 Ω load into 1 MΩ load | (-6 to 6) V (-130 to 130) V | 0.2 % of reading + 31 µV 0.04 % of reading + 31 µV | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|--|--|
| Scope Voltage – Source ¹ Square Wave into 50 Ω load | 10 Hz to 100 kHz 1 mV p-p to 6.6 Vp-p | 0.19 % of reading + 31 µV | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |
| Scope Voltage – Source ¹ into 1 MΩ load | 10 Hz to 1 kHz 1 mV p-p to 6.6 Vp-p (1 kHz to 10) kHz 1 mV p-p to 6.6 Vp-p | 0.08 % of reading + 31 µV 0.19 % of reading + 31 µV | |
| Scope – Time Markers ¹ into 50 Ω load | 1 ns to 20 ms 50 ms 0.1 s 0.2 s 0.5 s 1 s 2 s 5 s | 0.0002 % of reading 2.3 µs 7.6 µs 28 µs 0.16 ms 0.62 ms 2.4 ms 15 ms | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |
| Scope Rise Time – Source ^{1,2} into 50 Ω load Rate: 1 kHz to 2 MHz Rate: 2 MHz to 10 MHz | 5 mVp-p to 2.5 Vp-p 250 ps (nominal) 250 ps (nominal) | 50 ps 50 ps | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |
| Scope Leveled Sine Wave – Source ¹ into 50 Ω load | 5 mVp-p to 5 Vp-p 50 kHz 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (600 to 1 100) MHz | 1.8 % of reading + 0.23 mV 2.8 % of reading + 0.23 mV 3.2 % of reading + 0.23 mV 4 % of reading + 0.23 mV 4.9 % of reading + 0.23 mV | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |
| Scope Bandwidth/Flatness – Source ¹ into 50 Ω load (50 kHz Reference) | 5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (600 to 1 100) MHz | 1.4 % of reading + 78 µV 1.8 % of reading + 78 µV 3.2 % of reading + 78 µV 3.9 % of reading + 78 µV | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |
| Scope Input Impedance – Measure ¹ | (40 to 60) Ω (0.5 to 1.5) MΩ | 0.082 % of reading 0.081 % of reading | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|--|--|
| Scope Input Capacitance – Measure ¹ | (5 to 50) pF | 3.9 % of reading + 0.39 pF | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |
| Scope Waveform Generator – Source ¹ | | | |
| Amplitude (Sine, Square, Triangle) into 50 Ω load into 1 MΩ load | 10 Hz to 10 kHz 1.8 mVp-p to 2.5 Vp-p 1.8 mVp-p to 55 Vp-p | 2.3 % of reading + 78 μV 2.3 % of reading + 78 μV | Comparison to Fluke 5520A/1100 Multiproduct Calibrator |
| Frequency (Sine, Square, Triangle) | 10 Hz to 10 kHz | 0.001 9 % of reading + 12 mHz | |
| LF Phase – Source ¹ | (0 to 90)° (10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz | 0.11° 0.2° 0.39° 1.9° 3.9° 7.8° | Comparison to Fluke 5520A/1100 Calibrator |
| DC Power – Source ¹ | | | |
| (0.33 to 330) mA | 11 μW to 1.1 mW 1.1 mW to 0.11 W (0.11 to 110) W (110 to 330) W | 0.024 % of reading 0.027 % of reading 0.024 % of reading 0.018 % of reading | |
| (0.33 to 3) A | 11 μW to 110 mW (0.11 to 990) W (0.99 to 3) kW | 0.044 % of reading 0.053 % of reading 0.009 6 % of reading | Comparison to Fluke 5520A/1100 Calibrator |
| (3 to 20.5) A | 99 mW to 0.99 W 0.99 W to 6.8 kW (6.8 to 20.5) kW | 0.088 % of reading 0.07 % of reading 0.04 % of reading | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|---|--|
| AC Power – Source ^{1,6} PF = 1 | | | |
| (3.3 to 9) mA | (10 to 65) Hz (0.11 mW to 3) mW 3 mW to 9 W | 0.13 % of reading 0.077 % of reading | |
| (9 to 33) mA | (10 to 65) W (0.3 to 10) mW 10 mW to 33 W | 0.089 % of reading 0.077 % of reading | |
| (33 to 90) mA | (10 to 65) Hz (1 to 30) mW 30 mW to 90 W | 0.071 % of reading 0.057 % of reading | |
| (90 to 330) mA | (10 to 65) Hz (3 to 100) mW 100 mW to 300 W | 0.089 % of reading 0.078 % of reading | |
| (0.33 to 0.9) A | (10 to 65) Hz (11 to 300) mW (0.3 to 900) W | 0.071 % of reading 0.081 % of reading | |
| (0.9 to 2.2) A | (10 to 65) Hz (30 to 720) mW 0.72 W to 2 kW | 0.089 % of reading 0.079 % of reading | |
| (2.2 to 4.5) A | (10 to 65) Hz 80 mW to 1.4 W 1.4 W to 4.5 kW | 0.088 % of reading 0.18 % of reading | |
| (4.5 to 20.5) A | (10 to 65) Hz 150 mW to 230 kW | 0.17 % of reading | |
| | | | Comparison to Fluke 5520A/11 Multiproduct Calibrator |

Electrical – RF/Microwave

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-------------------------------------|---|---|--|
| Total Harmonic Distortion – Measure | (-80 to 0) dB 20 Hz to 20 kHz (20 to 100) kHz | 1.2 dB 2.3 dB | Comparison to Agilent 8903B Audio Analyzer |
| Harmonic Distortion | 9 kHz to 100 MHz | 1.7 dB | Comparison to Agilent 8592L Spectrum Analyzer |
| Rise Time – Measure ¹ | \geq 700 ps | 0.81 ns | Comparison to Tektronix TDS3052 Digital Oscilloscope |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|--|--|
| Angle Measuring Devices ³ | (0.017 to 5)° (5 to 20)° (20 to 35)° (35 to 45)° (45 to 60)° (60 to 75)° (75 to 85)° 90° | 1.7" 2.4" 3.8" 5.1" 8.5" 18" 55" 1.7" | Comparisons to Sine Bar, Gage Blocks, Surface Plate Granite Master Square |
| Bubble Levels ³ | (0 to 140)" | 3" | Comparison to Gage Blocks |
| Micrometers and Calipers—Outside, Inside, Depth ^{1,3} | (0.05 to 1) in (1 to 9) in (5 to 15) in (15 to 40) in | 13 μ in (10 + 4L) μ in (11 + 4.6L) μ in (14 + 4.6L) μ in | Comparison to Gage Blocks, Long Blocks |
| Anvil Flatness ¹ | Up to 1 in Diameter | 4.7 μ in | Comparison to Optical Flats |
| Anvil Parallelism ¹ | Up to 1 in | 6.1 μ in | Comparison to Optical Parallels |
| Dial Indicators ^{1,3} | Up to 0.1 in (0.1 to 6) in | 4.5 μ in (4 + 4L) in | Comparison to Gage Blocks, Surface Plate |
| Height Gages, Digital Indicators ³ | (0.05 to 1) in (1 to 9) in (4 to 15) in (15 to 24) in | (10 + 3L) μ in (12 + 4L) μ in (11 + 4.6L) μ in (14 + 4.6L) μ in | Comparison to Gage Blocks, Long Blocks, Surface Plate |
| Length – Single Axis ³ Outside Dimension | Up to 1 in (1 to 7) in (7 to 12) in | (6 + 1L) μ in (4 + 4L) μ in (4.5L) μ in | Comparison to Universal Length Measuring Machine |
| Inside Dimension | (0.04 to 1) in (1 to 2.5) in (2.5 to 10) in (10 to 14) in | (10 + 1L) μ in (9 + 4L) μ in (12 + 4L) μ in (26 + 3L) μ in | |
| Linear Dimensions – Two Axis (X-Y) | 12 in x 12 in | 320 μ in | Comparison to Vision System |
| Master 1-2-3 Blocks, Caliper Masters, Parallels ³ | Up to 6 in (6 to 24) in | (10 + 3L) μ in (12 + 4L) μ in | Comparison to Gage Blocks, Surface Plate, Gage Amplifier |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|---|---|
| Optical Comparators ^{1,3} X-Y Length | Up to 2 in (2 to 12) in | (42 + 36L) μ in (75 + 27L) μ in | Comparison to Linear Glass Scale |
| Squareness | (0.001 to 10) in | (90 + 1L) μ in | Glass Scale |
| Magnification | (10 to 50) X | (120 + 10L) μ in | Magnification Checker |
| Parallelism, Straightness | Up to 12 in | 20 μ in | Comparison to Gage Amplifier, Surface Plate |
| Cylindrical Plug Gages ³ Outside Diameter | Up to 1 in (1 to 7) in | 12 μ in (10 + 3.5L) μ in | Comparison to Universal Length Measuring Machine |
| Cylindrical Ring Gages ³ Inside Diameter | (0.4 to 1) in (1 to 2.5) in (2.5 to 10) in (10 to 14) in | (10 + 1L) μ in (9 + 4L) μ in (12 + 4L) μ in (26 + 3L) μ in | Comparison to Universal Length Measuring Machine |
| Steel Rules | Up to 12 in | 320 μ in | Comparison to |
| Surface Plates ^{1,3} Overall Flatness | Up to 168.4 in DL | 1.7 $\sqrt{DL} + 5.5$ μ in | In accordance with Fed Spec GGG-P-463 utilizing Electronic Levels |
| Local Area Flatness (Repeat Readings) | Up to 0.001 in | 32 μ in | Supramess Indicator |
| Thread Plug Gages ³ Pitch Diameter | Up to 1 in (1 to 4) in (4 to 7) in | 79 μ in 80 μ in 84 μ in | Comparison to Thread Wires, Universal Length Measuring Machine |
| Major Diameter | Up to 1 in (1 to 7) in | 13 μ in (10 + 3.5L) μ in | |
| Threaded Ring Pitch Diameter | Up to 1 in (1 to 4) in (4 to 7) in | 79 μ in 80 μ in 84 μ in | Comparison to Master Setting Plugs |
| Thread Wires | (2 to 120) TPI (0.008 33 to 0.5) in | 12 μ in | Comparison to Universal Length Measuring Machine |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|---|--|
| Durometers | | | |
| Spring Force | | | |
| Type A, B, E, O | Up to 100 Duro | 0.79 Duro | Partial Verification per ASTM D2240 using Durometer Calibrator |
| Type D, C, DO | Up to 100 Duro | 0.8 Duro | |
| Indenter Length | Up to 1 in | 320 μ in | Vision System |
| Force Measuring Equipment | (10 to 100) gf | 0.04 % of reading | Comparison to ASTM E617 Class 2 Weights |
| Force Measuring Equipment | (0.2 to 500) lbf | 0.025 % of reading + 0.001 lbf | Comparison to NIST Class F Weights |
| Force Measuring Equipment | (500 to 1 000) lbf | 0.58 lbf | Comparison to Master Load Cells |
| Mass – Measure | 1 g to 1 kg (1 to 5.1) kg | 18 mg 0.18 g | Comparison to Mettler PR5003 DR Electronic Balance |
| Balances and Scales ^{1,4} (Metric) | Up to 500 mg (0.5 to 10) g 10 g to 3 kg (3 to 8) kg (8 to 13) kg (13 to 19) kg (19 to 27) kg (27 to 38) kg (38 to 40) kg | 6 μ g 22 μ g 6 mg 8 mg 13 mg 16 mg 76 mg 77 mg 78 mg | SET 1: ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system. |
| Balances and Scales ^{1,4} (Metric) | Up to 500 mg (0.5 to 5) g (5 to 10) g (10 to 20) g 50 g to 3 kg (3 to 5) kg (5 to 10) kg (10 to 14) kg (14 to 19) kg (19 to 23) kg (23 to 26) kg (26 to 29) kg (29 to 33) kg (33 to 35) kg (35 to 40) kg | 20 μ g 40 μ g 60 μ g 90 μ g 10 mg 15 mg 20 mg 34 mg 37 mg 76 mg 77 mg 78 mg 82 mg 83 mg 84 mg | SET 2: ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system. |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|---|--|
| Balances and Scales ^{1,4} Avoirdupois | Up to 1 lb (1 to 1 600) lb | 0.024 % of reading 0.012 % of reading | NIST Class F weights and internal calibration procedure utilized for the calibration of the weighing system. |
| | Up to 500 g 500 g to 726 kg | 0.024 % of reading 0.012 % of reading | |
| Torque Devices ¹ | (3 to 15) ozf·in (15 to 200) ozf·in (1 to 12.5) lbf·ft (12.5 to 600) lbf·ft (600 to 2 000) lbf·ft | 1.7 % of reading + 0.006 ozf·in 0.44 % of reading + 0.3 ozf·in 0.44 % of reading 0.34 % of reading 1.3 % of reading | Comparison to Torque Calibrator |
| Torque Calibration Equipment | (2.5 to 15) ozf·in (15 to 80) ozf·in | 0.055 % of reading 0.06 % of reading | Comparison to Torque Wheels, Weights |
| Torque Calibration Equipment | (0.42 to 50) lbf·ft (50 to 2 000) lbf·ft | 0.06 % of reading 0.06 % of reading | Comparison to Torque Arm, Weights |
| Absolute Pressure – Source | (0 to 30) psia (30 to 1 000) psia | 0.002 4 psi 0.006 6 % of reading + 0.000 1 psi | Fluke/DHI RPM 4 Comparison to Pressure Controller/Calibrator |
| Pressure – Source ¹ | (0.14 to 25) psig | 0.017 % of reading + 0.000 041 psi | Comparison to Ametek RK-1000 WC Deadweight Tester |
| Pressure – Source | (-15 to 30) psig (30 to 1 000) psig | 0.002 1 psi 0.006 6 % of reading + 0.000 1 psi | Comparison to Fluke/DHI RPM 4 Pressure Controller/Calibrator |
| Pressure – Source ¹ (Hydraulic) | (5 to 15 000) psig | 0.018 % of reading | Comparison to Fluke RPM4-E-DWT Deadweight Tester |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|---|---|
| Relative Humidity – Source | (-10 to 15) °C (10 to 75) %RH (75 to 95) %RH (15 to 35) °C (10 to 95) %RH (35 to 70) °C (50 to 70) %RH (70 to 95) %RH | 0.5 %RH 0.65 %RH 0.5 %RH 0.7 %RH 0.85 %RH | Comparison to Thunder Scientific 2500 Two-Pressure Humidity Generator |
| Relative Humidity – Measure ¹ | (10 to 30) °C (10 to 90) %RH (90 to 99) %RH | 1.3 %RH 2.3 %RH | Comparison to Vaisala HMI41/HMP46 Temperature/Humidity Indicator w/ Probe |
| Temperature – Measure ¹ | (-196 to 0) °C (0 to 420) °C (420 to 660) °C | 0.011 °C 0.026 °C 0.035 °C | Comparison to AccuMac AM1760 Secondary SPRT, Black Stack Indicator |
| Temperature – Measure ¹ | (660 to 1 000) °C (1 000 to 1 200) °C | 0.93 °C 1.2 °C | Comparison to AccuMac AM1210 Type S Thermocouple Reference Standard, Black Stack Indicator |
| Temperature – Source | (-20 to 120) °C | 0.028 °C | Comparison to RTD Probe, Temperature Indicator; Liquid Bath |
| Temperature – Source | (120 to 600) °C | 0.13 °C | Comparison to RTD Probe, Temperature Indicator; Dry-block Calibrator |
| Infrared Measuring Devices | (-15 to 0) °C (0 to 50) °C (50 to 100) °C (100 to 120) °C (120 to 200) °C (200 to 350) °C (350 to 500) °C | 0.98 °C 0.67 °C 0.71 °C 0.77 °C 0.94 °C 1.7 °C 2.1 °C | Comparison to Hart Black Body (flat plate) $\varepsilon = (0.1 \text{ to } 1)$, $\lambda = (8 \text{ to } 14) \mu\text{m}$ |

Time and Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|--|---|
| Frequency – Reference ⁵ | 10 MHz | 6.4 nHz/Hz | Comparison to SRS FS725 Rubidium Frequency Standard |
| Frequency – Source/Measure ¹ | 30 Hz to 225 MHz | 2.1 μ Hz/Hz | Comparison to HP 53131A (10) Frequency Counter, SRS FS725 Rubidium Frequency Standard |
| Time – Measure ¹ | Up to 599 s/month | 58 ms/d | Comparison to Vibrograf 4500 Timometer |
| AC Duty Cycle – Source ¹ Square Wave: < 3.3 Vp-p Freq: 0.1 Hz to 100 kHz | (1 to 10) % Duty Cycle 10 μ s to 100 s (10 to 49) % Duty Cycle 10 μ s to 100 s 50 % Duty Cycle 10 μ s to 100 s (51 to 90) % Duty Cycle 10 μ s to 100 s (90 to 99) % Duty Cycle 10 μ s to 100 s | 0.62 % of reading + 78 ns 0.039 % of reading + 78 ns 0.001 6 % of reading + 78 ns 0.039 % of reading + 78 ns 0.62 % of reading + 78 ns | Comparison to Fluke 5522A Multiproduct Calibrator |

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. The stated uncertainty is the laboratory's ability to source a fast rise pulse that is approximately 250 ps. In the typical application of measuring rise time of an oscilloscope, this value is one of the contributing factors, but other factors are derived from the DUT. The known source rise time is mathematically removed from the total measured rise time measured on the DUT.
3. L = length in inches; DL = diagonal length in inches; $''$ = arc-second.
4. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
5. As frequency & amplitude deviate from the listed values, uncertainty may be higher than stated. If needed, contact laboratory for more information regarding uncertainties at frequency and range combinations other than the ones shown.
6. The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact the laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
7. Unless otherwise specified in the far-right column, the laboratory is utilizing an in-house developed calibration procedure.
8. The legal entity for this location is Transcat, Inc.
9. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.01.



A handwritten signature in black ink, appearing to read "Jason Stine".

Jason Stine, Vice President

