



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat – San Juan

281 Calle Matadero
Urb Puerto Nuevo
San Juan, PR 00920

Fulfills the requirements of

ISO/IEC 17025:2017

and national standards

ANSI/NCSL Z540-1-1994 (R2002) AND

ANSI/NCSL Z540.3-2006 (R2013)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.

The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 07 September 2025
Certificate Number: AC-2489.05



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

ANSI/NCSL Z540.3-2006 (R2013)

Transcat – San Juan

281 Calle Matadero

Urb Puerto Nuevo

San Juan, PR 00920

John Strouth 787-706-8855

CALIBRATION

Valid to: **September 7, 2025**

Certificate Number: **AC-2489.05**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	(29 to 329.99) μ A	0.16 % of reading + 80 nA	Fluke 5520A Multiproduct Calibrator
	(10 to 20) Hz	0.12 % of reading + 80 nA	
	(20 to 45) Hz	0.097 % of reading + 0.12 μ A	
	45 Hz to 1 kHz	0.23 % of reading + 0.16 μ A	
	(1 to 5) kHz	0.62 % of reading + 0.16 μ A	
	(5 to 10) kHz	1.2 % of reading + 0.31 μ A	
	(10 to 30) kHz		
	(0.33 to 3.299 9) mA	0.16 % of reading + 0.12 μ A	
	(10 to 20) Hz	0.097 % of reading + 0.12 μ A	
	(20 to 45) Hz	0.078 % of reading + 0.12 μ A	
	45 Hz to 1 kHz	0.16 % of reading + 0.16 μ A	
	(1 to 5) kHz	0.39 % of reading + 0.23 μ A	
	(5 to 10) kHz	0.78 % of reading + 0.47 μ A	
	(10 to 30) kHz		
	(3.3 to 32.999) mA	0.14 % of reading + 1.6 μ A	
	(10 to 20) Hz	0.07 % of reading + 1.6 μ A	
	(20 to 45) Hz	0.031 % of reading + 1.6 μ A	
	45 Hz to 1 kHz	0.062 % of reading + 1.6 μ A	
	(1 to 5) kHz	0.16 % of reading + 2.3 μ A	
	(5 to 10) kHz	0.31 % of reading + 3.1 μ A	
	(10 to 30) kHz		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	(33 to 329.99) mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (0.33 to 1.1) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (1.1 to 3) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (3 to 11) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz (11 to 20) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.14 % of reading + 16 μ A 0.07 % of reading + 16 μ A 0.031 % of reading + 16 μ A 0.078 % of reading + 39 μ A 0.16 % of reading + 78 μ A 0.31 % of reading + 0.16 mA 0.14 % of reading + 78 μ A 0.039 % of reading + 78 μ A 0.47 % of reading + 0.78 mA 1.9 % of reading + 4 mA 0.14 % of reading + 78 μ A 0.049 % of reading + 78 μ A 0.47 % of reading + 0.78 mA 1.9 % of reading + 4 mA 0.049 % of reading + 2 mA 0.08 % of reading + 2 mA 2.3 % of reading + 2 mA 0.094 % of reading + 4 mA 0.11 % of reading + 4 mA 2.3 % of reading + 4 mA	Fluke 5520A Multiproduct Calibrator
AC Clamp-on Ammeters (Toroidal Type) Transformer Type Sensor ¹	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.3 % of reading + 26 mA 0.83 % of reading + 47 mA 0.35 % of reading + 0.12 A 1.1 % of reading + 0.22 A	Fluke 5520A Multiproduct Calibrator, Fluke 5500A/COIL 50-turn Coil
AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹	20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.57 % of reading + 0.26 A 1 % of reading + 0.29 A 0.6 % of reading + 0.9 A 1.3 % of reading + 0.92 A	Fluke 5520A Multiproduct Calibrator, Fluke 5500A/COIL 50-turn Coil

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ¹	Up to 100 μ A		Agilent 3458A Opt 002 8.5 Digit Multimeter
	(10 to 20) Hz	0.46 % of reading + 35 nA	
	(20 to 45) Hz	0.17 % of reading + 35 nA	
	(45 to 100) Hz	0.072 % of reading + 35 nA	
	100 Hz to 5 kHz	0.072 % of reading + 35 nA	
	(0.1 to 1) mA		
	(10 to 20) Hz	0.46 % of reading + 0.23 μ A	
	(20 to 45) Hz	0.17 % of reading + 0.23 μ A	
	(45 to 100) Hz	0.071 % of reading + 0.23 μ A	
	100 Hz to 5 kHz	0.038 % of reading + 0.23 μ A	
	(1 to 10) mA		
	(10 to 20) Hz	0.46 % of reading + 2.3 μ A	
	(20 to 45) Hz	0.17 % of reading + 2.3 μ A	
	(45 to 100) Hz	0.071 % of reading + 2.3 μ A	
	100 Hz to 5 kHz	0.038 % of reading + 2.3 μ A	
	(10 to 100) mA		
DC Current – Source ¹	(10 to 20) Hz	0.48 % of reading + 23 μ A	Fluke 5520A Multiproduct Calibrator
	(20 to 45) Hz	0.17 % of reading + 23 μ A	
	(45 to 100) Hz	0.071 % of reading + 23 μ A	
	100 Hz to 5 kHz	0.037 % of reading + 23 μ A	
	100 mA to 1 A		
	(10 to 20) Hz	0.46 % of reading + 0.23 mA	
	(20 to 45) Hz	0.19 % of reading + 0.23 mA	
	(45 to 100) Hz	0.097 % of reading + 0.23 mA	
DC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor ¹	(0 to 330) μ A	0.012 % of reading + 20 nA	Fluke 5520A Multiproduct Calibrator, Fluke 5500A/Coil 50-turn Coil
	(0.33 to 3.3) mA	0.008 2 % of reading + 50 nA	
	(3.3 to 33) mA	0.009 3 % of reading + 0.19 μ A	
	(33 to 330) mA	0.007 8 % of reading + 1.9 μ A	
	(0.33 to 1.1) A	0.016 % of reading + 31 μ A	
	(1.1 to 3) A	0.03 % of reading + 31 μ A	
	(3 to 11) A	0.051 % of reading + 0.39 mA	
	(11 to 20) A	0.083 % of reading + 0.58 mA	
DC Current – Measure ¹	(20 to 150) A	0.5 % of reading + 0.14 A	Agilent 3458A Opt 002 8.5 Digit Multimeter
	(150 to 1 000) A	0.51 % of reading + 0.5 A	
	Up to 100 μ A	33 μ A/A + 0.92 nA	
	(0.1 to 1) mA	29 μ A/A + 5.8 nA	
	(1 to 10) mA	29 μ A/A + 58 nA	
	(10 to 100) mA	46 μ A/A + 0.58 μ A	
	(0.1 to 1) A	0.13 % of reading + 12 μ A	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance – Source ¹ (Simulation)	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) k Ω (1.1 to 3.3) k Ω (3.3 to 11) k Ω (11 to 33) k Ω (33 to 110) k Ω (110 to 330) k Ω (0.33 to 1.1) M Ω (1.1 to 3.3) M Ω (3.3 to 11) M Ω (11 to 33) M Ω (33 to 110) M Ω (110 to 330) M Ω (0.33 to 1) G Ω	34 $\mu\Omega/\Omega$ + 1 m Ω 27 $\mu\Omega/\Omega$ + 1.2 m Ω 25 $\mu\Omega/\Omega$ + 1.1 m Ω 26 $\mu\Omega/\Omega$ + 1.6 m Ω 25 $\mu\Omega/\Omega$ + 1.6 m Ω 25 $\mu\Omega/\Omega$ + 16 m Ω 22 $\mu\Omega/\Omega$ + 16 m Ω 23 $\mu\Omega/\Omega$ + 0.16 Ω 23 $\mu\Omega/\Omega$ + 0.16 Ω 28 $\mu\Omega/\Omega$ + 1.6 Ω 27 $\mu\Omega/\Omega$ + 1.6 Ω 66 $\mu\Omega/\Omega$ + 23 Ω 0.11 m Ω/Ω + 39 Ω 0.2 m Ω/Ω + 1.9 k Ω 0.41 m Ω/Ω + 2.8 k Ω 0.23 % of reading + 0.1 M Ω 1.2 % of reading + 0.4 M Ω	Fluke 5520A Multiproduct Calibrator
DC Resistance – Source ¹ (Artifact)	10 G Ω 100 G Ω	63 M Ω 1.3 G Ω	Decade Resistance Standard
DC Resistance – Measure ¹	Up to 10 Ω (10 to 100) Ω (0.1 to 1) k Ω (1 to 10) k Ω (10 to 100) k Ω (0.1 to 1) M Ω (1 to 10) M Ω (10 to 100) M Ω (0.1 to 1) G Ω	18 $\mu\Omega/\Omega$ + 58 $\mu\Omega$ 15 $\mu\Omega/\Omega$ + 0.58 m Ω 13 $\mu\Omega/\Omega$ + 0.58 m Ω 12 $\mu\Omega/\Omega$ + 5.8 m Ω 13 $\mu\Omega/\Omega$ + 58 m Ω 21 $\mu\Omega/\Omega$ + 2.3 Ω 62 $\mu\Omega/\Omega$ + 0.12 k Ω 0.059 % of reading + 1.2 k Ω 0.82 % of reading + 12 k Ω	Agilent 3458A Opt 002 8.5 Digit Multimeter
AC Voltage – Source ¹	(1 to 32.999) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 329.990) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.063 % of reading + 4.7 μ V 0.015 % of reading + 4.7 μ V 0.018 % of reading + 4.7 μ V 0.079 % of reading + 4.7 μ V 0.27 % of reading + 9.3 μ V 0.62 % of reading + 39 μ V 0.023 % of reading + 6.2 μ V 0.012 % of reading + 6.2 μ V 0.013 % of reading + 8 μ V 0.028 % of reading + 8 μ V 0.062 % of reading + 25 μ V 0.16 % of reading + 54 μ V	Fluke 5520A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(0.33 to 3.299 99) V		Fluke 5520A Multiproduct Calibrator
	(10 to 45) Hz	0.023 % of reading + 39 μ V	
	45 Hz to 10 kHz	0.012 % of reading + 47 μ V	
	(10 to 20) kHz	0.015 % of reading + 47 μ V	
	(20 to 50) kHz	0.024 % of reading + 39 μ V	
	(50 to 100) kHz	0.054 % of reading + 97 μ V	
	(100 to 500) kHz	0.19 % of reading + 0.47 mV	
	(3.3 to 32.999 9) V		
	(10 to 45) Hz	0.023 % of reading + 0.51 mV	
	45 Hz to 10 kHz	0.012 % of reading + 0.47 mV	
	(10 to 20) kHz	0.019 % of reading + 0.47 mV	
	(20 to 50) kHz	0.027 % of reading + 0.47 mV	
	(50 to 100) kHz	0.07 % of reading + 1.2 mV	
	(33 to 329.999) V		
	45 Hz to 1 kHz	0.015 % of reading + 1.6 mV	
	(1 to 10) kHz	0.016 % of reading + 4.7 mV	
	(10 to 20) kHz	0.02 % of reading + 4.7 mV	
	(20 to 50) kHz	0.025 % of reading + 4.7 mV	
	(50 to 100) kHz	0.16 % of reading + 39 mV	
AC Voltage – Measure ¹	(330 to 1 020) V		Agilent 3458A Opt 002 8.5 Digit Multimeter
	45 Hz to 1 kHz	0.023 % of reading + 7.7 mV	
	(1 to 5) kHz	0.02 % of reading + 7.7 mV	
	(5 to 10) kHz	0.023 % of reading + 7.7 mV	
	Up to 10 mV		
	(1 to 40) Hz	0.04 % of reading + 3.5 μ V	
	40 Hz to 1 kHz	0.03 % of reading + 1.2 μ V	
	(1 to 20) kHz	0.04 % of reading + 1.2 μ V	
	(20 to 50) kHz	0.15 % of reading + 1.2 μ V	
	(50 to 100) kHz	0.59 % of reading + 1.2 μ V	
	(100 to 300) kHz	4.6 % of reading + 2.3 μ V	
	300 kHz to 1 MHz	1.5 % of reading + 5.8 μ V	
	(1 to 4) MHz	8.1 % of reading + 8.1 μ V	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	(10 to 100) mV		Agilent 3458A Opt 002 8.5 Digit Multimeter
	(1 to 40) Hz	0.013 % of reading + 4.6 μ V	
	40 Hz to 1 kHz	0.009 7 % of reading + 2.3 μ V	
	(1 to 20) kHz	0.017 % of reading + 2.3 μ V	
	(20 to 50) kHz	0.038 % of reading + 2.3 μ V	
	(50 to 100) kHz	0.093 % of reading + 2.3 μ V	
	(100 to 300) kHz	0.36 % of reading + 12 μ V	
	300 kHz to 1 MHz	1.2 % of reading + 12 μ V	
	(1 to 2) MHz	1.8 % of reading + 12 μ V	
	(2 to 4) MHz	4.7 % of reading + 81 μ V	
	(4 to 8) MHz	4.7 % of reading + 92 μ V	
	(8 to 10) MHz	17 % of reading + 0.12 mV	
	(0.1 to 1) V		
	(1 to 40) Hz	0.008 8 % of reading + 46 μ V	
	40 Hz to 1 kHz	0.008 3 % of reading + 23 μ V	
	(1 to 20) kHz	0.017 % of reading + 23 μ V	
	(20 to 50) kHz	0.036 % of reading + 23 μ V	
	(50 to 100) kHz	0.093 % of reading + 23 μ V	
	(100 to 300) kHz	0.35 % of reading + 0.12 mV	
	300 kHz to 1 MHz	1.2 % of reading + 0.12 mV	
	(1 to 2) MHz	1.8 % of reading + 0.12 mV	
	(2 to 4) MHz	4.6 % of reading + 0.81 mV	
	(4 to 8) MHz	4.6 % of reading + 0.92 mV	
	(8 to 10) MHz	17 % of reading + 1.2 mV	
	(1 to 10) V		
	(1 to 40) Hz	0.009 5 % of reading + 0.46 mV	
	40 Hz to 1 kHz	0.023 % of reading + 0.23 mV	
	(1 to 20) kHz	0.017 % of reading + 0.23 mV	
	(20 to 50) kHz	0.036 % of reading + 0.23 mV	
	(50 to 100) kHz	0.093 % of reading + 0.23 mV	
	(100 to 300) kHz	0.35 % of reading + 1.2 mV	
	300 kHz to 1 MHz	1.2 % of reading + 1.2 mV	
	(1 to 2) MHz	1.8 % of reading + 1.2 mV	
	(2 to 4) MHz	4.6 % of reading + 8.1 mV	
	(4 to 8) MHz	4.6 % of reading + 9.2 mV	
	(8 to 10) MHz	17 % of reading + 12 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	(10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.024 % of reading + 4.6 mV 0.024 % of reading + 2.3 mV 0.024 % of reading + 2.3 mV 0.041 % of reading + 2.3 mV 0.14 % of reading + 2.3 mV 0.46 % of reading + 12 mV 1.7 % of reading + 12 mV 0.048 % of reading + 46 mV 0.048 % of reading + 23 mV 0.071 % of reading + 23 mV 0.19 % of reading + 23 mV 0.35 % of reading + 23 mV	Agilent 3458A Opt 002 8.5 Digit Multimeter
DC Voltage – Source ¹	(0 to 330) mV (0.33 to 3) V (3 to 33) V (33 to 330) V (330 to 1 000) V	16 μ V/V + 0.78 μ V 9.1 μ V/V + 1.6 μ V 10 μ V/V + 16 μ V 16 μ V/V + 0.12 mV 15 μ V/V + 1.2 mV	Fluke 5520A Multiproduct Calibrator
DC Voltage – Measure ¹	(0 to 100) mV (0.1 to 10) V (10 to 100) V (100 to 500) V (500 to 800) V (800 to 1 000) V	8.3 μ V/V + 0.58 μ V 5.3 μ V/V + 0.58 μ V 7.7 μ V/V + 35 μ V 15 μ V/V + 0.12 mV 18 μ V/V + 0.12 mV 21 μ V/V + 0.12 mV	Agilent 3458A Opt 002 8.5 Digit Multimeter
Capacitance – Source ¹ (Variable Artifact)	1 kHz (50 to 190) pF	0.6 % of reading + 5 pF	Decade Capacitance Standard

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source ¹ (Simulation)	10 Hz to 10 kHz (0.19 to 1.099 9) nF	0.39 % of reading + 6.1 pF	Fluke 5520A Multiproduct Calibrator
	10 Hz to 3 kHz (1.1 to 3.299 9) nF	0.39 % of reading + 6.1 pF	
	10 Hz to 1 kHz (3.3 to 10.999 9) nF	0.21 % of reading + 6.1 pF	
	(11 to 109.999) nF	0.21 % of reading + 61 pF	
	(110 to 329.999) nF	0.21 % of reading + 0.18 nF	
	(10 to 600) Hz (0.33 to 1.099 99) μ F	0.2 % of reading + 0.61 nF	
	(10 to 300) Hz (1.1 to 3.299 9) μ F	0.2 % of reading + 1.9 nF	
	(10 to 150) Hz (3.3 to 10.999 9) μ F	0.2 % of reading + 6.1 nF	
	(10 to 120) Hz (11 to 32.999 9) μ F	0.32 % of reading + 18 nF	
	(10 to 80) Hz (33 to 109.999 9) μ F	0.35 % of reading + 61 nF	
	DC to 50 Hz (110 to 329.999) μ F	0.35 % of reading + 0.18 μ F	
	DC to 20 Hz (0.33 to 1.099 99) mF	0.35 % of reading + 0.61 μ F	
	DC to 6 Hz (1.1 to 3.299 99) mF	0.35 % of reading + 1.8 μ F	
	DC to 2 Hz (3.3 to 10.999 9) mF	0.35 % of reading + 6.1 μ F	
	DC to 0.6 Hz (11 to 32.999 9) mF	0.58 % of reading + 18 μ F	
	DC to 0.2 Hz (33 to 110) mF	0.85 % of reading + 61 μ F	
Capacitance – Source ¹ (Simulation)	DC to 0.6 Hz (11 to 32.999 9) mF DC to 0.2 Hz (33 to 110) mF	0.58 % of reading + 18 μ F 0.85 % of reading + 61 μ F	Fluke 5520A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹	Type B		Ectron 1140A Thermocouple Calibrator/Simulator
	(250 to 350) °C	1.2 °C	
	(350 to 445) °C	0.9 °C	
	(445 to 580) °C	0.71 °C	
	(580 to 750) °C	0.55 °C	
	(750 to 1 000) °C	0.45 °C	
	(1 000 to 1 820) °C	0.35 °C	
	Type E		
	(-270 to -245) °C	1.6 °C	
	(-245 to -195) °C	0.24 °C	
	(-195 to -155) °C	0.12 °C	
	(-155 to -90) °C	0.095 °C	
	(-90 to 0) °C	0.08 °C	
	(0 to 15) °C	0.076 °C	
	(15 to 890) °C	0.064 °C	
	(890 to 1 000) °C	0.074 °C	
	Type J		
	(-210 to -180) °C	0.15 °C	
	(-180 to -120) °C	0.12 °C	
	(-120 to -50) °C	0.093 °C	
	(-50 to 990) °C	0.08 °C	
	(990 to 1 200) °C	0.094 °C	
	Type K		
	(-270 to -255) °C	2.5 °C	
	(-255 to -195) °C	0.85 °C	
	(-195 to -115) °C	0.16 °C	
	(-115 to -55) °C	0.12 °C	
	(-55 to 1 000) °C	0.087 °C	
	(1 000 to 1 372) °C	0.096 °C	
	Type N		
	(-270 to -260) °C	5.4 °C	
	(-260 to -200) °C	1.5 °C	
	(-200 to -140) °C	0.29 °C	
	(-140 to -70) °C	0.18 °C	
	(-70 to 25) °C	0.14 °C	
	(-25 to 160) °C	0.12 °C	
	(160 to 1 300) °C	0.11 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹	Type R		Ectron 1140A Thermocouple Calibrator/Simulator
	(-50 to -30) °C	0.8 °C	
	(-30 to 45) °C	0.69 °C	
	(45 to 160) °C	0.49 °C	
	(160 to 380) °C	0.35 °C	
	(380 to 775) °C	0.3 °C	
	(775 to 1 768) °C	0.26 °C	
	Type S		
	(-50 to -30) °C	0.76 °C	
	(-30 to 45) °C	0.68 °C	
	(45 to 105) °C	0.49 °C	
	(105 to 310) °C	0.41 °C	
	(310 to 615) °C	0.35 °C	
	(615 to 1 768) °C	0.31 °C	
	Type T		
	(-270 to -255) °C	1.9 °C	
	(-255 to -240) °C	0.6 °C	
	(-240 to -210) °C	0.36 °C	
	(-210 to -150) °C	0.22 °C	
	(-150 to -40) °C	0.15 °C	
	(-40 to 100) °C	0.095 °C	
	(100 to 400) °C	0.08 °C	
DC Power – Source ¹			Fluke 5520A Multiproduct Calibrator
(0.33 to 330) mA	11 µW to 1.1 mW	0.024 % of reading	
	1.1 mW to 0.11 W	0.027 % of reading	
	(0.11 to 110) W	0.024 % of reading	
	(110 to 330) W	0.018 % of reading	
(0.33 to 3) A	11 µW to 110 mW	0.044 % of reading	
	(0.11 to 990) W	0.053 % of reading	
	(0.99 to 3) kW	0.009 6 % of reading	
(3 to 20.5) A	99 mW to 0.99 W	0.088 % of reading	
	0.99 W to 6.8 kW	0.07 % of reading	
	(6.8 to 20.5) kW	0.04 % of reading	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Source ^{1,2} PF = 1 (3.3 to 9) mA (9 to 33) mA (33 to 90) mA (90 to 330) mA (0.33 to 0.9) A (0.9 to 2.2) A (2.2 to 4.5) A (4.5 to 20.5) A	(10 to 65) Hz (0.11 mW to 3) mW 3 mW to 9 W (10 to 65) W (0.3 to 10) mW 10 mW to 33 W (10 to 65) Hz (1 to 30) mW 30 mW to 90 W (10 to 65) Hz (3 to 100) mW 100 mW to 300 W (10 to 65) Hz (11 to 300) mW (0.3 to 900) W (10 to 65) Hz (30 to 720) mW 0.72 W to 2 kW (10 to 65) Hz 80 mW to 1.4 W 1.4 W to 4.5 kW (10 to 65) Hz 150 mW to 230 kW	0.13 % of reading 0.077 % of reading 0.089 % of reading 0.077 % of reading 0.071 % of reading 0.057 % of reading 0.089 % of reading 0.078 % of reading 0.071 % of reading 0.081 % of reading 0.089 % of reading 0.079 % of reading 0.088 % of reading 0.18 % of reading 0.17 % of reading	Fluke 5520A Multiproduct Calibrator
Phase – Source ¹	(0 to 180)° (10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 20) kHz	0.11° 0.2° 0.4° 1.9° 3.9° 7.8°	Fluke 5520A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,3}			
Amplitude – DC into 50 Ω load into 1 M Ω load	(-6 to 6) V (-130 to 130) V	0.2 % of reading + 31 μ V 0.039 % of reading + 31 μ V	
Amplitude – Square Wave into 50 Ω load	10 Hz to 100 kHz 1 mV p-p to 6.6 Vp-p	0.19 % of reading + 31 μ V	
into 1 M Ω load	10 Hz to 1 kHz 1 mV p-p to 6.6 Vp-p (1 kHz to 10) kHz 1 mV p-p to 6.6 Vp-p	0.078 % of reading + 31 μ V 0.19 % of reading + 31 μ V	
Time Markers into 50 Ω load	1 ns to 20 ms 50 ms 0.1 s 0.2 s 0.5 s 1 s 2 s 5 s	0.000 22 % of reading 2.3 μ s 7.6 μ s 28 μ s 0.16 ms 0.62 ms 2.4 ms 15 ms	
Rise Time – Source into 50 Ω load Rate: 1 kHz to 2 MHz Rate: 2 MHz to 10 MHz	5 mVp-p to 2.5 Vp-p 250 ps (nominal) 250 ps (nominal)	51 ps 51 ps	
Leveled Sine Wave into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz (50 to 100) MHz (100 to 300) MHz (300 to 600) MHz	1.8 % of reading + 0.23 mV 2.8 % of reading + 0.23 mV 3.2 % of reading + 0.23 mV 4 % of reading + 0.23 mV	
Bandwidth Flatness – Source (50 kHz Reference) into 50 Ω load	5 mVp-p to 5.5 Vp-p (50 to 100) MHz (100 to 300) MHz (300 to 600) MHz	1.4 % of reading + 78 μ V 1.8 % of reading + 78 μ V 3.2 % of reading + 78 μ V	

Fluke 5520A/6
Multiproduct Calibrator
w/ 600 MHz Scope Option

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,3}			
Wave Generator – Amplitude (Sine, Square, Triangle) into 50 Ω load	10 Hz to 10 kHz 1.8 mVp-p to 2.5 Vp-p	2.3 % of reading + 78 μ V	Fluke 5520A/6 Multiproduct Calibrator w/ 600 MHz Scope Option
into 1 M Ω load	10 Hz to 10 kHz 1.8 mVp-p to 55 Vp-p	2.3 % of reading + 78 μ V	
Wave Generator – Frequency (Sine, Square, Triangle)	10 Hz to 10 kHz	0.001 9 % of reading + 12 mHz	
Wave Generator – Rise Time	250 ps	63 ps	
Input Impedance Measure into 50 Ω	(40 to 60) Ω	0.082 % of reading	
into 1 M Ω	(0.5 to 1.5) M Ω	0.081 % of reading	
Input Capacitance Measure	(5 to 50) pF	3.9 % of reading + 0.39 pF	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Micrometers and Calipers ^{1,4} (Outside, Inside, Depth)	(0.05 to 1) in (1 to 10) in	(15 + 7L) μ in (9 + 14L) μ in	Gage Blocks
Anvil Flatness ¹	Up to 1 in	6.1 μ in	Optical Flats

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination (SI Units)	1 mg 2 mg 3 mg 5 mg 10 mg	1.7 μ g 1.7 μ g 1.7 μ g 1.7 μ g 1.7 μ g	Echelon II

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination (SI Units)	20 mg	1.7 µg	Echelon II
	30 mg	1.7 µg	
	50 mg	1.7 µg	
	100 mg	1.7 µg	
	200 mg	1.7 µg	
	300 mg	1.7 µg	
	500 mg	1.7 µg	
	1 g	4.3 µg	
	2 g	4.3 µg	
	3 g	4.4 µg	
	5 g	5.2 µg	
	10 g	8.6 µg	
	20 g	12 µg	
	30 g	13 µg	
	50 g	17 µg	
	100 g	35 µg	
	200 g	0.1 mg	
	300 g	0.13 mg	
	500 g	0.17 mg	
	1 kg	0.33 mg	
	2 kg	0.98 mg	
	3 kg	1.2 mg	
	5 kg	1.7 mg	
	10 kg	3.4 mg	
	1 mg	3.6 µg	Echelon III
	2 mg	3.6 µg	
	3 mg	3.6 µg	
	5 mg	3.6 µg	
	10 mg	3.6 µg	
	20 mg	3.6 µg	
	30 mg	3.6 µg	
	50 mg	3.6 µg	
	100 mg	3.6 µg	
	200 mg	3.6 µg	
	300 mg	3.6 µg	
	500 mg	3.6 µg	
	1 g	11 µg	
	2 g	11 µg	
	3 g	11 µg	
	5 g	11 µg	
	10 g	18 µg	

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination (SI Units)	20 g 30 g 50 g 100 g 200 g 300 g 500 g 1 kg 2 kg 3 kg 5 kg 10 kg 20 kg 25 kg	26 µg 26 µg 41 µg 84 µg 0.2 mg 0.27 mg 0.41 mg 0.83 mg 1.9 mg 2.7 mg 4.1 mg 8.3 mg 0.23 g 0.23 g	Echelon III
Mass Determination ⁵ (Avoirdupois)	1 lb 2 lb 3 lb 5 lb 10 lb 20 lb 30 lb 50 lb	34 mg 34 mg 34 mg 34 mg 34 mg 34 mg 35 mg 41 mg	ASTM E617 Class 0 Weights, Electronic Balances
Scales and Balances ^{1,6} (SI Units)	Up to 500 mg (0.5 to 5) g (5 to 10) g (10 to 30) g (30 to 50) g (50 to 100) g (0.1 to 1) kg (1 to 2) kg (2 to 3) kg (3 to 5) kg (5 to 10) kg	1.7 µg 5.2 µg 8.4 µg 22 µg 24 µg 39 µg 2 mg 2.2 mg 2.3 mg 2.6 mg 3.9 mg	ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system.
Scales and Balances ^{1,6} (Avoirdupois)	Up to 0.5 oz (0.5 to 1) oz (1 to 2) oz (2 to 4) oz (4 to 8) oz 8 oz to 1 lb	0.000 049 oz 0.000 102 oz 0.000 201 oz 0.000 423 oz 0.000 811 oz 0.001 623 oz	ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales and Balances ^{1,6} (Avoirdupois)	(1 to 2) lb (2 to 3) lb (3 to 5) lb (5 to 10) lb (10 to 20) lb (20 to 22) lb	0.000 201 lb 0.000 309 lb 0.000 507 lb 0.001 014 lb 0.002 006 lb 0.003 086 lb	ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system.
	(50 to 1 200) lb	0.012 % of reading	NIST Class F weights and internal calibration procedure utilized for the calibration of the weighing system.
Pneumatic Absolute Pressure Devices ¹	Up to 25 psia (25 to 500) psia	0.001 9 psi 0.006 5 % of reading + 0.001 psi	Ruska 7250xi Pressure Controller/Calibrator
Pneumatic Pressure Gages ¹	(-15 to 15) psi (25 to 500) psi	0.001 6 psi 0.006 5 % of reading	Ruska 7250xi Pressure Controller/Calibrator
Hydraulic Pressure Gages ¹	(100 to 15 000) psig	0.009 % of reading	Ametek T-150 Deadweight Tester

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity – Source	(0 to 15) °C (10 to 75) %RH (75 to 95) %RH	0.5 %RH 0.65 %RH	Two-Pressure Humidity Generator
	(15 to 35) °C (10 to 95) %RH (35 to 70) °C (10 to 50) %RH (50 to 75) %RH (75 to 95) %RH	0.5 %RH 0.5 %RH 0.7 %RH 0.85 %RH	
Temperature – Source ¹	(-30 to 125) °C	0.036 °C	Liquid Bath, SPRT w/Indicator
Temperature – Measure ¹	(-190 to 0) °C (0 to 420) °C	0.012 °C 0.024 °C	SPRT w/Indicator
	(420 to 660) °C	0.035 °C	


Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source/Measure ¹	10 MHz	58 mHz	Rubidium Oscillator
Stopwatches, Timers	Up to 599 s/mon	59 ms/d	Vibrograf 4500 Timometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one, but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
3. The stated uncertainty is the laboratory's ability to source a fast rise pulse that is approximately 250 ps. In the typical application of measuring rise time of an oscilloscope, this value is one of the contributing factors, but other factors are derived from the DUT.
4. L = length in inches.
5. The reported measurand on the calibration certificate will be in the difference of the nominal and actual mass value in appropriate SI unit.
6. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
7. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.05.



Jason Stine, Vice President

