



Certificate of Compliance

Certificate: 70046465

Master Contract: 162029

Project: 70046465

Date Issued: 2015-10-07

Issued to: **Fluke Corporation**
6920 Seaway Blvd
Everett, Washington 98203
USA

Attention: Gary L. Allen

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: *Paul Tanaka*
Paul Tanaka, ASCT.

PRODUCTS

CLASS - C363106 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS - C363186 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE-Certified to US Standards

Pressure Calibrator/Controller, Model 2271A, Rated 100 – 240 V, 47 – 63 Hz, 100 W MAX.

Notes:

1. The above model is cord connected, Equipment Class 1, Pollution Degree 2
2. Mode of operation: Continuous
3. Environmental Conditions: Normal: 15 to 35 C, 2000 m max, 5% to 95% non-condensing



Certificate: 70046465

Project: 70046465

Master Contract: 162029

Date Issued: 2015-10-07

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CONDITIONS OF ACCEPTABILITY

- (1) The main supply cord set provided with the equipment must be an approved type acceptable to the authorities in the country where the equipment is sold.
- (2) The user replaceable mains (line) fuse must be an approved type acceptable to the authorities where the equipment is sold.
- (3) If at any time there is a conflict between the system safety provisions and any relevant local (national or regional) requirements, the local requirements always take precedence.



Supplement to Certificate of Compliance

Certificate: 70046465

Master Contract: 162029

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70046465	2015-10-07	Evaluation of Fluke Calibration 2271A Pressure Calibrator/Controller according to the requirements of CSA C22.2 No 61010-1-12 and UL 61010-1 (3rd Edition). (FCal)(C/US).